



CERTIFICATE OF CONFORMANCE

Calibration with NIST Traceable Certification

Customer	York
Cust Ref #	
Instrument	DPT810
Inst. SN.	A010007031

Date:	8/13/2026
Examiner:	TP

Ambient Temperature:	21.1°C
Relative Humidity	54%RH

Comments & Observations:

Test point Number	NIST Traceable Standard °C dp	Analyzer Under Test mA
1	-89.00	6.58
2	-80.12	7.62
3	-70.81	8.75
4	-60.23	10.14
5	-50.82	11.27
6	-40.91	12.65
7	-30.45	13.78
8	-21.90	14.86
9	-12.26	16.03
10	10.25	18.80

Data Traceable to
NIST Certified1
Chilled Mirror,
Buck Research
Model CR-1A
Serial #07
Measuring range:
(-110.0°C to 20.0°C)

We hereby certify that the analyzer detailed above has been tested and calibrated by the Undersigned, and its performance was found to meet our specifications unless noted otherwise in the comments section.

For and on behalf of PhyMetrix Ltd

For Quality Assurance Manager

Note: 1.For Chilled Mirror Certificate refer to the attached Manufacturer's Certificate