

Customer	YORK
Cust Ref #	
Instrument	DPT-990-Ex
Inst. SN.	EX10260902

Date:	8/28/2026
-------	-----------

Test Flow	2L/min
Ambient Temperature:	22.1°C
Relative Humidity	50%RH

Comments & Observations:

Test point Number	NIST Traceable Standard °C dp	Analyzer °C dp
1	-80.75	-80.1
2	-70.95	-70.2
3	-60.42	-59.7
4	-52.43	-51.9
5	-42.15	-41.8
6	-31.76	-31.4
7	-21.24	-21.1
8	-12.56	-11.8
9	12.19	12.5

We hereby certify that the analyzer detailed above has been tested and calibrated by the Undersigned, and its performance was found to meet our specifications unless noted otherwise in the comments section.

For and on behalf of PhyMetrix Ltd.



For Quality Assurance Manager