

Corrective and Preventive Action Record

Quality Management System · 2025-04 · Issue 2

1. Action

Corrective and preventive action logged against the field safety action recorded under reference R-2025-014.

Field	Entry
Record	CAPA-2025-019
Logged	2025-04-02
Action reference	R-2025-014
Action	Revise the seal interface qualification test sequence
Owner	Quality Management System

The action requires the qualification test sequence for the seal interface to be revised to include sustained thermal cycling at the upper end of the declared operating range, and the revised sequence to be adopted into the approved test plan.

2. Status

Progress against the action is reviewed monthly and recorded below. The entry at the most recent review is the current position.

Review	Position
2025-07	Draft sequence prepared
2025-10	Draft sequence under technical review
2026-01	Technical review comments being addressed
2026-03	In progress

The action is in progress. Adoption into the approved test plan follows completion of the technical review and is not yet scheduled.