

Supplier Quality Report — Batch 4471

Supplier Quality · 2024-08 · Issue 1

1. Batch identification

This report records the assessment of batch 4471 at incoming inspection and the disposition applied to the batch.

Field	Entry
Batch	4471
Material	Elastomer seal, housing interface
Received	2024-08-14
Quantity received	900
Quantity sampled	32

The batch was sampled in accordance with the incoming inspection plan for the material category. Sampling was performed on receipt and before the material was made available to production.

2. Test results

Sampled items were assessed against the material specification for hardness, compression set and dimensional conformity. Dimensional conformity and compression set were within specification across the sample.

Hardness measurements fell below the lower specification limit in 11 of the 32 items sampled. The batch was recorded as **seal hardness out of tolerance** and the sample extended in accordance with the inspection plan. Extended sampling confirmed the original finding across the batch.

Characteristic	Specification	Result
Hardness	Within declared band	Below lower limit, 11 of 32
Compression set	Within declared band	Conforming
Dimensional	Within drawing tolerance	Conforming
Extended sample	64 items	Finding confirmed

No further testing was performed. The material was not released to production at any point and no assemblies were built from this batch.

3. Disposition

The batch was rejected in full and quarantined pending return to the supplier. A supplier notification was raised on the date below and the material was collected against the return note recorded here.

Field	Entry
Disposition	Reject in full
Quarantined	2024-08-16
Supplier notified	2024-08-16
Return note	RN-2024-0812

Replacement material was requested against the original order. Receipt of the replacement is recorded separately on the goods receipt register.