

Sequencing assay report

The Illumina run QC gate failed on flow cell FC-4412 because the cluster density fell below the acceptance threshold.

The operator repeated the library preparation with fresh adapters and the second flow cell passed every quality gate.