A-A-0051948(DM)
22 April 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

DEFIBRILLATOR AND CARDIOSCOPE
(Three Lead, with Recorder)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a defibrillator monitor-recorder with three lead system that will be required to respond to numerous cardiac crisis and/or trauma situations in a hospital environment.

Salient characteristics:

The defibrillator system shall be an integral or modular defibrillator ECG monitor-recorder, which functions from self - contained batteries and AC line power. Units which are modular in design and can be separated shall consist of only two separate components: a defibrillator and a monitor-recorder. Both modules shall be independently powered and shall securely interconnect into a physically contiguous unit without the use of tools, without the need for auxiliary cables, and in such a manner as not to cause excessive stress on the electrical connectors. The system, or both modules, shall have an integral battery charger. The system shall include a monitor quality ECG recorder, a heart rate meter, defibrillator paddles to include one pair of external pediatric paddles, and one pair external adult paddles (the paddle sizes should conform with the American Heart Association specifications for minimum width, length, and surface area), synchronized cardioversion capability and accessories as specified herein. Provision shall be made for out-of-the way paddle storage.

Reference Documents. The system as either independent modules and/or as an integral unit shall comply with applicable sections and requirements (not addressed herein) of the following publications (current editions):

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

a. American National Standards Institute/Association for the Advancement of Medical Instrumentation DF2, ANSI/AAMI DF2 - American National Standard for Cardiac Defibrillator Devices, with the exception of paragraph 3.2.1.2, Energy Range.

b. ANSI/AAMI ESl - American National Standard, Safe Current Limits for Electromedical Apparatus.

c. ANSI/AAMI ECl3 - American National Standard for Cardiac Monitors, Heart Rate Meters and Alarms.

d. Underwriters Laboratories Standard 544, UL 544 - Standard for Safety of Electric Medical and Dental Equipment.

e. National Fire Protection Association 99, NFPA 99 - Standard for Health Care Facilities.

f. American Heart Association, AHA publication titled, "Recommendations for Standardization of Leads and of Specifications for Instruments in Electrocardiography and Vectorcardiography".

g. Association for the Advancement of Medical Instrumentation, AAMI ECGC, Standard for ECG Connectors.

Environmental conditions: All equipment must be capable of enduring environments the equipment is expected to experience in shipment, storage, service and use, as follows:

Atmospheric Pressure:	500 mm Hg to 700 mm Hg
Relative Humidity (R.H.):	95 percent (non-condensing)
Temperature:	
Operational	32°F (0°C) to 114°F (45°C)
Storage	-23°F (-30°C) to 149°F (65°C)

All units shall be easily cleaned with available hospital cleansing solutions, adequately protected from moisture, and be designed to withstand normal abuse, such as shock and impact, frequently encountered in the hospital environment.

Construction. The unit shall be of a rugged construction capable of withstanding hazards of normal use in a hospital setting.

Weight. The total weight of the defibrillator/monitor-recorder including the batteries, charger, paddles, etc., shall not exceed forty (40) pounds.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:03Z
Check the source to verify that this is the current version before use.

ECG Monitor/Recorder.

Input. The instrument shall be capable of providing an ECG through the defibrillator paddles or via a three (3) lead patient cable having a six (6) pin cable connector, conforming to AAMI ECGC.

The patient cable leads supplied with the monitor shall be color coded in accordance with AHA publication.

The patient cable shall be shielded and have an approximate minimum length of 10 feet (3.03m). Individual lead length shall be approximately three feet (.91m). Total cable length with patient leads should not exceed 13 feet (3.95m). Individual leads shall be color coded at the end of connectors.

Cardioscope-display. The cardioscope display shall have the following characteristics:

1. Shall be non-fade type display measuring no less than 4" diagonally for ease of operator view.

2. Unit shall have either a freeze control or a delayed recording capability for detailed analysis of arrhythmias.

3. Shall have a monitor quality frequency response of no less than 0.5 Hz to no less than 30 Hz and no greater than 50 Hz (measured at -3 db).

4. Shall have sweep speed of 25 mm/second.

5. Shall have an easily adjustable audible systole control.

6. Shall show a representation of heart pacemaker pulse on the CRT, but not count the signal on the digital heart rate display to avoid false alarms or operator error.

7. Shall have a lead selector switch enabling the operator to select lead I, II, and III without switching patient electrode sites.

8. Shall have an ECG gain control that is variable and operator adjustable.

ECG Amplifier. The ECG amplifier shall have the following characteristics:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

1. Unit shall have a three (3) lead (I, II, III) select and a frequency response of 0.5 to no less than 30 Hz and no greater than 50 Hz (measured at -3 db point).

2. Shall have common mode rejection ratio of no less than 100 dB with respect to chassis ground with 5 K ohm imbalance at 60 Hz; 65 dB minimum with respect to isolated ground.

3. Input system, amplifier electronics and mechanical configuration of the ECG amplifier shall comply with requirements of UL 544.

4. Amplifier shall be protected from defibrillator counter shocks up to a level of 400 delivered joules.

5. Unit shall have a 1 mV reference signal for ECG calibration.

6. Shall have an electrically isolated patient input circuit in accordance with ANSI/AAMI ES1.

Recorder. The recorder shall have the following characteristics:

1. Recorder shall have a frequency response of 0.5 to no less than 30 Hz and no greater than 50 Hz (-3 db point).

2. Sweep speed shall be 25 mm/sec.

3. Shall utilize federal standard stock ECG paper with 50 mm width by 30 m (100 ft) length.

4. The recorder shall be an integral part of the monitor/cardioscope. Stand alone recorders requiring electronic cable interface are not acceptable.

5. A low paper indication shall be provided, either by the instrument or in the form of a mark on the paper, which indicates the needs to replace paper supply.

Heart Rate. The heart rate meter shall have the following characteristics:

1. The heart rate detector shall operate with normal R waves of either positive or negative polarity.

2. The heart rate shall be displayed continuously via an LED/LCD digital display.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

3. Shall be capable of displaying a range of 30 - 240 BPM with an accuracy of ±5 percent.

4. The update frequency of heart rate display shall be 2 - 5 seconds and shall not be affected by freezing the ECG trace.

Defibrillator. The defibrillator shall have the following characteristics:

1. Defibrillator pulse shape shall be clinically proven (i.e. Lown, Edmark or Trapezoidal) as defined in AAMI DF2.

2. The defibrillator shall provide for AC and battery operation, with integral battery charging system. When modular units are provided, each must have its own recharging system.

3. The defibrillator shall provide a continuously adjustable energy range of at least 0 to 360 delivered Joules or a selection of at least six discrete energy steps from 5 to 360 Joules. When delivered into a 50 ohm load, at least three of these outputs shall be less than 110 Joules.

4. Charge time to maximum energy output shall not exceed 12 seconds or less at 82.5°F (28°C)

5. The unit shall have a synchronizer with visual indication on activation of this control. Defibrillator shall discharge within 30 ms after a patient - generated R-wave.

6. Synchronization marker shall appear on the cardioscope and the paper ECG strip when cardioversion is attempted.

7. Unit shall switch back to defibrillation mode after each cardioversion attempt.

8. The defibrillator shall be capable of delivering 10 pulses with maximum energies into a 50 ohm resistive load within 5 minutes, at 32°F (0°C).

9. Unit shall have pushbutton discharge controls on both paddle handles which must be simultaneously activated for discharge. (Inadvertent/accidental discharges eliminated).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

10. Paddle shall have a remote button to initiate energy charge.

11. Unit shall internally discharge stored energy when power is turned off.

12. Paddles shall be capable of ECG pick-up through the paddle surface.

13. Defibrillator output shall be isolated from ground when operated in the AC power source mode.

ELECTRICAL CHARACTERISTICS.

A. The defibrillator/monitor-recorder shall operate over the full tolerance range of one or more of the following line voltage and frequency combinations as specified:

1. 115 VAC ± 10 percent, 50 ± 2 Hz.

2. 115 VAC ± 10 percent, 60 ± 2 Hz.

3. 230 VAC ± 10 percent, 50 ± 2 Hz.

Dual voltage and hertz on all units are desireable but not mandatory.

B. The defibrillator/monitor-recorder shall be capable of full operation while powered by internal batteries for the following functions:

1. Monitor - minimum of 2.25 consecutive hours during continuous Monitor operation.

2. Monitor-recorder - minimum of 45 minutes during use of the monitor and strip chart recorder.

3. Defibrillator:

a. Thirty defibrillator discharges at maximum energy into a 50 ohm load at an ambient temperature of 82.5°F (28°C).

b. Charge time to maximum energy output shall not exceed 12 seconds at 82.5°F (28°C).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

c. Ten defibrillator discharges at maximum energy into a 50 ohm load at an ambient temperature of 32°F(0°C) within a period of five minutes.

C. Battery Charging System:

1. The charging system must provide sufficient energy to fully power the unit while simultaneously charging fully depleted batteries.

2. Charge time for fully depleted batteries to fully charged state shall not exceed sixteen hours. The unit shall have a level indicator to show the level to which the batteries have been recharged after the 16 hours.

3. A charging lamp/indicator shall be provided for each battery used in the system.

4. A low voltage indicator shall be provided for each battery used in the system.

Accessories. As a minimum, the following accessories shall be provided:

(a) Two (2) each patient cables.

(b) One (1) spare stylus, if applicable (w/replacement tool, if required).

(c) One ECG ruler.

(d) Three (3) pre-jelled electrodes.

(e) Two (2) spare batteries of each type used.

(f) Three (3) packages of recording paper, marked with Federal/National Stock Number.

(g) One (1) tube electrode jelly.

(h) One (1) accessory bag.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Operational and Service Data Manuals shall be furnished as follows:

a. Information Content:

The contractor shall furnish manuals, handbooks and/or brochures containing complete operation, installation and service/maintenance instructions (including pictures or illustrations, as necessary) with complete schematics and wiring diagrams. All manuals, handbooks, and/or brochures will be written in the English language and all schematics and wiring diagrams will use American electrical and electronics symbols. The manuals will include electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name and quantity required. All service notes, service memos, etc, pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included. These instructions shall be as a minimum, that furnished to service engineers (or servicemen) that normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components.

b. Quantities, Distribution and Schedules:

Two (2) complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, then only a total of 2 service manuals need to be supplied to each consignee. However, 2 operation manuals still need to be supplied with each unit.

Workmanship. The defibrillator system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One defibrillator system, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

First article inspection. First article inspection shall be performed on the defibrillator system when a first article sample is required. This inspection shall include examination and tests specified herein.

Sampling.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one defibrillator system.

TABLE I. Sampling for Examination

	Inspection level	AQL (defects per hundred units)
Major Defects	II	1.0
Total Defects (Major and Minor combined)	II	4.0
For Dimensional Defects	S-1	2.5
For All Tests	S-2	1.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

For tests. Sampling for tests shall be conducted as follows:

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a first article sample has successfully passed all tests, performance test shall not be required at acceptance.

Examination. The defibrillator/monitor-recorder shall be examined to determine compliance with the requirements of this specification in accordance with the classification of defects in table II.

TABLE II. Categories and Defects *

Categories	Defects
Major	
101	Component missing.
102	Component damaged including (but not restricted to) cracks, checking, breaks and deformed areas.
103	Exposed sharp edges or corners accessible to operator during normal use or cleaning.
104	Synchronization switch not recessed or protected from accidental engagement.
105	Power switch and charge switch/button not clearly labeled and readily visible.
106	Paddle storage not neat and "out of the way".
107	Modules do not fit together correctly.
108	Any knobs, switches, or buttons missing or misaligned.
109	Stylus bent or misaligned.
Minor	
201	Identification marking not complete, permanent, and legible.
202	Finish not free of chips, cracks, scratches, and other appearance defects.
203	Any component part not secured in place.

*Examination shall not be limited to the classified possible defects listed herein.

Tests. Tests shall be conducted to determine compliance with table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Table III. Tests.

I. Leakage current:

A. Defibrillator

1. Chassis Leakage:
Leakage current measured between any exposed conductive surface and ground shall not exceed 100 uA.

2. Paddle Leakage:
Leakage current measured between an external paddle electrode and ground shall not exceed 100 uA.
Leakage current measured between an internal paddle electrode and ground shall not exceed 50 uA.

3. Conditions:
Leakage current measurements will be made under all of the following conditions:

a) Power on and off.
b) Defibrillator in both charged and uncharged state.

B. ECG Monitor/Recorder

1. Chassis Leakage:
Leakage current measured between any exposed conductive surface and ground shall not exceed 100 uA.

2. Lead leakage (source current):
Leakage current measured between each lead and ground, all leads and ground, and between any two leads shall not exceed 10uA.

3. Lead leakage (sink current):
Sink current measured between each lead at a potential of 120V and ground shall not exceed 20uA. This test shall be performed under the conditions listed above except the ground remains intact.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

C. Test Device

All leakage current measurements will be performed with a test device equipped with an AAMI Standard test load.

II. Strain Relief (Line Cord)

Gradually apply a 35 lb. pull in a direction that would be the most likely to pull the cord from the enclosure and maintain for one minute. The strain relief is not acceptable if there is such movement of the cord as to indicate that stress is being transmitted to terminals, splices or interior wiring.

III. Air Clearance Spacing

Measurements shall be made, using an appropriate measuring instrument, to verify isolated patient connected circuitry is spaced at a minimum of 1/8 inch from non-isolated circuitry.

IV. Defibrillator Paddle Insulation

A. Apply aluminum foil to the entire gripping surface of one paddle with sufficient pressure to make it conform to the surface.

NOTE: The gripping surface for verification purposes is defined as the approximately cylindrical area around each handle where the operator's fingers grasp the paddle, but not beyond any barrier or curvature that is obviously intended to limit hand placement, including areas one inch in diameter centered on any control located on the paddle, representing the location(s) of the operator's thumb and/or finger.

B. With the paddles connected to the defibrillator, connect the high potential lead of a 60 Hz (or 50 Hz, if preferred) high potential test source to the aluminum foil. Connect low-potential or ground lead to the paddle electrode. Slowly adjust the peak voltage amplitude of the test source to equal 1.5 times the maximum voltage occurring on the defibrillator energy storage capacitor. For a sine wave this corresponds to an rms valve of 1.06 times the capacitor voltage. Maintain this voltage for 1 minute. Repeat this test with the low-potential lead connected to the other paddle electrode.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

C. There shall be no voltage breakdown, and the 60 Hz leakage current shall not exceed 250 uA peak (177 uA rms); if a 50 Hz test source is used, the current shall not exceed 210 uA peak (148 uA rms).

D. Repeat (A) through (C)

V. ECG Monitor/Recorder Sweep/Chart Speed

Measure the monitor sweep speed and the recorder paper speed by applying a signal of known frequency to the patient cable. Monitor sweep speed shall be within ±10 percent of 25 mm/sec. and recorder chart speed within ±5 percent of 25 mm/sec.

VI. ECG Amplifier Calibration

The accuracy of the 1 mV internal calibration pulse shall be measured by comparing the amplitude of this pulse to that of a calibrated 1 mV pulse applied to the patient cable from a pulse generator. It shall be accurate to ±15 percent.

VII. ECG Amplifier Common Mode Rejection

Common mode rejection is measured by comparing the amplitude of a 1 mV differential signal applied to the monitor to the amplitude of a higher voltage sinusoidal signal of 60 Hz applied between the junction of the leads and ground. Measured common mode rejection shall be at least 100 dB with respect to chassis ground with 5K ohms imbalance between leads, and at least 65 dB with respect to isolated ground.

VIII. ECG Monitor and Recorder Frequency Response

Frequency response of the ECG monitor and recorder is measured by applying a 1 mV peak-to-peak differential sinusoidal signal to the patient cable, and varying the frequency. The bandwidth is defined by the upper and lower frequencies at which the trace deflection amplitude decreases to 0.707 times the maximum value. This bandwidth shall be 0.5 Hz to anywhere between 30 to 50 Hz inclusive, on the monitor and recorder.

IX . Defibrillator Overload Protection.

Overload protection shall comply with those requirements specified in ANSI/AAMI EC13,except the defibrillator energy level shall be 400J.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

X. Battery Capacity

Battery Capacity will be tested by verifying 2.25 hours of continuous cardioscope monitoring or 45 minutes of continuous recording of a heart rate of 60 BPM. Defibrillator battery capacity will be tested by verifying 30 maximum energy discharges at 28°C.

XI. Defibrillator Energy Accuracy

Connect the defibrillator paddles to a defibrillator analyzer equipped with a 50 ohm resistive load. Charge the defibrillator to a preselected energy level, then discharge into the analyzer. Repeat for each available energy level setting.

Selected Energy Setting (Joules)	Measured Energy (Joules)	Percent Error

To be in compliance, the measured energy must be within ±15 percent (or 4 Joules, whichever is greater) of the selected energy.

XII. Defibrillator Charge Time

Using a stopwatch, measure the time between activation of the charge control and full charge indication for the first and last of 10 sequential charges and discharges all at maximum energy setting. This time shall not exceed 12 seconds.

XIII. Defibrillator Discharge Pulse Rate

Apply the paddle electrodes to a defibrillator tester and select the maximum energy level. Simultaneously initiate the charge and a timer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

XIV. Synchronized Discharge

Connect a signal generator to the ECG monitor. Measure the time between the leading edge of square pulse and the leading edge of the defibrillator's synchronized discharge pulse. This time can be displayed automatically on a defibrillator analyzer if so equipped, or measured directly from an oscilloscope, and shall be within 30 milli-seconds.

XV. Humidity

Manufacturer shall supply the following, verifying compliance with the test described below:

a. Copy of test procedure as specified below in test method.

b. Unit type.

c. Serial number.

d. Manufacturer and model number of chamber used.

e. Leakage current test results after 48 hours.

f. Energy accuracy test results after 48 hours

g. Test Method

1. Place the unit in a test chamber such that leakage current and delivered energy accuracy tests can be performed external to the chamber.

2. Adjust the temperature of the chamber to 32°F (0°C).

3. Permit sufficient time for the temperature "inside the unit" to stabilize prior to proceeding.

4. After temperature stabilization, adjust the R.H. of the chamber to 95 (±5) percent and permit sufficient time for humidity stabilization.

5. After 48 hours of conditioning, source and sink leakage current tests and energy accuracy tests are to be performed. The unit is in compliance if leakage currents and energy accuracy are within the specified limits.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, satisfy the performance tests, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4085

Location: CID/DEF/H1D-21

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-0051948(DM)	DEFIBRILLATOR AND CARDIOSCOPE (Three Lead, with Recorder)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.