MIL-B-17675B
AMENDMENT-2 1/
10 July 1969
SUPERSEDING
Amendment-1
28 July 1967

MILITARY SPECIFICATION

BREATHING APPARATUS, OXYGEN GENERATING

This amendment forms a part of Military Specification MIL-B-17675B, 29 April 1966, and is mandatory for use by all Departments and agencies of the Department of Defense.

Pages 1 and 2

2.1: Under "SPECIFICATIONS, FEDERAL" add: "L-P-516 - Plastic Sheet and Plastic Rod, Thermosetting, Cast."

2.1: Under "SPECIFICATIONS, MILITARY" delete "MIL-P-77".

2.1: After "DRAWINGS" delete "BUREAU OF SHIPS" and substitute "MILITARY".

2.1: After "DRAWINGS, MILITARY" delete "S9300-921878" and "S9300-921879" and substitute "608-921878" and "608-921879".

Page 2

2.2, "OFFICIAL CLASSIFICATION COMMITTEE": Delete and substitute:

"UNIFORM CLASSIFICATION COMMITTEE
Uniform Freight Classification Rules

(Application for copies should be addressed to the Uniform Classification Committee, 202 Union Station, 516 West Jackson Boulevard, Chicago, Illinois 60606.)"

Page 3

3.3.1.1, listing, column 2: Delete "5.50" and substitute "550".

3.3.2, line 6: Delete "MIL-P-77" and substitute "L-P-516".

Page 5

3.3.8, item (b): Delete "plunger" and substitute "breastplate".

3.3.9, line 3: Delete "S9300-921878" and "S9300-921879" and substitute "608-921878" and "608-921879".

3.3.9: Delete the last sentence.

3.3.10: Delete and substitute:

"3.3.10 Timing device. The timing device shall be of the general type and design shown on Drawing S9300-921869. The timing device shall sound a warning signal in the form of a continuous ring of 8 minus 1 seconds minimum duration. This signal shall operate when a 1-hour period has elapsed or any intermediate setting when the timer is fully wound and backed off to the appropriate time setting. The timer shall be spring-actuated. The spring shall be made of materials known for the ability to resist salt water corrosion. The timer shall not deviate more than plus or minus 5 minutes for the 1-hour period."

1/ CHANGES FROM PREVIOUS ISSUE. THE OUTSIDE MARGINS OF THIS DOCUMENT HAVE BEEN MARKED "#" TO INDICATE WHERE CHANGES (DELETIONS, ADDITIONS, ETC.) FROM THE PREVIOUS ISSUE HAVE BEEN MADE. THIS HAS BEEN DONE AS A CONVENIENCE ONLY AND THE GOVERNMENT ASSUMES NO LIABILITY WHATSOEVER FOR ANY INACCURACIES IN THESE NOTATIONS. BIDDERS AND CONTRACTORS ARE CAUTIONED TO EVALUATE THE REQUIREMENTS OF THIS DOCUMENT BASED ON THE ENTIRE CONTENT AS WRITTEN IRRESPECTIVE OF THE MARGINAL NOTATIONS AND RELATIONSHIP TO THE LAST PREVIOUS ISSUE.

FSC 4240

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

Page 10

4.6.2.2.2.7.1: Delete and substitute:

"4.6.2.2.2.7.1 Rubber irritation test shall be conducted under surveillance of a properly qualified physician. Should the contractor wish to avail himself of a Government physician, then the contractor shall notify the Government physician via the Defense Contract Administration Service representative of the time that this test will be conducted at least ten days prior to commencement of the test."

Page 11

4.6.6.2, line 2: Delete "anr" and substitute "air".

Custodians:
Army - MU
Navy - SH
Air Force - 84

Review activities:
Army - MU, MD
Navy - SH, MS

User activities:
Army - ME
Navy - MC, YD

Preparing activity:
Navy - SH
(Project 4240-0305)

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969-393-065/S-1385

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.