MIL-C-17671E
23 September 1986
SUPERSEDING
MIL-C-17671D(SH)
24 September 1981
(See 6.6)

MILITARY SPECIFICATION

CANISTER, OXYGEN BREATHING APPARATUS, FOR TYPE A-4

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a canister for the type A-4 oxygen breathing apparatus.

1.2 Part number. The part number for the canister covered by this specification shall be M17671-1.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

FEDERAL

PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-640 - Boxes, Fiberboard, Corrugated, Triple-Wall.

MILITARY

MIL-P-116 - Preservation, Methods of.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commander, Naval Sea Systems Command, SEA 55Z3, Department of the Navy, Washington, DC 20362-5101 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 4240
DISTRIBUTION STATEMENT A Approved for public release; distribution unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-313 - Material Safety Data Sheets, Preparation and the Submission of.
FED-STD-595 - Colors.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-171 - Finishing of Metal and Wood Surfaces.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.
MIL-STD-831 - Test Reports, Preparation of.
MIL-STD-1472 - Human Engineering Design Criteria for Military Systems, Equipment and Facilities.

2.1.2 Other Government drawing and publication. The following other Government document and drawing form a part of this specification to the extent specified herein. Unless otherwise specified, the issues shall be those in effect on the date of the solicitation.

DRAWING

NAVAL SEA SYSTEMS COMMAND (NAVSEA)
803-5184258 - Oxygen Breathing Apparatus, Canister, Quick Starting.

PUBLICATION

NIOSH/OSHA Publication No. 78-210
Pocket Guide to Chemical Hazards.

(Application for copies should be addressed to Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.)

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein (except for associated detail specifications, specification sheets or MS standards), the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 First article. When specified in the contract or purchase order, sample canisters shall be subjected to first article inspection (see 4.3 and 6.3).

3.2 Design and details. The design of the canister shall be in accordance with Drawing 803-5184258. The canister shall not utilize compressed gas as a source of breathable gas.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

3.3 Markings. Markings on the canisters shall be in accordance with Drawing 803-5184258. If the external surface temperature exceeds 140 degrees Fahrenheit (°F) (60 degrees Celsius (°C)) when tested in 4.6.2.3, a caution shall be imprinted on the canister label warning of the burn hazard when unprotected skin is exposed to this temperature in accordance with MIL-STD-1472.

3.4 Canister performance.

3.4.1 Canister output. The canister shall supply breathable gas for approximately 60 minutes with a 50-minute minimum. Breathable gas shall be defined as no less than 20.0 (dry atmosphere) volume percent oxygen, no more than 1.5 percent inert gases, as defined by the periodic table and no more than 0.5 percent impurities as defined by NIOSH Publication No. 78-210. The canister shall contain no impurities which will generate any gas or product which will have any harmful physiological effect on personnel under any conditions of use. The average carbon dioxide volumetric content of the gas produced by the canister shall not exceed 1.5 percent and shall conform to the limits as specified in 4.6.2.2.1. The temperature of the gas shall be less than 115°F (46°C) when measured as specified in 4.6.2.2.3.

3.4.1.1 Canister external surface temperature. The external surface temperature of the canister shall not exceed 475°F (246°C) when tested as specified in 4.6.2.3.

3.4.2 Gasflow resistance. The resistance of each fully assembled canister to gasflow shall not be greater than 0.72 inch 18 millimeters (mm) of water column height when tested in accordance with 4.6.2.1. This requirement shall apply only to canisters that utilize a flow-through concept (for example, KO_2 canisters) to produce breathable gas.

3.5 Breathable gas starting process. The canister shall contain a breathable gas starting process which shall allow for continuous production of breathable gas.

3.5.1 Overall requirements. The starting process shall operate within an ambient temperature range of minus 22 to 212°F (minus 30 to 100°C), when tested in accordance with 4.6.3.1 and 4.6.3.2. The process shall supply a minimum of 0.66 gallon (2.5 liters (L)) of breathable gas at standard temperature and pressure (STP), 30.4 inches (760 mm) Hg, 32°F (0°C) in the first minute of operation and shall continue to evolve breathable gas until 2.5 gallons (9.5 L) of breathable gas at STP have been evolved within a maximum of 3.5 minutes. When the process is initiated at 212°F, it shall evolve 2.5 gallons of breathable gas at STP in not less than 45 seconds. Carbon monoxide generated from the starting process shall be less than 0.02 percent by volume when tested in accordance with 4.6.3.3.

3.5.2 Sparking. The breathable gas starting process shall be non-sparking when tested in accordance with 4.6.4.

3.5.3 Salt spray. The breathable gas starting process shall be resistant to salt water corrosion when tested in accordance with 4.6.5.

3.6 Weight. The fully assembled canister shall weigh no more than 75.4 ounces (2155 grams).

Source: https://assist.dla.mil -- Downloaded: 202610-18T10:27Z
Check the source to verify that this is the current version before use.

3.7 Color. The color of the canister shall be green, in accordance with color no. 14110 or 14187 of FED-STD-595 and as specified in Drawing 803-5184258.

3.8 Finish. The finish for the canister shall be in accordance with finish no. 22.9 of MIL-STD-171 and as specified in Drawing 803-5184258.

3.9 Identification. The bottom of each canister shall be permanently imprinted with the lot number, date of manufacture, and contract number without damaging the material.

3.10 Material safety data sheet. The contracting activity shall be provided a material safety data sheet (MSDS) at the time of contract award. The MSDS is form OSHA-20, found in and part of FED-STD-313. The MSDS shall be included with each shipment of the material covered by this specification.

3.11 Workmanship. The canister shall be free from any trace of contamination such as dirt, grease, oil, foreign matter, flammable substance, and damage such as cracks, dents, burrs, tears, deformations, punctures, scratches, rust, and any other visible defects.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling in quality conformance does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to acceptance of defective material.

4.2 Classification of inspections. The inspection requirements specified herein are classified as follows:

(a) First article inspection (see 4.4 and 6.3).
(b) Quality conformance inspection (see 4.4).

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

4.3 First article inspection. One hundred sample canisters shall be furnished for first article inspection when specified (see 6.3). If any sample fails any examination or test, it shall constitute failure of first article inspection.

4.3.1 For examinations and nondestructive test. Forty canisters shall be randomly selected from the 100 sample canisters, and subjected to the examinations and test in table I.

TABLE I. First article inspection (examinations and nondestructive test).

Examination or test	Requirement paragraph	Test paragraph	Number tested
Canister weight	3.6	4.6.1	40
Visual and dimensional examination	3.2, 3.3, 3.7, 3.8, 3.9, 3.11	4.5	40

4.3.2 For destructive tests. From the 40 representative canisters selected at random in 4.3.1, 26 will be selected for the tests specified in table II.

TABLE II. First article inspection (destructive tests).

Examination or test	Requirement paragraph	Test paragraph	Number tested
Resistance to gasflow	3.4.2	4.6.2.1	13
Duration	3.4.1	4.6.2.2.1	13
Performance purity	3.4.1	4.6.2.2.2	13
Temperature of canister produced gas	3.4.1	4.6.2.2.3	13
External surface temperature	3.4.1.1	4.6.2.3	13
Breathable gas starting process			
Breathable gas evolution at minus 22°F (minus 30°C)	3.5.1	4.6.3.1	13
Breathable gas evolution at 212°F (100°C)	3.5.1	4.6.3.2	13
Purity	3.5.1	4.6.3.3	13
Canister sparking	3.5.2	4.6.4	13
Salt spray	3.5.3	4.6.5	13

4.4 Quality conformance inspection. Quality conformance inspection shall consist of the visual and nondestructive examination specified in 4.4.1.1 and group A, B, and D tests specified in tables III, IV, and V. Group C tests are not applicable. When specified in the contract or purchase order, a test report shall be prepared in accordance with MIL-STD-831 (see 6.2.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

TABLE III. Group A examinations test.

Examinations and test	Requirement paragraph	Test paragraph
Canister weight	3.6	4.6.1
Visual and dimensional examination	3.2, 3.3, 3.7, 3.8 and 3.11	4.5

TABLE IV. Group B test.

Test	Requirement paragraph	Test paragraph
Resistance to gasflow	3.4.2	4.6.2.1

TABLE V. Group D tests.

Test	Requirement paragraph	Test paragraph
External surface temperature	3.4.1.1	4.6.2.3
Temperature of canister produced gas	3.4.1	4.6.2.2.3
Canister sparking	3.5.2	4.6.4
Duration	3.4.1	4.6.2.2.1
Performance purity	3.4.1	4.6.2.2.2
Breathable gas starting process		
Breathable gas evolution at minus 22°F (minus 30°C)	3.5.1	4.6.3.1
Breathable gas evolution at 212°F (100°C)	3.5.1	4.6.3.2
Purity	3.5.1	4.6.3.3
Salt spray	3.5.3	4.6.5

4.4.1 Lot. The lot shall consist of all canisters offered for delivery at one time not to exceed a quantity of 500 canisters.

4.4.1.1 Sampling for examination. A random sample of assembled canisters shall be selected from each lot in accordance with MIL-STD-105, inspection level II, for the examination specified in table VI. The acceptable quality level (AQL) for single tightened, shall be 1.0 for major defects and 2.5 for minor defects. Any item in the sample containing one or more defects shall not be offered for delivery and if the number of nonconforming items in any sample exceeds the acceptance number for that sample, this shall be cause for rejection of the entire lot. If any sample contains a critical defect, the entire lot shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

TABLE VI. Classification of defects.

Categories	Defects
Critical:	No defects permitted
1	Evidence of oil, grease, or flammable substances.
Major:	AQL 1.0
101	Breathable gas starting process defective or damaged.
102	Lot number, date of manufacture, or contract number missing.
103	Instruction markings missing, illegible, incorrect, or not as specified.
104	Identification and caution markings missing, illegible, incorrect, or not as specified.
105	Finish missing, incomplete, damaged, or not as specified.
106	Body and neck bent, cracked, or not as specified.
Minor:	AQL 2.5
201	Tear-off cap defective, not secure on neck, or not as specified.
202	Color not as specified.
203	Workmanship not as required.

4.4.1.2 Sampling for tests.

4.4.1.2.1 Sampling for group A examinations and test (nondestructive). A random sample of assembled canisters shall be selected from each lot in accordance with MIL-STD-105 inspection level S-3, AQL 2.5, for single tightened, and tested as specified in table III. Any item in the sample containing one or more defects shall not be offered for delivery. If the number of nonconforming items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

4.4.1.2.2 Sampling for group B tests (non-destructive). A random sample of assembled canisters shall be selected from each lot in accordance with MIL-STD-105 inspection level S-3, AQL 2.5, for single tightened, and tested as specified in table IV. Any item in the sample containing one or more defects shall not be offered for delivery. If the number of nonconforming items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

4.4.1.2.3 Sampling for group C tests. None required.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

4.4.1.2.4 Sampling for group D tests (destructive). A random sample of assembled canisters shall be selected from each lot in accordance with MIL-STD-105 inspection level S-3, AQL 2.5, for single tightened, and tested as specified in table V. Any item in the sample containing one or more defects shall not be offered for delivery. If the number of nonconforming items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

4.5 Examinations.

4.5.1 Visual. Sample canisters shall be visually examined for conformance to the requirements specified in 3.3, 3.7, 3.8, 3.9 and 3.11.

4.5.2 Dimensions. Sample canisters shall be examined for conformance to the requirement specified in 3.2.

4.6 Tests.

4.6.1 Weight. Sample canisters shall be tested for weight to determine conformance to the requirement in 3.6.

4.6.2 Resistance to gasflow and performance tests.

4.6.2.1 Resistance to gasflow. Each sample canister selected shall be tested using a continuous gasflow of 22.5 gallons per minute (gal/min) (85 L per minute (L/min)). Resistance shall be measured within the first minute of gasflow through the canister. Each canister shall be tested for conformance to 3.4.2. A dry or inert gas test shall be used in making this determination.

4.6.2.2 Performance test. After the gasflow test, the canister performance test shall be conducted. For this test, the canister shall be properly installed in a type A-4 oxygen breathing apparatus that has had the facepiece removed from the point where the valve housing attaches to the corrugated breathing tubes. The ends of the corrugated tubes shall be connected to an automatic breathing metabolic simulator (ABMS). The canister shall be tested for at least 50 minutes using the following metabolic parameters:

(a) Oxygen consumption rate - 0.36 gal/min (1.35 L/min)
(b) CO_2 production rate - 0.30 gal/min (1.15 L/min)
(c) Ventilation rate - 8.4 gal/min (32.00 L/min)
(d) Respiratory frequency - 22.0 breaths/min
(e) Exhaled humidity - 80 percent saturated
(f) Exhalation temperature - 95°F (35°C)

Test samples shall be taken at the point where the corrugated breathing tubes and the ABMS are connected. The canister performance shall meet the requirements of 4.6.2.2.1 through 4.6.2.2.3.

4.6.2.2.1 Duration. The carbon dioxide volume content of the gas produced by the canister shall be monitored and recorded continuously during the performance test cycle. If carbon dioxide limits specified herein are not exceeded, the duration of the test shall be determined when the breathing resistance reaches 1.6 inches (40 mm) of water column height. The duration of the test will

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

be determined by carbon dioxide levels in the event that the average carbon dioxide volume content over the period of the test has reached 1.5 percent or any instantaneous reading exceeding 5 percent. Sample canisters shall conform to the requirements of 3.4.1.

4.6.2.2.2 Performance purity. The test samples of the canister output shall be monitored for purity as specified in 3.4.1. The test samples shall be taken within the first minute of the performance test cycle and shall continue to be drawn at 10 minute intervals until completion of the test cycle as specified in 4.6.2.2.1. Samples shall conform to the requirements of 3.4.1.

4.6.2.2.3 Temperature. At the point where the test samples are taken, the temperature of the canister output shall be monitored and recorded continuously to determine conformance to 3.4.1.

4.6.2.3 Canister external surface temperature. The external surface temperature of the canister shall conform to 3.4.1.1 at any point during the gas evolution process. Test samples shall be taken at contact points around the body of the canister as follows: on the bottom, near the oxygen starting process; on the top, at two points equally distant from the neck; on the concave face, at two distinct points along the vertical center line separated by 4 inches (10.16 centimeters (cm)) and on the convex faces, at two distinct points along the vertical centerline separated by 4 inches.

4.6.3 Breathable gas starting process. The breathable gas starting process shall be tested by installing the canister in the ABMS to conduct tests in 4.6.3.1 through 4.6.3.3. Test samples shall be taken at the same point specified in 4.6.2.2. Test samples shall be taken within the first minute after starting the canister and shall continue to be drawn at 10 minute intervals until completion of the test.

4.6.3.1 Breathable gas evolution at minus 22°F (minus 30°C). Each sample canister selected for test shall be thoroughly chilled to minus 22°F. When the starting process of each canister has been initiated, it shall conform to 3.5.1. The starting process shall allow the continuous production of breathable gas as specified in 3.4.1.

4.6.3.2 Breathable gas evolution at 212°F (100°C). Each sample canister selected for test shall be thoroughly heated to plus 212°F. When the starting process of each canister has been initiated, it shall conform to 3.5.1. The starting process shall allow the continuous production of breathable gas as specified in 3.4.1.

4.6.3.3 Purity. The gas samples collected in performing tests specified in 4.6.3.1 and 4.6.3.2 shall be tested for purity as specified in 3.4.1. Carbon monoxide shall not exceed 0.02 percent by volume at any one time, as specified in 3.5.1.

4.6.4 Canister sparking. To determine conformance to 3.5.2, the canister breathable gas starting process shall not produce a spark. Samples shall be tested in a darkened area to allow any sparking to be observed during the activation process. No more than one sample shall be tested at a time.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

4.6.5 Salt spray test. To determine the conformance to 3.5.3, sample canisters shall be exposed to a continuous 5 percent salt water spray, or equivalent, for 96 hours, rinsed with fresh water, and wiped dry. Failure of the breathable gas starting process to operate after this conditioning shall be cause for rejection of the lot.

4.7 Packaging inspection. Sample packages and packs, and the inspection of the preservation-packaging, packing and marking for shipment and storage shall be in accordance with the requirements of section 5 and the documents specified therein.

5. PACKAGING

(The packaging requirements specified herein apply only for direct Government acquisition.)

5.1 Preservation-packaging. Preservation-packaging shall be level A or C as specified (see 6.2.1).

5.1.1 Level A. Canisters shall be cleaned, dried, and unit-protected in accordance with the requirements of method III of MIL-P-116, and as follows: Six canisters, placed in a vertical attitude with necks upward, shall be contained in a fiberboard box meeting the requirements of type CF, grade V3C of PPP-B-636, with style at the contractor's option. Each box shall be furnished with interior fiberboard partitions providing each canister with a separate cell. Canisters' necks shall be provided additional protection by use of fiberboard pads to prevent damage during handling, shipment, and storage. Box closure, waterproofing, and reinforcing shall be in accordance with the appendix to the box specification, method V of PPP-B-636.

5.1.2 Level C. Six canisters, cleaned and dried, shall be packaged in a single container in a manner which will afford protection against physical and mechanical damage during shipment and handling from the supply source to the first receiving activity for immediate use.

5.2 Packing. Packing shall be level B or C as specified (see 6.2.1).

5.2.1 Level B. Canisters packaged as specified shall be packed in containers conforming to the following specifications at the contractor's option.

Specification	Container	Type or class
PPP-B-636	Fiberboard	Weather-resistant
PPP-B-640	Fiberboard-corrugated, triple wall	Class 2

Fiberboard boxes shall be closed, waterproofed, and reinforced in accordance with the appendix to the applicable box specification, except that method V of the appendix shall apply for boxes conforming to PPP-B-636. Unit containers conforming to class weather-resistant of PPP-B-636, when closed, waterproofed, and reinforced as specified herein, and used as a shipping container, need not be overpacked.

Source: https://assist.dla.mil -- Downloaded: 202640-18T10:27Z
Check the source to verify that this is the current version before use.

5.2.2 Level C. Canisters packaged as specified shall be packed in containers conforming to the following specifications at the contractors' option.

Specification	Container	Type or class
PPP-B-636	Fiberboard	Class - domestic
PPP-B-640	Fiberboard-corrugated, triple wall	Class 1

Fiberboard boxes shall be closed, in accordance with the appendix to the applicable box specification. Unit containers conforming to class weather resistant or domestic of PPP-B-636, closed, waterproofed, and reinforced, when required, and used as shipping containers, need not be overpacked.

5.3 Palletization. When specified (see 6.2.1), unit packages or shipping containers, as applicable, shall be palletized for shipment in accordance with MIL-STD-147.

5.4 Marking. In addition to any special marking required (see 3.3 and 6.2.1), canisters, unit packages, shipping containers, and palletized loads shall be marked for shipment in accordance with the requirements of MIL-STD-129.

6. NOTES

6.1 Intended use. The canister covered by this specification is intended to be used with the type A-4 oxygen breathing apparatus to supply breathable gas for use in shipboard fire fighting and rescue operations.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) When first article is required (see 3.1).
(c) Level A or C preservation-packaging as specified (see 5.1).
(d) Packing level required (see 5.2).
(e) Palletization required (see 5.3).
(f) Marking required (see 5.4).

6.2.2 Data requirements. When this specification is used in an acquisition and data are required to be delivered, the data requirements identified below shall be developed as specified by an approved Data Item Description (DD Form 1664) and delivered in accordance with the approved Contract Data Requirements List (CDRL), incorporated into the contract. When the provisions of DoD FAR Supplement, Part 27, Sub-Part 27.410-6 (DD Form 1423) are invoked and the DD Form 1423 is not used, the data specified below shall be delivered by the contractor in accordance with the contract or purchase order requirements. Deliverable data required by this specification are cited in the following paragraph.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

Paragraph no.	Data requirement title	Applicable DID no.	Option
4.4	Reports, test	DI-T-2072	----

(Data item descriptions related to this specification, and identified in section 6 will be approved and listed as such in DoD 5010.12-L., Vol. I, AMSDL. Copies of data item descriptions required by the contractors in connection with specific acquisition functions should be obtained from the Naval Publications and Forms Center or as directed by the contracting officer.)

6.2.2.1 The data requirements of 6.2.2 and any task in sections 3, 4, or 5 of this specification required to be performed to meet a data requirement may be waived by the contracting/acquisition activity upon certification by the offeror that identical data were submitted by the offeror and accepted by the Government under a previous contract for identical item acquired to this specification. This does not apply to specific data which may be required for each contract regardless of whether an identical item has been supplied previously (for example, test reports).

6.3 First article. When first article inspection is required, the canisters should be a first article sample or it may be a standard production item from the contractor's current inventory. The first article should consist of 100 units. The contracting officer should include specific instructions in acquisition documents regarding arrangements for examinations, approval of first article test results and disposition of first articles. Invitations for bids should provide that the Government reserves the right to waive the requirement for samples for first article inspection to those bidders offering a product which has been previously acquired or tested by the Government, and that bidders offering such products, who wish to rely on such production or test, must furnish evidence with the bid that prior Government approval is presently appropriate for the pending contract.

6.4 Material safety data sheets. Contracting officers will identify those activities requiring copies of completed Material Safety Data Sheets prepared in accordance with FED-STD-313. The pertinent Government mailing addresses for submission of data are listed in appendix B of FED-STD-313.

6.5 Subject term (key word) listing.

Breathing apparatus
Oxygen equipment

6.6 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue due to the extensiveness of the changes.

Preparing activity:
Navy - SH
(Project 4240-0515)

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEPARTMENT OF THE NAVY

COMMANDER
NAVAL SEA SYSTEMS COMMAND (SEA 55Z3)
DEPARTMENT OF THE NAVY
WASHINGTON, DC 20362-5101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 12503 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE NAVY

COMMANDER
NAVAL SEA SYSTEMS COMMAND (SEA 55Z3)
DEPARTMENT OF THE NAVY
WASHINGTON, DC 20362-5101

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T10:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-C-17671E	CANISTER, OXYGEN BREATHING APPARATUS, FOR TYPE A-4

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-18T10:27Z
Check the source to verify that this is the current version before use.