GG-B-00511a(DSA-DM)
August 19, 1969
Interim Revision of
Fed. Spec. GG-B-511
July 31, 1952

INTERIM FEDERAL SPECIFICATION

BURNER OUTFIT, GAS, DENTAL LABORATORY

This Interim Federal Specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this Interim Federal Specification as a valid exception to Federal Specification GG-B-511.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers an acetylene, gas-operated burner outfit capable of producing a flame suitable for soldering and melting procedures involving dental gold and silver alloys.

1.2 Classification.

1.2.1 Burner outfit, complete. The burner outfit shall be furnished complete with all component parts (see 3.3 and 6.1).

1.2.2 Replacement parts. Replacement parts for the burner outfit shall be of the following designated types (see 3.4.2 and 6.1).

Type I. Blowpipe, soldering and melting.
Type II. Burner, bunsen, with spider.
Type III. Valve assembly, needle, bench.
Type IV. Regulator, pressure, acetylene.
Type V. Tube assembly, brass.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents, of the issue in effect on date of invitations for bid or requests for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical, General Specifications.
QQ-A-601	Aluminum, Alloy Sand Castings.
QQ-B-626	Brass Leaded & Non-Leaded, Rod, Shaped, Forgings & Flat Products.
QQ-S-698	Steel, Sheet & Strip, Low Carbon.
QQ-S-763	Steel, Bars, Shapes & Forgings, Corrosion-Resistant.
ZZ-T-831	Tubing, Rubber, Synthetic-Rubber, and Plastic.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Fort Worth, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

regulator or bench valve. Tubing shall be furnished in 8-foot lengths, neatly coiled to a diameter of approximately 3½ inches.

3.2.6 *Rubber tubing.*—See 3.1.1 and 3.2.1.5.

3.3 Finish.—Unless otherwise specified, all brass parts shall be chrome plated, in conformance with class 4, composition N of Federal Specification GG–I–526.

3.4 Identification marking.—Blowpipe, regulator, Bunsen burner, and bench valve shall be stamped with the name or registered trade-mark of the manufacturer. Blowpipe tips shall be stamped with the number on each tip.

3.5 Service data.—The contractor shall enclose with each unit two copies of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of spare parts. Wiring diagrams shall be included for electrical items or components. Each part, other than common fasteners, etc., shall be identified by the manufacturer's part number. In addition, the contractor shall mail one copy of the service data to Specifications Branch, Armed Services Medical Procurement Agency, 84 Sands Street, Brooklyn 1, N. Y., and one copy to Medical Technical Maintenance Division, Armed Services Medical Procurement Agency, St. Louis Medical Depot, 12th and Spruce Streets, St. Louis 2, Mo.

3.6 Workmanship.—Workmanship shall be first class throughout. Blowpipe outfits shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.—Unless otherwise specified, at least two acetylene blowpipe outfits shall be taken at random from each 100, or fraction thereof, delivered.

4.2 Inspection.—Samples shall be inspected for materials, workmanship, dimensions, component parts, finish, identification marking, packaging, packing, and marking.

4.3 Tests.—Tests shall be conducted in accordance with the applicable specifications referenced in 2.1.

5. PREPARATION FOR DELIVERY

5.1 Packaging.—Unless otherwise specified, each blowpipe outfit shall be packaged in accordance with accepted commercial practice.

5.2 Packing.—The subject commodity shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate, and to afford maximum protection from normal hazards of transportation.

5.3 Marking.

5.3.1 *Packages.*—Unless otherwise specified, packages shall be marked with the name of the commodity, the quantity contained therein, and the names of the manufacturer and contractor.

5.3.2 *Shipping containers.*—Unless otherwise specified, shipping containers shall be marked with the name of the commodity as defined by the contract or order under which the shipment is made, the quantity contained therein, the name of the contractor, the number of the contract or order, the name and address of the consignee, and the gross weight.

6. NOTES

6.1 Purchasers should exercise any desired option offered herein. (See 3.3, 4.1, 5.1, 5.3, 6.2, and 7.2.6.)

6.2 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the current version before use.

that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.3 This specification covers the following items appearing in the Armed Services Catalog of Medical Materiel:

5–022–000—Blowpipe Outfit, Acetylene.
5–022–200—Blowpipe, Acetylene.
5–022–400—Burner, Bunsen, Acetylene.
5–022–600—Regulator, Pressure, Acetylene.
5–022–800—Tubing, Brass, Acetylene.
5–023–000—Valve, Bench, 3-Way.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.5 Transportation description.

6.5.1 *Transportation description applicable to these items is:*

Outfits, oxy—acetylene cutting or welding.
In boxes.
Carload weight 12,000 pounds subject to Rule 34, Consolidated Freight Classification.
Truckload weight 12,000 pounds.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and standards, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

7.2 Army, Navy, and Air Force.

7.2.1 *Applicable specifications and standard.*

Federal Specifications:
NN–B–601—Boxes, Wood-Cleated-Plywood, for Domestic Shipment.
NN–B–621—Boxes, Wood, Nailed and Lock-Corner.
NN–B–631—Boxes, Wood, Wirebound (for Domestic Shipment).
LLL–B–631—Boxes; Fiber, Corrugated (for Domestic Shipment).
LLL–B–636—Boxes, Fiber, Solid (for Domestic Shipment).

Military Specifications:
JAN–P–105—Packaging and Packing for Overseas Shipment—Boxes, Wood, Cleated, Plywood.
JAN–P–106—Packaging and Packing for Overseas Shipment—Boxes, Wood, Nailed.
MIL–B–107—Boxes, Wood, Wirebound (Overseas Type).
JAN–P–125—Packaging and Packing for Overseas Shipment—Barrier-Materials, Waterproof, Flexible.
MIL–A–140—Adhesive, Water-Resistant, Waterproof Barrier-Material.

Navy Department Specification:
General Specifications for Inspection of Material.[1]

Military Standard:
MIL–STD–129—Marking of Shipments.

[1] Applicable to Navy purchases only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring agency or as directed by the contracting officer.)

7.2.2 *Packaging.*—One blowpipe outfit, as specified, constitutes one unit. Each unit shall be packaged in a solid fiberboard or double-faced corrugated fiberboard carton, of appropriate size and design, conforming to the requirements of Federal Specification LLL–B–631 or LLL–B–636. Carton shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Each carton shall be adequately secured to prevent accidental opening.

7.2.3 *Packing.*

7.2.3.1 *Domestic.*—Four unit cartons (4 units) shall be packed in a shipping container conforming to the requirements of Federal Specification NN–B–601, NN–B–621, or NN–B–631.

7.2.3.2 *Overseas.*—Four unit cartons (4 units) shall be packaged in a shipping container designed for type 2 load and conforming to the requirements of Military Specification JAN–P–105, JAN–P–106, or MIL–B–107. Closure and strapping shall conform to the requirements of the appendix to the applicable box specification. Each shipping container shall be lined with a sealed waterproof bag, or its equivalent, made from material conforming to the requirements of Military Specification JAN–P–125, for case liners; seams and closures shall be sealed with a continuous seam at least ¾ inch wide of water-resistant adhesive conforming to the requirements of Military Specification MIL–A–140.

7.2.3.3 *Packing variation.*—If the number of units in any shipment is not the exact multiple of the number of units specified to be contained in each shipping container, the units not constituting the nearest such multiple shall be packed in a shipping container of suitable size and design and so constructed as to insure safe delivery to the specified destination.

7.2.4 *Marking.*—Each unit carton and each shipping container shall be marked in accordance with Military Standard MIL–STD–129.

7.2.5 *Inspection procedures (Navy only).*—The general inspection procedures shall be in accordance with Navy Department General Specifications for Inspection of Material.

7.2.6 *Ordering data.*—Procurement documents should specify whether the subject commodity is to be packed for domestic or overseas shipment. (See 7.2.3.)

NAVY INTEREST: Md

This specification is a part of Section IV, Part 5, of the Federal Standard Stock Catalog.

For sale by the Superintendent of Documents, Washington 25, D. C. Price 5 cents.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.