GG-H-575b

MAY 14, 1957

SUPERSEDING
Int. Fed. Spec. GG–H–00575a (GSA–FSS)
September 6, 1956, and
Fed. Spec. GG–H–575
November 20, 1946

FEDERAL SPECIFICATION

HOLDER, NAPKIN, DENTAL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers napkin holders suitable for use in a dental office.

1.2 Classification.

1.2.1 Types.—Dental napkin holders covered by this specification shall be of the following types:

Type I.—Alligator-jaw clasp.

Type II.—Cross-action clasp.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 Specifications and standards.—The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

GG–I–526 — Instruments, Dental and Surgical; General Specifications.

QQ–W–465 — Wire, Steel (Carbon); Hard-Drawn, Spring and Zinc-Coated High-strength.

Federal Standards:

No. 102—Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL–P–12902 — Packaging of Surgical, Dental and Veterinary Instruments, Diagnostic Equipment, and Related Items.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

402174/57/No. 355

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS.

3.1 Material.

3.1.1 *Brass.*—All parts of the napkin holders, except the spring, shall be made of brass conforming to the requirements of Federal Specification GG–I–526 for class 4, composition N.

3.1.2 *Steel.*—The spring shall be of spring steel wire conforming to the requirements of Federal Specification QQ–W–465.

3.2 **Chain.**—Chain shall be of the ball-chain type, approximately 10.50 inches in length. Each ball shall be approximately 0.125 inch in diameter. The balls of the chain shall be linked together by "dumbbell-type" links, and the chain shall be extremely flexible.

3.3 **Finish.**—The napkin holders shall be finished in accordance with the requirements of Federal Specification GG–I–526.

3.4 **Type I.**—The dental napkin holder shall consist of a ball chain with an alligator-jaw-type clasp and sleeve securely attached at each end. Each clasp shall have hole in base for attachment of the ends of the chain.

3.4.1 *Clasps.* — Clasps shall be approximately 2 inches in length and 0.25 inch in width, tapering to 0.125 inch in width at the distal end. The clasps shall be serrated on the jaw edges. There shall be not less than 15 serrations to each jaw. The serrations shall mesh properly when closed, so as to hold the napkin securely. Clasps shall be made of brass as specified in 3.1.1, or of carbon steel, suitable for the purpose, in conformance with class 1 of Federal Specification GG–I–526.

3.4.2 *Sleeves.*—Sleeves shall be approximately 0.375 inch in diameter and 0.687 inch in length. There shall be two sleeves, one on each clasp, covering the joint of the clasp and the ball chain.

3.4.3 *Spring.*—Each clasp shall have one spring so wound as to give a spring tension of 3-inch-pounds. The spring steel wire shall be approximately 0.037 inch in diameter.

3.5 **Type II.** — The dental napkin holder shall consist of two cross-action-type clasps permanently connected by a ball chain.

3.5.1 *Clasps.* — Clasps shall be approximately 0.437 inch long and 0.187 inch wide. Each clasp shall have a hole in the base for attachment of the ends of the chain.

3.5.2 *Joints.*—The attachment of the chain to each clasp shall be made by passing the wire end of the chain through the hole in the base of the clasp and riveting the wire securely.

3.6 **Workmanship.**—Workmanship shall be first class in every respect. Napkin holders shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Sampling.**—Unless otherwise specified, samples shall be submitted in accordance with instructions furnished by the procuring activity. In addition, sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105.

4.2 **Inspection.**—Inspection may be made throughout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibility for faulty workmanship or materials which may be discovered at any time prior to final acceptance. Final inspection of the finished article shall be made either at point of production or at point of delivery designated in the contract or purchase order of procuring agency. In case of factory inspection, every

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 **Tests.**—Tests shall be conducted in accordance with the applicable specifications, as necessary, to determine compliance with this specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging and packing.

5.1.1 *Unit of issue.*—One holder, as specified, constitutes one unit.

5.1.2 *Packaging quantities.*—The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

TABLE I.—*Packaging quantities*

Federal Stock Number	Unit Package	Intermediate Package	Shipping Container
6520–536–0000–	1 Unit	12 Units	288 Units

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.* — Packaging and packing shall conform to Procedure Code No. 38 as specified in Military Specification MIL–P–12902.

5.1.4 *Level B.* — Packaging and packing shall conform to Procedure Code No. 22 as specified in Miiltary Specification MIL–P–12902.

5.1.5 *Level C.* — The subject commodity shall be packaged and packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 **Marking.**—Marking shall be as specified in Military Specification MIL–P–12902, except that the ASMPA stock number will not be required.

6. NOTES

6.1 Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 **Ordering data.** — Purchasers should specify type, class and size required, and should exercise any desired options offered herein. (See 1.2.1, 4.1, and 5.1).

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 This specification does not include all types, classes, grades, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

6.5 Transportation description.—Transportation description applicable to these items is:

Brass Articles, NOIBN, Flat.
Carload minimum weight 30,000 pounds.
Motor volume minimum weight 30,000 pounds.

6.6 This specification covers the following item appearing in the Armed Services Medical Stock List:

6520–536–0000—Holder, Napkin, Dental.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force—AF.

Copies of this specification may be purchased for 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.