GG-R-563b
December 9, 1957
SUPERSEDING
Int. Fed. Spec. GG–R–00563a(GSA–FSS)
October 15, 1956 and
Fed. Spec. GG–R–563
January 23, 1946

FEDERAL SPECIFICATION

ROLL, ABSORBENT COTTON AND ROLL, CELLULOSE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers dental absorbent rolls generally used in dental procedures.

1.2 Classification.

1.2.1 *Types and sizes.* — Absorbent rolls shall be supplied in the following types and sizes, as specified in the invitation for bids:

Type I.—Cotton rolls, starched or braided surfaces.
- Sizes.—3/8-inch diameter in 6-inch lengths.
- Sizes.—1/2-inch diameter in 1 1/2-inch lengths.
- Sizes.—1/2-inch diameter in 6-inch lengths.

Type II.—Cotton filler, knitted cover.
- Size.—3/8-inch diameter in 22 1/2 foot lengths.

Type III.—Cellulose filler, knitted cover.
- Sizes.—7/16-inch diameter in 22 1/2 foot lengths.
- Sizes.—7/16-inch diameter in 45-foot lengths.

Type IV.—Cellulose filler, gauze cover.
- Sizes.—7/16-inch diameter in 22 1/2 foot lengths.
- Sizes.—7/16-inch diameter in 45-foot lengths.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND OTHER PUBLICATIONS

2.1 Specifications and standards.—The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

L–C–166—Cellulose, Absorbent, Surgical.

CCC–G–101—Gauze, Absorbent.

JJJ–C–561 — Cotton; Sterile and Nonsterile.

PPP–B–566 — Boxes, Folding, Paperboard.

PPP–B–585—Boxes; Wood, Wirebound.

PPP–B–601—Boxes, Wood, Cleated-Plywood.

PPP–B–621—Boxes, Wood, Nailed and Lock-Corner.

PPP–B–636—Boxes, Fiber.

PPP–B–676—Boxes, Set-up, Paperboard.

Federal Standard:

Fed. Std. No. 102 — Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL–B–10377 — Box, Wood, Cleated, Veneer, Paper Overlaid.

MIL–L–10547 — Liners, Case, Waterproof.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications.—The following publication, of the issue in effect on date of invitation for bids, forms a part of this specification:

United State Pharmacopoeia Convention Publication:

Pharmacopoeia of the United States.

(The Pharmacopoeia of the United States is published by the Mack Printing Company, Easton, Pa.)

3. REQUIREMENTS

3.1 Material.

3.1.1 *Absorbent cellulose.*—Cellulose shall conform to the requirements of Federal Specification L–C–166.

3.1.1.1 *Gauze.*—The gauze shall conform to the requirements for type I in Federal Specification CCC–G–101.

3.1.2 *Absorbent cotton.* — The absorbent cotton shall conform to the requirements for grade B in Federal Specification JJJ–C–561.

3.2 Absorbent rolls shall be readily adaptable to mouth contour and shall have a high degree of absorbency.

3.3 **Sterility.** — Absorbent rolls shall be sterilized. Sterilization shall be accomplished by high pressure steam. Rolls shall be thoroughly dried by vacuum. Chemical sterilization is not permissible.

3.3.1 *Certificate of sterility.*—Unless otherwise specified, a certificate of sterility, listing all lot or control numbers, stating that samples from each sterilizer charge have been tested and found to be sterile shall be forwarded by the contractor to the procuring activity (see 6.3).

3.4 Construction.

3.4.1 *Type I.* — Absorbent cotton shall be formed in rolls 1½ and 6 inches in length, of the specified diameter. The rolls shall be uniform in firmness and consistency, shall be non-collapsible, and shall not adhere to tissue. The outer surface shall be either starched or braided with a soft, synthetic yarn.

3.4.1.1 *Weight.* — The minimum weight shall be as follows:

⅜-inch diameter, 70 grams per 50 rolls, 6-inch lengths.

½-inch diameter, 25 grams per 50 rolls, 1½-inch lengths.

½-inch diameter, 100 grams per 50 rolls, 6-inch lengths.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

3.4.2 *Type II.*—Absorbent rolls shall consist of a filler made of absorbent cotton, running approximately 9 grams per yard, formed in a roll of the specified diameter; and an outer covering circular-knit, seamless, with a plain stitch, having a loop count of 32 loops in the circumference. Twelve single carded knitting twist bleached yarns shall be used in the knitting.

3.4.2.1 *Dispensing package.*—One continuous length of 22½ feet of absorbent roll shall be coiled in a blue-surfaced paper bag with an opening at the top for ready withdrawal of the free end. This bag shall be enclosed in an outer protective bag of white paper, closed by stapling through a double fold or by sealing.

3.4.2.2 *Weight.* — The minimum weight shall be 80 grams per 22½ foot roll.

3.4.3 *Type III.*—Absorbent rolls shall consist of a filler made of absorbent cellulose of single ply, approximately 26 inches wide, formed in a roll of the specified diameter, with an outer covering circular-knit, seamless, with a plain stitch, having a loop count of 32 loops in the circumference. Twelve single carded knitting twist bleached yarn shall be used in the knitting.

3.4.3.1 *Dispensing package.*—One continuous length of 22½ feet of absorbent roll, or one continuous length of 45 feet, shall be packaged as described in 3.4.2.1.

3.4.3.2 *Weight.*—The weight of the material per running yard shall be as follows:

Absorbent cellulose	*Knit casing*	*Absorbent roll*
115-145 grains	35-55 grains	150-190 grains

3.4.4 *Type IV.*—Absorbent rolls shall consist of a filler made of absorbent cellulose of approximately 14 ply, 7/16-inch wide, with an outer covering of gauze 2 inches wide. The gauze shall be sewed to the cellulose. The gauze seam shall have its raw edges on the inside and shall be stitched by machine with No. 40-3 cord (left twist) using a chain stitch (one needle and one looper thread).

3.4.4.1 *Dispensing package.*—One continuous length of 22½ feet of absorbent roll, or one continuous length of 45 feet, shall be packaged as described in 3.4.2.1.

3.4.4.2 *Weight.*—The weight of the material per running yard shall be as follows:

Absorbent cellulose	*Gauze 44-36*	*Absorbent roll*
170-200 grains	36-40	200-240 grains

3.5 **Workmanship.** — The workmanship shall be first class throughout. Absorbent rolls shall be free from all defects which affect their appearance or impair their serviceability.

4. SAMPLING, INSPECTION AND TEST PROCEDURES

4.1 **Sampling.**—Unless otherwise specified, samples shall be submitted in accordance with instructions furnished by the procuring activity (see 6.2). In addition, sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105.

4.2 **Inspection.**—Inspection may be made throughout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibility for faulty workmanship or materials which may be discovered at any time prior to final acceptance. Final inspection of the finished article shall be made either at point of production or at point of delivery, as specified (see 6.2). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the currentversion before use.

4.3 Tests.

4.3.1 *Component materials.* — Cellulose, gauze, and cotton shall be tested as prescribed in the applicable specifications.

4.3.2 *Sterility.*—All sterility tests shall be conducted in accordance with the latest revision of the U. S. Pharmacopoeia.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 *Unit of issue.*

5.1.1.1 *Stock No. 6510–518–1900.* — Five hundred (500) rolls, as specified, constitute one unit of issue.

5.1.1.2 *Stock No. 6510–518–2050.* — One thousand (1000) rolls, as specified, constitute one unit of issue.

5.1.2 *Packaging quantities.* — The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

TABLE I.—*Packaging quantities*

Federal stock number	Unit Package	Intermediate package	Shipping container
6510–518–1900 – or 6510–518–2050 –	1 unit	6 units	24 units

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*

5.1.3.1.1 *Stock No. 6510–518–1900.*—Fifty rolls shall be contained in a suitable paper bag, or wrapped in suitable paper. Bag or wrapper shall be secured. Ten such packages (1 unit), shall be packaged in a carton of appropriate size constructed in accordance with Federal Specifications PPP–B–566 or PPP–B–676. Each carton shall be adequately secured to prevent accidental opening.

5.1.3.1.2 *Stock No. 6510–518–2050.* — One hundred twenty-five (125) rolls shall be contained in a suitable paper bag, or wrapped in suitable paper. Bag or wrapper shall be secured. Eight (8) such packages (1 unit) shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Each carton shall be adequately secured to prevent accidental opening.

5.1.3.2 *Intermediate package.*—Intermediate package shall be a carton of appropriate size constructed in accordance with Federal Specification PPP–B–636, type I, class 1. Each carton shall be adequately secured to prevent accidental opening.

5.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 3 use; PPP–B–601, table II, using type III, class 1 plywood; PPP–B–621, table III; or Military Specification MIL–B–10377, table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

5.2.1.2 *Case liner.*—Each level A box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.* — Strapping, when required, shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall conform to Federal Specification PPP–B–601, table I, using type III, class 1 plywood, or shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use; PPP–B–621, table I; or Military Specification MIL–B–10377, table I.

5.2.3 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size, and kind, commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 *Interior packages.*—Each unit package and intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.2 *Shipping container.*—Shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Federal Standard 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 Ordering data. — Purchasers should specify type and size required and should exercise any desired options offered herein. (See 1.2.1, 3.3.1, 4.1, 5.1 and 5.2).

6.3 For Military Medical Supply Agency procurement. — Certificate of sterility, required in 3.3.1, shall be submitted to Head, Quality Control Division, Technical Department, Military Medical Supply Agency, 84 Sands Street, Brooklyn 1, New York.

6.4 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.5 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.6 Transportation description.—Transportation description applicable to this item is:

Dental goods, not otherwise indexed by name.

Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Motor volume minimum weight 12,000 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

6.7 This specification covers the following items appearing in the Armed Services Medical Stock List:

6510–518–1900—Roll, Absorbent Cotton, ⅜ by 6 inches, 500s.

6510–518–2050—Roll, Absorbent Cotton, ½ by 1½ inches, 1000s.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.