GG-R-571b
MAY 24, 1957
SUPERSEDING
Int. Fed. Spec. GG-R-00571a (GSA-FSS)
March 14, 1957, and
Fed. Spec. GG-R-571
July 26, 1945

FEDERAL SPECIFICATION

ROLLER, BANDAGE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. — This specification covers a hand-operated device for rolling strip bandages of different widths into rolls.

1.2 Classification.

1.2.1 *Type, grade, and sizes.*—This specification covers one type and grade of bandage rollers, which shall be furnished in the following two sizes, as specified in the invitation for bids:

Size 1 — Bandage rollers, for 4½-inch maximum width bandages (single clamp).

Size 2 — Bandage rollers, for 6- to 8-inch bandages (double clamp).

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

GG-I-526 — Instruments, Dental and Surgical; General Specifications.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

3. REQUIREMENTS

3.1 Material. — Shall be as hereinafter specified.

3.2 Description. — Bandage rollers shall consist of two side frames, clamp or clamps, a bandage roller, three guide rods and a removable spindle with off-set handle.

3.3 Construction. — Bandage rollers shall be sturdily constructed in all respects for heavy duty. Frames shall be of conventional

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

design with table clamp or clamps which will permit an opening of at least 1 inch for secure attachments. Three guide rods shall be provided on the front of the bandage rollers. The spindle shall be removable and made of approximately $\frac{5}{32}$-inch square rod having an off-set hardwood handle on the distal end to facilitate manipulation. Frame, spindle, and guide rods shall be of cold-rolled steel or cast brass.

3.3.1 *Screws.*

3.3.1.1 *Screws for attaching guide rods.*—Shall be of brass, except when frame is cast brass when screws may be of brass or steel.

3.3.1.2 *Clamp screws.*—Shall be of brass, or of malleable iron or forged steel hardened and tempered to provide strength and wearing qualities. Each screw shall have a swivel button securely attached to the inner end.

3.3.1.3 *Screw heads.* — The heads of all screws shall be winged, corrugated or milled for ease in handling.

3.3.2 *Clamps.*

3.3.2.1 Size 1 bandage roller shall have a single table clamp as an integral part of the right frame.

3.3.2.2 Size 2 bandage roller shall have two table clamps formed as integral parts of side frames.

3.4 **Dimensions.**

3.4.1 Size 1 bandage roller shall be designed to roll bandages of various widths up to 4½ inches.

3.4.2 Size 2 bandage roller shall be designed to roll bandages of widths from 6 to 8 inches.

3.5 **Finish.**—All metal surfaces shall have a composite plating of chromium over nickel or chromium over nickel and copper, complying with the requirements of class I plated instruments of Federal Specification GG-I-526.

3.6 **Workmanship.** — Shall be first class throughout. The complete bandage roller shall be free from defects which detract from its appearance or which may impair its serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.—Unless otherwise specified, at least two bandage rollers, of like size, shall be taken at random from each delivery for inspection and tests.

4.2 Inspection.—Samples shall be carefully inspected for defects in material, workmanship, dimensions and other applicable requirements.

4.3 Tests.

4.3.1 *Plating.*—Shall be tested in conformance with the applicable requirements of Federal Specification GG-I-526.

5. PREPARATION FOR DELIVERY

5.1 **Packaging.**—Unless otherwise specified, commercial packages are acceptable, (level C.).

5.2 **Packing.**—Unless otherwise specified, the subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate to the point of delivery called for in the contract or purchase order.

5.3 **Marking.**

5.3.1 *Packages.*—Each unit container shall be plainly labeled with the name of the item, size, quantity contained, and the name of the manufacturer or contractor.

5.3.2 *Shipping containers.*—Unless otherwise specified, the shipping containers shall be marked with the name of the item, the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

size, and quantity contained therein, as defined by the contract or purchase order under which the shipment is made, the name of the contractor, the number of the contract or order, and the gross weight.

6. NOTES

6.1 Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 Purchasers should exercise any desired options offered herein and should specify size. (See 1.2.1, 4.1, 5.1, 5.2, and 5.3.2.)

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 Federal specifications do not include all types, classes, grades, sizes, etc., of the commodities indicated by the titles of the specifications, or which are commercially available, but is intended to cover the types, etc., which are suitable for Federal Government purchases.

6.5 Transportation description.—Transportation description applicable to this item is:

Tools, NOIBN (by analogy)
Carload minimum weight 30,000 pounds.

Motor Volume minimum weight 30,000 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

★ U. S. GOVERNMENT PRINTING OFFICE: 1967

Copies of this specification may be purchased for 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

5090 M

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

NOTICE OF CANCELLATION

GG-R-571B
NOTICE 1
31 JANUARY 1996

FEDERAL SPECIFICATION

ROLLER, BANDAGE

GG-R-571B, dated 24 MAY 1957, is hereby cancelled without replacement.

Preparing Activity

DOD-MB

Project Number: 6510-0875

AMSC/NA

FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.