GG-S-530A
November 17, 1975
SUPERSEDING
Int. Fed. Spec. GG-S-00530
April 16, 1964

FEDERAL SPECIFICATION

SNAKE BITE KIT (SUCTION)

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1 Scope. This specification covers a small snake bite first aid kit of the suction type.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issues in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

Federal Specifications:

PPP-B-566 - Boxes, Folding Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-Up.

Federal Standards:

Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issue in effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

National Motor Freight Traffic Association, Inc., Agent:

National Motor Freight Classification

(Application for copies should be addressed to the American Trucking Associations, Inc., Tariff Order Section, 1616 P Street, N.W., Washington, DC 20036.)

Uniform Classification Committee, Agent:

Uniform Freight Classification.

(Application for copies should be addressed to the Uniform Classification Committee, Room 1106, 222 South Riverside Plaza., Chicago, IL 60606.)

3. REQUIREMENTS

3.1 Design. Each kit is self-packaging and consists of three suction cups, one vial of antiseptic, one scapel, one lymph constrictor, and one instruction sheet.

3.2 Components.

3.2.1 Suction cups. The two large suction cups shall form the outer container for all the components. The small cup shall be designed for application to fingers or other narrow gripping areas and shall fit snugly into the large cups. The suction cups shall be molded of pliable rubber, or plastic, compounded to provide long service and of suitable resiliency to provide air-tight contact with the affected area in a manner to produce positive suction.

3.2.2 Vial. The vial of antiseptic shall fit into the small suction cup. The vial shall contain antiseptic of a 0.1% solution of thimerosal or iodine to make sterile the complete area of the bite and to sufficiently moisten the area to provide an effective air-tight seal upon application of the suction cup.

3.2.3 Scapel. The scapel, contained in the small suction cup, shall be ready for immediate use. The cutting edge shall be protected by a short length of plastic tubing that also serves as a handle.

3.2.4 Lymph constrictor. The lymph constrictor shall be contained within one large suction cup. The lymph constrictor shall be a braided cord with a minimum of 24 inches in length, and of such strength as to stop the flow of blood in the veins in the area affected when applied 1-1/2 inches above the wound and in accordance with the directions for use.

3.3 Dimensions. The overall dimensions of the assembled kit shall not exceed 3-1/2 inches in length and shall be 1 inch plus or minus 1/8 inch in diameter.

3.4 Weight. The weight of the kit shall not exceed 2 ounces.

3.5 Instruction sheet. An instruction sheet, contained within the small suction cup, shall be included in each kit. This sheet may also be used as protective packaging for the vial of antiseptic and knife blade.

3.6 Workmanship. The kit shall be free from defects which affect its appearance or which may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsiblity for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

4.2. Sampling for inspection and tests. Sampling for inspection and test shall be in accordance with the provisions of MIL-STD-105. For purposes of sampling and testing, a lot shall consist of kits covered by this specification submitted for inspection and delivery at one time.

4.2.1 Inspection of component and materials. In accordance with 4.1 the supplier shall be responsible for insuring that the quality assurance provisions for components or materials shall be in accordance with this specification and with reference subsidiary specifications to the extent applicable. In the event of conflict, this specification shall govern.

4.3 Inspection of the end item.

4.3.1 Examination of the end item. Examination of the end item shall be in accordance with the classification of defects and acceptable quality levels (AQL's) specified in table I. The Government reserves the right to require examination for any defect not included herein and to classify such defects in accordance with the definitions contained in MIL-STD-105.

TABLE I. Classification of defects

Examine	Defects
Major	AQL 1.0 percent defective
101	Part missing
102	Part damaged effecting use
103	Instructions missing
Minor	AQL 4.0 percent defective
201	Length (over 3-1/2 inches)
202	Diameter (not 1 inch plus or minus 1/8 inch)
203	Weight (over 2 ounces)

4.4 Examination of preparation for delivery. An examination shall be made to determine compliance with the packaging, packing and marking requirements of section 5. Defects shall be scored as specified in table II. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for delivery at one time. The inspection level shall be S-2 with an AQL of 4.0 expressed in terms of percent defective.

TABLE II. Examination of preparation for delivery

Examine	Defects
Container	Not as specified
Contents	Not as specified
Marking	Omitted; incorrect; illegible; improper size, location, sequence or method of application.
Material	Component missing or damaged
Workmanship	Bulging or distortion of containers

4.4.1 Examination of closure, waterproofing, and banding of containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure, waterproofing and banding shall be in accordance with the appendix of that specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, B, or C as specified (6.2).

5.1.1 Level A and B. Unit container. Twelve (12) snake bite kits, each assembled as specified in 3.2.1, shall be packaged in a box conforming to PPP-B-566 or PPP-B-676, variety 1. The type, style, class, and process is at the option of the supplier. The box shall be secured to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. The snake bite kits, each assembled as specified in 3.2.1 shall be packaged to afford adequate protection against damage during shipment from the supply source to the first receiving agency.

5.2 Packing. Packing shall be level A, B, or C as specified (6.2).

5.2.1 Level A. Twelve unit packages shall be packed in a box conforming to PPP-B-636, class weather resistant. The box shall be closed, waterproofed and banded in accordance with the appendix to the box specification.

5.2.2 Level B. Forty unit packages shall be packed in a box conforming to PPP-B-636, class domestic. The box shall be closed in accordance with the appendix to the box specification.

5.2.3 Level C. Snake bite kits in quantities as specified (see 6.2), shall be packed in a container complying with the Uniform Freight or National Motor Freight Classification, as applicable. The containers shall be packed to assure carrier acceptance and safe arrival at destination.

5.3 Marking. The unit and shipping containers shall be marked in accordance with Fed. Std. No. 123 or MIL-STD-129, as applicable (see 6.2).

6. NOTES

6.1 Intended use. The product covered in this specification is intended as an emergency first aid treatment for snake bite when conditions are such that there is a delay in reaching a physician.

6.2 Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this specification.
(b) Level of packaging and packing (5.1 and 5.2).
(c) Quantity to be packaged (5.1.1).
(d) Marking document required (see 5.3).

6.3 Contractor's inspection. The contractor should be required to assure that the product conforms with the specification prior to submission to the Government for final inspection.

6.4 Destination inspection. When the finished product has been inspected and passed at point other than destination, the contract should require that the product be inspected as destination for condition and quantity only.

CIVIL AGENCY COORDINATING ACTIVITY

VA-DMS

Preparing activity:

GSA-FSS

U. S. GOVERNMENT PRINTING OFFICE : 1976 - 210-814/1109

Orders for this publication are to be placed with General Services Administration, acting as agent for the Superintendent of Documents. See Section 2 of this specification to obtain extra copies and other documents referenced herein. Price 25 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.