GG-S-00571a
23 May 1967
Interim Revision of
Fed. Spec. GG-S-571
February 26, 1954

INTERIM FEDERAL SPECIFICATION

SOUNDS, URETHRAL (Curved, LeFort)

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-S-571.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers curved urethral sounds of the LeFort design capable of union with urethral filiform catheters and suitable for use in urethral dilation.

1.2 Classification. Sounds shall be of the following sizes; as specified (see 6.2):

Size 12 - 12 French
Size 14 - 14 French
Size 16 - 16 French
Size 18 - 18 French
Size 20 - 20 French

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawings. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-F-340	Filiforms, Urethral, Woven (Braided)
GG-I-526	Instruments, Dental and Surgical; General Specifications
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.
MIL-P-116	Preservation, Methods of

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

23025	Sound, Urethral, Curved, LeFort

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The sounds shall be fabricated of class 3, type 302, 303, or 304, stainless steel conforming to the requirements of GG-I-526.

3.2 Style, design and dimensions. The sound shall be curved, of the LeFort design, suitable for dilation of the urethra. Sounds shall be so constructed that when fitted with a urethral filiform conforming to Federal Specification GG-F-340, the filiform shall be capable of butting against the end of the sound. Style, design, and dimensions shall be in accordance with DPSC Drawing No. 23025, revised 22 March 1967.

3.3 Finish. The sounds shall have a mirror finish, conforming to the requirements of GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

3.4 Hardness. The hardness for the material specified in 3.1 shall be Rockwell B-75 minimum conforming to the requirements of GG-I-526.

3.5 Identification marking. Sounds shall be permanently and legibly marked with the French scale size; "stainless"; the name of the country of origin (if other than U.S.); and the name or registered trademark of the contractor.

3.6 Workmanship. Workmanship shall be first class throughout. Sounds shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboraotry appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specificed test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the current version before use.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The Acceptable Quality Levels (AQL's) shall be as follows:

Type defect	Inspection level	AQL (percent defective)
Major defects	II	1.0
Minor defects	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table I.

Table I. Sampling Criteria.

Component	Characteristic	Requirement	Test procedure	Unit of product	Insp. level	AQL (percent defective)
Sound	Hardness	3.4 and GG-I-526	GG-I-526	1 sound	S-2	*
	Boil test	3.1 and GG-I-526	GG-I-526	1 sound	S-2	*

* Acceptance number shall be zero for all sample sizes.

4.2.3 Raw materials. Four ounces of each lot of each metal used in construction of the sounds shall be selected for chemical analysis.

4.2.4 Test samples. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The sounds shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in table II. Any sounds in the sample containing one or more defects shall be rejected, and if the number of defective sounds in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects in accordance with MIL-STD-105.

Categories	Defects
Critical	None defined
Major	
101	Sound not free from cracks, laps and folds.
102	Sound not free from burrs and sharp edges.
103	Surfaces not smooth and free from pits, corrosion, roughness and deformation or uspecified bends.
104	Soldered or brazed joint not clean and free from excess flux and discoloration (applicable for two piece construction only).
105	Dimension A not within tolerance.
106	.140 $\pm$ .000 .005 dimension (see DPSC Drawing No. 23025, Detail 1) oversize.
107	Thread not American National Standard for Class 2 fit as follows: Threads per inch 56; Major diameter .073 - .004; Pitch diameter .0614 - .002; Minor diameter .0511 maximum.
108	Thread not uniform in dimensions (within tolerances) over its entire length.
109	Undercut shown in detail 1 of DPSC drawing No. 23025 not present or suitable to assure that metal fittings of Filiform will butt against end of sound.
110	Labels or printed marking on unit container not correct (particularly Stock No. and item identification).
Minor	
201	Dimensions (those not listed under major defects) not within tolerances shown in DPSC Drawing No. 23025.
202	Identification marking not present, legible or permanent.
203	Finish not mirror.
204	Finish not free of coarse grind marks or excessively scratched surface areas.
205	Labels or marking on unit container not legible and free of defacing marks, scratches or tears.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each sound, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	216 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the currentversion before use.

1.3.4 Unit package. Each unit shall be packaged in accordance th MIL-P-116, method 1A-15.

1.3.5 Intermediate package. Intermediate package shall be box of appropriate size constructed in accordance with PPP-566 or PPP-B-676, except commercial colors will be acceptable, PPP-B-636, type CF, class domestic. Closure shall be adequate prevent accidental opening under normal handling.

1.4 Level C. Units shall be packaged in standard commercial ntainers of the size and kind commonly used, which will afford e degree of protection required for shipment and use of the oduct for its intended purpose.

2 Packing.

2.1 Level A.

2.1.1 Exterior container. Exterior container shall be designed r type 2 load, and constructed in accordance with PPP-B-585, yle 3, for class 3 use; PPP-B-601, table II, using type III, ass 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, ass weather-resistant.

2.1.2 Case liner. Each level A wood box shall be lined with waterproof case liner conforming to MIL-L-10547. Closure and aling shall conform to applicable paragraphs of appendix thereto. se liner shall not be required for fiberboard boxes.

2.1.3 Closure. Closure of wood boxes shall be in accordance th appendix of applicable box specification. Closure and terproofing of each fiberboard box shall be as specified in e appendix of PPP-B-636.

2.1.4 Strapping. Strapping, when required, shall be in accord-ce with appendix of applicable box specification.

TE: Strapping shall not be required for shipments forwarded a receiving activity within the continental limits of the ited States for storage and redistribution.

2.2 Level B.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Sounds shall be for use in dilation of the urethra.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size (see 1.2).
(c) Selection of applicable levels of packing.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the currentversion before use.

.3 This specification does not cover all types, classes, rades, or sizes of the commodity indicated by the title of ıis specification, or those which are commercially available, ıt is intended to cover the types which are normally procured ɔ meet requirements.

ıstodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

ɜview activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 86

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of *products for ultimate use* by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements .

SPECIFICATION

GG-S-00571a

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.