MILITARY SPECIFICATION SHEET

CABLE ASSEMBLY, (TYPE I), ELECTRONIC TEST EQUIPMENT, (2 POLES, 3 WIRES, 125 VOLTS, 13 AMPERES, 50/60 HERTZ), GROUNDING PLUG CONNECTOR AND MIL-C-5015 CONNECTOR WITH CABLE CLAMP

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The complete requirements for acquiring the cable assembly described herein shall consist of this specification and the latest issue of MIL-C-28777.




FIGURE 1. Cable assembly.

REQUIREMENTS:

Dimensions and configuration: See figure 1.

Cable-plug connector interface: The green lead (with or without a yellow stripe) shall be connected to the safety ground terminal, the black lead shall be connected to the line terminal, and the white for neutral shall be connected to the remaining terminal of the plug connector.

Type I equipment connector: The MS3106A-14S-7S shall have the safety ground lead (green) connected to pin B, the live lead (black) to pin C, and the neutral lead (white) to pin A.

Part number: M28777/1-1.

© denotes change

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T15:12Z
Check the source to verify that this is the current version before use.

FSC 6150

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Custodians:
Army - CR
Navy - EC
Air Force - 99

Review activities:
Navy - AS, MC
Air Force - 11, 85
DLA - GS, ES

Agent:
DLA - ES

Preparing activity:
Navy - EC

(Project 6150-0170-01)

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T15:12Z
Check the source to verify that this is the current version before use.