
FED. SUP CLASS
6640



PART NUMBER	NOMINAL RESERVOIR DIMENSIONS MM (IN)			APPROXIMATE CAPACITY L (GAL)	MAXIMUM WATTS
	LENGTH	WIDTH	DEPTH		
MS36002-1	31 (12)	19 (7.5)	19 (7.5)	13 (3.5)	800
MS36002-2	45 (18)	31 (12)	19 (7.5)	32 (8.6)	1600
MS36002-3	40 (16)	40 (16)	31 (12)	64 (17.0)	2400
MS36002-4	76 (30)	31 (12)	19 (7.5)	53 (14.1)	2400
MS36002-5	90 (36)	45 (18)	24 (9.5)	117 (31.1)	2750

MATERIAL: Reservoir, exterior walls and cover shall be of corrosion-resisting steel in accordance with Federal Specification QQ-S-766, Class 304.

FINISH: Corrosion-resisting steel shall be ground to a No. 4 commercial finish.

DESIGN: Bath shall be of double wall construction and designed with an insulated water leakproof chamber enclosed in a moisture-proof jacket that will not settle and that is a minimum of 2.5 cm thick. Bath shall be equipped with a gabled cover, a heater, thermoregulator, switches, pilot light and a water circulation system. Also, bath shall accommodate a graduated glass thermometer.

Bath shall have a solid state control that will regulate temperature from 5°C above ambient to +100°C within + 0.2°C or better. Bath shall have a diffuser plate that when reversed will reduce the depth of the chamber to 127 mm for supporting test tube racks. A drain cock shall be included for draining the bath.

All electrical components and insulation shall conform to the standards of Underwriters Laboratories, Incorporated.

Bath shall operate on 120 or 240V and 50 or 60 Hz.

NOTES:
1. Voltage and cycles shall be specified on date of invitations for bid.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 15486.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
AIR FORCE - 82

APPROVED 16 JAN 1984 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		WATER BATH, ELECTRIC, CIRCULATING	MS36002
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 COORDINATED PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1210

Source: https://assist.dla.mil -- Downloaded: 2026-10-28T17:30Z
Check the source to verify that this is the current version before use.