
FED. SUP CLASS
6640



PART NUMBER	CAPACITY LITERS	OUTSIDE DIMENSIONS (CM) LENGTH	WIDTH	HEIGHT
MS36282-1	113.6	77.5	54.0	45.7
MS36282-2	227.1	77.5	95.3	45.7

MATERIAL: All materials of construction in contact with the water shall be non-toxic and noncorrodible. The viewing windows shall be made of thermal-insulating glass. All other materials shall be suitable for construction that will enable proper function and strength.

DESIGN: MS36282-1 shall have one 63.5 x 26.7 cm viewing window and MS36282-2 shall have three 63.5 x 26.7 cm viewing windows. The inside shall be coated with ocean blue-colored epoxy and the exterior antique white-colored high-impact laminated plastic on marine plywood. The aquarium shall have two independent water circulation systems, each with its own pump. A refrigeration system with a self-lubricating, hermetically sealed compressor shall thermostatically control the temperature range from 3°C to ambient +1°C (lower limit may vary with room temperature). Also, a cover plate, hydrometer/thermometer and operating manual shall be included.

The power requirements are 110 volts AC, 60 Hz, 6 amp (grounded plug). The electrical wiring shall conform to all the requirements of the Underwriters Laboratories, Code Number UL73 and UL1262. The approximate filled weight of MS36282-1 is 195Kg and MS36282-2 is 347Kg.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

APPROVED 5 JAN 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		AQUARIUM, LABORATORY	MS36282
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1132

Source: https://assist.dla.mil -- Downloaded: 2026-10-23 17:37:2
Check the source to verify that this is the current version before use.