INCH-POUND

A-A-54889
23 July 1993

COMMERCIAL ITEM DESCRIPTION

BLOOD FLUID WARMER AND PRESSURE INFUSION SYSTEM

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a blood-fluid warmer and pressure infusion system for trauma and massive fluid replacement.

Salient characteristics:

The unit shall be a fluid management system for trauma and massive fluid replacement. It shall provide blood and IV fluid warming plus high speed infusion. It shall be able to mount completely to an IV pole.

The unit shall warm cold blood to normothermic temperature at flow rates up to at least 450 ml/minute and physiological fluids at flow rates up to at least 750 ml/minute.

The system shall consist of:

1. A fluid reservoir and warmer that maintains circulating water temperature of 40°C plus or minus 0.2°C, alarms and displays for maintaining proper water temperature and level, and an automatic high temperature shut off at 41°C.

2. A pressure infusion system with two pressure chambers that accept both 500 ml or 1000 ml fluid bags, and a built-in compressor.

The system shall allow for fast, easy, disposable set-up, shall have an alarm that signals improperly installed disposables, and shall not function unless disposables are properly installed.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The system, excluding pressure chambers, shall not exceed 40 inches in height, 6 inches in width, and 7 inches in depth. Weight shall not exceed 20 pounds.

The system shall function from 115 VAC, 50/60 Hz electrical supplies.

Workmanship. The blood-fluid warmer and pressure infusion system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One blood fluid warmer and pressure infusion system, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one blood-fluid warmer and pressure infusion system.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:39Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one blood-fluid warmer and pressure infusion system.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

* Maximum flow rate of blood for one minute shall measure at least 450 ml, and for physiological fluid at least 750 ml.

* Circulating water temperature shall measure 40°C, plus or minus 0.2°C. High temperature shut off shall operate at 41°C.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1. This system is intended to be fully compatible with a full range of sterile blood and fluid warming disposable sets:

Disposable Set	NSN
Blood/Fluid Administration Set	6515-01-344-6612
Blood/Fluid Administration Set	6515-01-334-0853
Blood/Fluid Administration Set	6515-01-344-6106
Filter Air Eliminator for D-50/100	6515-01-334-0855
Pre-Filter	6515-01-334-0856
High-Flow Injection Port and Extension	6515-01-344-5249
High-Flow Extension Y-Type	6515-01-334-0854
Extension Line	6515-01-344-5255
Irrigating Set	6515-01-344-5254

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.

Plus the following patient rewarming disposable system:

Arterial to Venous Rewarming set, capable of rewarming cold patients at rates up to 1 C every 15 minutes.

2. The following National Stock Number is covered by this document.

6515-01-370-5019

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5620

Location: WARMR17C.DOC

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54889	2. DOCUMENT DATE (YYMMDD) 23 July 1993

3. DOCUMENT TITLE
BLOOD FLUID WARMER AND PRESSURE INFUSION SYSTEM

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	d. IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 9879C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.