INCH-POUND

A-A-54925
15 October 1993

COMMERCIAL ITEM DESCRIPTION

ANALYZER, MIDDLE EAR

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a middle ear analyzer to be used in testing for tympanometry, ipsilaternal and contralateral acoustic reflex testing, and screening audiometry.

Salient characteristics:

The middle ear analyzer shall be capable of testing for tympanometry, ipsilateral and contralateral acoustic reflex testing and screening audiometry. It shall have a probe tone frequency of 226 Hz, plus or minus 3 percent, a probe tone level of 85 dB SPL, plus or minus 1 dB, measured in a 2 cm coupler.

The analyzer shall have the following typanometry mode:

* Pressure range: plus 200 to minus 400 daPa

* Accuracy: plus or minus 15 percent or plus or minus 10 daPa whichever is greater

* Rate of sweep: 600 daPa/s except near typmpanogram peak where sweep slows down to 200 daPa/s

* Test time: approximately 1 second

* Direction of sweep: positive to negative pressure.

* Compliance ranges: 0 to 1.5 ml, 0 to 3.0 ml.

* Compliance accuracy: plus or minus 5 percent or plus or minus 0.1 ml, whichever is greater.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

* Ear canal volume range: 0 to 6 ml

The analyzer shall have the following reflex modes:

* Frequency: 500, 4 K Hz
  IPSI: 78, 85, 95 dB HL
  Contra: 80, 90, 100 dB HL

* Frequency: 1 K, 2 K Hz
  IPSI: 85, 95, 105 dB HL
  Contra: 90, 100, 110, dB HL

The analyzer shall have the following audiometry mode:

* Frequency: 500, 1 K, 2 K, and 4 K Hz

* HL selector/tone bar: plus 25 dB and plus 35 dB HL

The analyzer shall have a printer plotter, shall be approximately 17 X 14 X 5 inches in size and approximately 15 lbs. in weight.

The analyzer shall be supplied with the following accessories:

* Hand-held probe

* Contralateral headset

* Eartips - (3 each, 4 sizes)

* Charts - (paper) 2 rolls (270 charts per roll)

* Pens - 1 pkg. (4 pens)

* Test cavity

* Instruction manual

The analyzer shall operate from 115/230 Vac, 50/60 Hz power.

Workmanship. The middle ear analyzer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One middle ear analyzer, together with accessories, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one middle ear analyzer.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one middle ear analyzer.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

* Verify probe tone frequency of 226 Hz, plus or minus 3 percent.
* Verify probe tone level of 85 dB SPL, plus or minus 1dB SPL.
* Verify typanometry mode pressure range of plus 200 to minus 400 daPa, plus or minus 15 percent or plus or minus 10 daPa whichever is greater.
* Verify compliance ranges of 0 to 1.5 ml, 0 to 3.0 ml, plus or minus 5 percent or plus or minus 0.1 ml, whichever is greater.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6625-01-372-7772

MILITARY INTERESTS:

Custodians:

Air Force-03
Army-MD
Navy-MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6625-0796
Location: MIDLEAR17

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54925	15 October 1993

3. DOCUMENT TITLE
ANALYZER, MIDDLE EAR

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.