INCH-POUND

A-A-54934
17 December 1993

COMMERCIAL ITEM DESCRIPTION

PLIERS, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-37083A which is cancelled.

Salient characteristics:

This Commercial Item Description covers the following three dental pliers suitable for holding and placing pins and points.

No. 102
No. 111
No. 137

Material. Pliers shall be fabricated from either carbon or stainless steel as specified. Stainless steel shall be in accordance with ASTM Designation F899 for composition only.

Hardness. After proper stress relieving, the pliers shall have the Rockwell hardness as specified below for the applicable material. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Distortion. The handle of the pliers shall not show pliability or distortion under normal usage.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Material, hardness and plating

	Material	Rockwell hardness	Plating
Carbon Steel	Carbon content: 0.40 to 1.10 present. Managanese content: 0.45 to 0.90 percent except when carbon content is over 0.90 percent, the manganese content shall be a maximum of 0.50 percent. Phosphorous content: Maximum of 0.040 percent. Sulfur content: Maximum of 0.050 percent.	C40 to C45, with exception of the screw or pin. When the carbon content is over 0.90 percent, the hardness shall be C48 to C52.	Plating shall be chromium over nickel.
Stainless Steel	Class 4, Type 420	C40 to C48, with exception of the screw or pin.	None
	Class 4, Type 410	C40 to C45, with the exception of the screw or pin.	None

Design and dimensions. The pliers shall be suitable for the purpose specified below and shall conform to the applicable figure as specified. The pliers shall be furnished in the number specified. Number 102 shall be a cone socket pliers. Number 111 shall be a crown pliers, with serrated beaks. Number 137 shall be contouring pliers.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Figure number and purpose of each instrument

Number	Figure	Purpose
No. 102	1	Holding, tightening and removal of cone socket points (in and from handles), and general laboratory use.
No. 111	2a, 2b	Ligating, placing and removing pins.
No. 137	3	General contouring work in the dental laboratory.

Locks. The joint of the pliers shall be box-lock or lap-joint type. The edges of joints shall be properly beveled to prevent pinching of or injury to tissue.

Finish. The pliers shall have a mirror, satin, matte or a smooth, non-reflective surface finish. The inside face of the jaws of pliers No. 102 and No. 111 shall be serrated as specified in the applicable figure. The inside face of jaws of pliers No. 137 shall be smoothly contoured surfaces as specified in the applicable figure.

Plating. Carbon steel pliers shall be plated as specified.

Corrosion resistance. All pliers shall meet the boil test requirements as specified in ASTM Designation F1089. Pliers fabricated of stainless steel shall also meet the copper sulfate test requirements as specified in ASTM Designation F1089.

Identification markings. The pliers shall be marked with the applicable instrument number and the contractor's name or registered trademark. Stainless steel pliers shall be marked with the word "stainless". As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless is acceptable. Pliers of foreign manufacture shall be marked with the country of origin. All markings shall be legible.

Workmanship. The entire pliers shall be free from defects such as cracks, coarse grind marks, pits, feathers or burrs which detracts from their appearance or impairs their serviceability, proper function, or intended use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One pliers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one pliers. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one pliers. Sampling shall be inspection level S-2. Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-542-7480	PLIERS, DENTAL, No. 102, 5-7/6 inch (13.81 cm)
6520-00-543-0150	PLIERS, DENTAL, No. 111, 5-1/8 inch (13.01 cm)
6520-00-543-3750	PLIERS, DENTAL, No. 137, 5-1/4 inch (13.33 cm)

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6520-2535
Location: word\doc\preston\pliers


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.



NOTES:
1 ALTERNATE WIDTH OF LOCK: 3/8 +1/16 - 1/64.
2 ALTERNATE LENGTH OF LOCK: 1/2 1/32.

FIGURE 1 PLIERS, DENTAL, NO. 102

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.



NSN 6520-00-543-0150

STYLE 2

FIGURE 2a PLIERS, DENTAL, NO. 111

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

	DIMENSIONS																					
	A	B	C	D	E	F	G	H	J	K	L	M	N	O	P	Q	R	S	T	U	V	W
STYLE 2	$5\frac{5}{16}$	$5\frac{7}{8}$	$1\frac{3}{4}$ $\pm\frac{1}{16}$	$\frac{3}{4}$	$\frac{5}{32}$	$\frac{3}{16}$ $+\frac{1}{32}-\frac{1}{16}$	$\frac{7}{32}$	.093	43°	$\frac{13}{16}$	$\frac{11}{64}$	$\frac{7}{16}$	$\frac{15}{32}$ $+\frac{1}{32}-\frac{1}{16}$	$1\frac{7}{8}$	1	$\frac{5}{16}$	$\frac{1}{8}$	$\frac{7}{16}$ $+\frac{1}{32}-\frac{3}{32}$	$\frac{1}{16}$	$\frac{7}{64}$	$\frac{3}{64}$ R.	$\frac{5}{32}$
ALT. STYLE 2	$4\frac{7}{8}$	$5\frac{3}{8}$	$1\frac{3}{8}$ $\pm\frac{1}{4}-\frac{1}{16}$	$\frac{5}{8}$	$\frac{5}{32}$	$\frac{1}{8}$	$\frac{7}{32}$	.093	47°	$\frac{13}{16}$	$\frac{9}{64}$	$\frac{13}{32}$	$\frac{13}{32}$	$2\frac{1}{16}$	$1\frac{9}{32}$	$\frac{9}{32}$	$\frac{5}{32}$ $+\frac{1}{32}-\frac{1}{16}$	$\frac{3}{8}$ $\pm\frac{1}{32}-\frac{1}{16}$	$\frac{3}{32}$	$\frac{1}{16}$	$\frac{3}{64}$ R.	$\frac{5}{32}$
TOLERANCES	$\pm\frac{1}{16}$	$\pm\frac{1}{8}$		REF.						BSC	$\pm\frac{1}{64}$			$\pm\frac{1}{8}$	$\pm\frac{1}{8}$				$\pm\frac{1}{64}$	$\pm\frac{1}{64}$	$\pm\frac{1}{64}$	

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES

TOLERANCES ON:

FRACTIONS	DECIMALS	ANGLES
± 1/32	± .015	± 2°

FIGURE 2b PLIERS, DENTAL, NO. 111

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.




FIGURE 3 PLIERS, DENTAL, NO. 137

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54934	17 December 1993

3. DOCUMENT TITLE

PLIERS, DENTAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.