INCH-POUND

A-A-54936
20 December 1993

COMMERCIAL ITEM DESCRIPTION

# GUIDE, CIRCUMCISION, PLASTIC, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-G-37061(1), which is cancelled.

This Commercial Item Description covers a sterile, disposable, plastic circumcision guide for use on male infants.

Salient characteristics:

General. Shall be a sterile, disposable, plastic device, suitable for performing circumcisions of infant male humans.

Material. Shall be fabricated of clear butyrate plastic and shall be regular size (1.3cm). Shall consist of a plastic bell which is inserted under the prepuce and a break away plastic handle.

Style, design and construction. Shall be shaped generally like a spherical sector to fit over the glans penis and under the prepuce. A break away handle shall be attached at opposite sides of the short diameter. A shoulder shall be present on the exterior near the short diameter. A shoulder shall be present on the exterior near the short diameter so that the ligature is maintained in position over the prepuce. The guide shall be in accordance with figure 1. The linen ligature shall be 12 to 14 inches long.

Sterility. The circumcision guide and linen shall be sterile when tested in accordance with the United States Pharmacopeia (USP) or American Association for Medical Instruments (AAMI) guidelines. Sterility Assurance Level shall be 0.0001.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Instructions. Each unit package shall be supplied with 125 copies of special information and instructions for parents. The information should include as a minimum the following:

a. The plastic rim usually drops off in 5 to 8 days following circumcision.

b. No special dressing is required and the baby can be bathed just as though he had not been circumcised.

c. Notify your doctor if healing does not proceed as described, if any unusual swelling is observed or if the plastic ring is not off within 8 days.

In addition, each box shall contain instructions. Instruction shall include anatomical illustrations and correct methods of insertion, ligation and excision.

Workmanship. The circumcision guides shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 125 individually packaged guides together with instructions, as specified constitutes one unit. Each guide shall be packaged in a sealed, peel-open or easy-open package capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one circumcision guide as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one circumcision guide. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications is covered by this document:

| NATIONAL STOCK NUMBER | ITEM IDENTIFICATION |
|---|---|
| 6515-00-364-7867 | GUIDE, CIRCUMCISION, PLASTIC 1.3 cm size, Sterile, Disposable, 125s |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5419
Location: tarun/guides/djb



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.





FIGURE 1 CIRCUMCISION GUIDE



Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER: A-A-54936 | 2 DOCUMENT DATE (YYMMDD): 931220 |
|---|---|---|

3 DOCUMENT TITLE: GUIDE, CIRCUMCISION, PLASTIC, DISPOSABLE

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION | |
|---|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7 DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a. NAME: Defense Personnel Support Center<br>ATTN: DPSC-MSE | b. TELEPHONE (Include Area Code)<br>(1) Commercial: 215-737-2870<br>(2) AUTOVON: 444-2870 |
|---|---|
| c. ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.