INCH-POUND

A-A-54937

25 January 1994

COMMERCIAL ITEM DESCRIPTION

VACUUM CLEANER, PLASTER

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a portable vacuum cleaner used to dispose of plaster dust during orthopedic procedures.

Salient characteristics:

The portable plaster vacuum cleaner, shall be 12 inches long x 11 inches high x 5 inches wide, suitable for use in collecting plaster dust during cast cutting/trimming orthopedic procedures. As an alternate, the vacuum cleaner may be approximately 9 inches diameter by 23 inches high. The unit shall operate on 115/230 volt, 50-60 Hz, AC. The vacuum cleaner shall be capable of generating a minimum of 46 inches of water operating with a motor drawing no more than 1.16 horsepower (865 watts).

Materials and constructions. The vacuum cleaner shall have a high impact plastic casing with a handle integrated into the top of the casing. A minimum of 5 feet of 1-1/4 inch I.D. plastic tubing shall be distally attached to an opening on the front of the vacuum cleaner with a twist-lock mechanism.

The opposite (proximal) end of the tubing shall have a metal male connecting piece that secures in place on the blade (distal) end of the cast cutter.

Input power. The cleaner shall be capable of operating from the following line sources: 110 volts, plus and minus 10 percent, and 230 volts, plus and minus 10 percent alternating current, multiple frequency, 50 and 60 Hz plus or minus 2 Hz.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Modifications to the offeror's standard commercial item which will allow operation on all the above voltage levels and 50/60 Hz may be acceptable.

The design modification may be integral to the equipment or may consist of a separate external equipment item(s). The conversion kit shall include adequate instructions for the Government to perform the modifications. Complex modifications requiring more than one manhour to accomplish are not acceptable.

Metal parts shall be fabricated of a suitable corrosion-resisting steel, properly hardened and passivated. Material used in the fabrication of the vacuum cleaners shall be of the highest quality throughout and suitable in all respects for the intended use.

The vacuum cleaner shall have an electric cable approximately 12 inches long attached to the lateral portion of the metal tube connection. This cable shall be equipped with a female plug end that is compatible with the 3-prong hospital grade male plug end on the cast cutter electrical cable.

The cleaner shall also be provided with a 5 to 6 foot power cable with proper provisions for plugging into 115 or 230 volts.

Accessories. The vacuum cleaner shall be supplied with a plastic or coated metal wall mounting bracket, a semi-disposable cloth plastic bag, and all other components necessary to perform its function.

Compatability. The distal end of the vacuum cleaner hose shall have a male connecting piece specifically designed to attach to the blade housing of an orthopedic cast cutting saw used to cut casts without patient injury. See Note 1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Identification marking. The following marking shall be furnished:

a. Name or registered trademark of the manufacturer.
b. Type or model number.
c. Serial number.
d. Electrical characteristics.
e. National Stock Number (NSN).
f. Federal Supply Code for Manufacturers (FSCM).
g. Quarter and year of manufacture (3Q81 - designates third quarter of 1981).

Nameplates shall be permanently and securely attached to the equipment. Adhesives shall not be used. Equipment shall be permanently and legibly marked in accordance with the general marking requirements of MIL-STD-130. The nameplate shall be located on an external equipment surface which is readily accessible.

Interchangeability. Parts having the same manufacturers part number shall be fully and functionally interchangeable. All major components of the assembled system having the same model number shall be physically and functionally interchangeable.

Engineering design changes. The contractor shall notify the provisioning activity of all changes whether of a production or modification type which are approved for incorporation into the end item furnished under the contract and which modify, add to, delete, or supersede parts in the end item or its supporting equipment.

Safety. All items shall conform to the applicable requirements of a nationally recognized standard for the safety of medical equipment such as UL544 or equal of the issue in effect on the date of solicitation. All items shall be approved as conforming to such a standard by an nationally recognized testing laboratory.

Service data. The contractor shall furnish, with each unit, two copies of service data containing complete instructions for assembly, operation, maintenance, and a list of component parts. All information contained in the manual shall be indexed so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identification. Identical or interchangeable replacement parts shall be identified by a single part number only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Reusable container. No reusable container is required.

Workmanship. Workmanship shall be first class throughout. Vacuum cleaners shall be free from defects which detract from their appearance or which may impair their serviceability.

Unit. Each (EA). One vacuum cleaner, complete with all accessories and manuals, as specified, constitutes one unit of issue.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 (percent defective) and an inspection level of II. Unit of product for examination purposes shall be one plaster vacuum cleaner, as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one plaster vacuum cleaner.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

a. Performance test. Unit shall be connected to a 115 volt, 60 Hertz power source and then operated. During operation, the unit shall be examined for suction at the inlet to determine that it meets or exceeds the specified 46 inches of water minimum. It shall also be determined at this time that power draw of unit does not exceed its rated maximum of 1.16 horsepower (865 watts). There shall be no air leaks at the tank gasket or hose connection. Unit is then connected successively to each of the DEPMEDS power sources and test repeated for suction only.

b. Operational test. Unit shall be connected to 115 volt, 60 Hertz power source and operated under a full load power of 750 watts for 5 minutes. When operated from a cold start, the temperature rise of the motor housing shall be no more than 40 degrees Fahrenheit over ambient. Also, the unit shall operate under a full load power of 750 watts of continuous operation for 100 hours without breakdown or failure. The 100 hour test may be waived upon request by the contractor, when the contractor has previously furnished items in accordance with the requirements of this specification, and there has been no major design change of motor or drive mechanism.

c. First article test. Where a first article sample is required, all tests shall be performed on that unit.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

NOTES 1. The vacuum cleaner is intended to be compatible with the three orthopedic cast cutters below, which attach to it by means of the vacuum hose as well as the electric power plug.

6515-01-291-5447 (J)
6515-01-315-3739 (J)
6515-01-206-6028 (W)
6515-00-323-4510 (D)

2. The following National Stock Number is covered by this document:

6515-01-167-6642 (W)

MILITARY INTERESTS:

Custodians:

Air Force-03
Army-MD
Navy-MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6515-5656
Location: VACCLNR1D - CID

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54937	2. DOCUMENT DATE (YYMMDD) 25 January 1994

3. DOCUMENT TITLE
VACUUM CLEANER, PLASTER

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.