METRIC

A-A-54938
12 January 1994

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC
with Filter

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a hypodermic needle used in filtering out particulate matter from intravenous medication.

Salient characteristics:

Except as specified herein, the needle shall be in accordance with ISO 7864, Sterile Hypodermic Needles for Single Use.

General. The needle shall consist of a hub, cannula, filter and sheath (needle guard).

Material. The cannula shall be fabricated from Class 3, Type 304 stainless steel in accordance with ASTM Designation F899. The hub shall be fabricated from rigid polypropylene or polycarbonate that meets the United States Pharmacopeia (USP) requirements for Class IV plastics or aluminum. The sheath shall be fabricated from rigid plastic. The exterior of the cannula shall be lubricated with inert non-toxic medical grade silicone compound. The filter shall be fabricated from stainless steel. As an alternate, the filter shall be fabricated from acrylic copolymer with a nylon mesh support.

Style and design. The needle shall be an 18 gauge disposable non-coring needle with a sintered, non-shedding cup shaped, 5 micron microporous filter in the hub. As an alternate, the needle cannula shall have an opening on the side. The needle shall have an overall length of 38.1 millimeters (mm) (1-1/2 inches) ± 1.6 mm (1/16 inch)

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The needle shall be furnished with a removable, protective sheath. The tip of the point shall be ground to maximum keenness. The heel of the needle point shall be dull to prevent coring. The needle shall have a Luer lock hub in accordance with ISO 594-2.

Color Coding. The hub of the needle shall be color coded. The color designators specified in ISO 6009 are optional. If colored the needle sheath is not required to meet ISO 6009.

Aluminum hub. Needles with alumimum hubs shall show no evidence of corrosion when covered with 3 percent sodium chloride solution at 20-25° for one hour. There shall be no discoloration visible or any sign of reaction.

Non-toxic and pyrogen-free. The needle shall be non-toxic and pyrogen-free when tested in accordance with USP.

Sterility test. The needle shall be tested for sterility in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Workmanship. The needle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred individually packaged sterile needles as specified, constitutes one unit. The individual package shall be a sealed peel-open type or rigid tamper proof container capable of maintaining sterility of contents unless the package is damaged or opened. The package design shall be such as to permit ready dispensing of the needle from the package by sterile techniques and by attaching the syringe to the hub of the needle without touching the needle with hands or other instruments.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:40Z
Check the source to verify that this is the currentversion before use.

specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one needle, as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one needle. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE (1): The following National Stock Number and its respective Item Identification is covered by this document:

National Stock Number	Item Identification
6515-01-173-8882	NEEDLE HYPODERMIC, Disposable 18 gauge

NOTE (2): ANSI and ISO standards are available from the American National Standards Institute, Inc., 11 West 42nd Street, New York, NY 10036.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5583
Location: c:mark\filter

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54938	940112

3. DOCUMENT TITLE

NEEDLE, HYPODERMIC with Filter

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. [illegible]	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: 215-737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil - Downloaded: 2010-26T09:40Z
Check the source to verify that this is the current version before use.