INCH-POUND

A-A-54943
10 February 1994

COMMERCIAL ITEM DESCRIPTION

CLEANER, VACUUM, ELECTRIC

Portable, Wet/Dry, Infectious Waste

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description a wet/dry, electric vacuum cleaner for general use, but also including surgical and sterile preparation areas for infectious waste pick-up.

Salient characteristics:

The unit shall be a compact, lightweight, self-contained, electric-motor-driven, portable vacuum cleaner capable of being used for both dry and wet pick-up. The tank shall have a dry capacity of not less than one bushel and a wet capacity of not less than 6.0 gallons.

The cleaner shall be suitable for general use and in the vacuum removal of wet and dry microbiological waste from surgical areas and sterile preparation areas. It shall be capable of removing water, other fluid waste materials, small solids, mud-type materials, and dust.

The wet/dry vacuum cleaner shall consist of, at least, an electric vacuum motor, tank, carriage, filter system with installed filter mechanism, positive seal hose adapter, easy emptying capability, and portability. It shall also include the following accessories:

One each - 12 inch master tool kit (includes tool for carpet, squeegee and bristle shoes)

One each - Lightweight, crushproof, non-marking hose, 10 feet x 1-1/2 inch diameter

AMSC N/A FSC 7910

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

One each - 5 foot x 1-1/2 inch double curved aluminum wand with quick disconnect

The vacuum cleaner tank shall be fabricated of corrosion-resistant stainless steel. The finish shall be durable and unaffected by contact with any cleaning and sanitizing agents. A 1-1/2 inch diameter positive seal hose adapter shall be provided to assure high flow rate.

The filter system shall incorporate a HEPA filter, DOP tested and certified to retain 99.97 percent of particles 0.3 micron and larger. Three spare filters, also DOP tested, certified and suitably packaged shall be provided with each vacuum cleaner. Four wheels/casters at least three-inch diameter in size shall be provided. An automatic shut-off system to prevent overfilling of tank and damage to motor shall be provided.

The cleaner shall be capable of at least 80 inch water lift and 95 CFM air flow, with a 1-1/2 inch diameter orifice, with specified filter in place.

The unit shall withstand storage and transit conditions in the temperature ranges of minus $80^{\circ}F$ to $155^{\circ}F$ and air pressure of 2.47 psi.

The motor shall be rated one HP (0.75 watt), shall be multi-stage, constant duty, by-pass type, 7.2 amps continuous capacity, and shall maintain noise level below Office of Safety and Health Administration (OSHA) standards for commercial vacuum cleaners.

All electrical components (including motor) shall be treated for moisture and fungus resistance.

The unit shall operate on 115/230 volts, 50-60 Hz for both voltages, AC electrical power.

The input power cable shall be non-marking, 18 gage, grounded three-wire cable with strain relief. The three-prong male plug shall be according to ANSI C73.11. The length of the power cable shall be a minimum of 20 feet.

Safety instructions on proper use of the unit shall be permanently affixed on the tank in a prominent location.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Workmanship. The vacuum cleaner shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One vacuum cleaner with accessories, as specified, constitutes one unit.

The vacuum cleaner shall be manufactured in accordance with UL requirements for safety. Acceptable evidence of this shall be a UL listing or a certified test report from an independent laboratory acceptable to the Government.

Manuals.

The contractor shall furnish manuals, handbooks and/or brochures containing complete operation, installation and service/maintenance instructions (including pictures of illustrations, as necessary) with complete schematics and wiring diagrams. All manuals, handbooks, and/or brochures shall be written in the English language and all schematics and wiring diagrams shall use American electrical and electronics symbols. The manual shall include electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name and quantity required. All service notes, service memos, etc., pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included. These instructions shall be, as a minimum, that furnished to service engineers (or servicemen) that normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components.

Two complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, then only a total of two service manuals need to be supplied to each consignee. However, two operational manuals still need to be supplied with each unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one vacuum cleaner with accessories.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one vacuum cleaner with accessories.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

a. Measure usable tank capacity.

b. Measure water lift capability with specified filter in place.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

c. Measure air flow capability with specified filter in place.

d. Measure noise level with specified filter in place.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

7910-01-177-0622 (W)
7910-01-327-0757 (J)
7910-01-260-9064 (J)
7910-01-236-0893 (J)

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 7910-0083

Location: CLNR1DC.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54953	2. DOCUMENT DATE (YYMMDD) 10 February 1994

3. DOCUMENT TITLE
CLEANER, VACUUM, ELECTRIC, Portable, Wet/Dry, Infectious Waste

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 9879C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.