INCH-POUND

A-A-54944
25 February 1994

COMMERCIAL ITEM DESCRIPTION

FOOD BLENDER, ELECTRIC

(110\220 V, 50\60 Hz, AC)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an electric food blender for laboratory or radiology use.

Salient characteristics:

The unit shall be an electric chopping mixer for laboratory or radiology use in pulping, mixing and blending specimens or solutions to obtain a homogeneous mass.

The blender shall consist of a base with a motor and speed control, stainless steel blade, and stainless steel container with a two-piece vinyl lid.

The base shall be made of zinc die cast capable of accommodating the motor with speed control. All components of the base shall be finished with corrosion resisting coating.

The motor shall be rated at 840 watts and shall have permanently lubricated sealed bearings. By means of a solid state speed control, the motor shall be capable of running at seven speeds ranging from 3500 to 23000 RPM. A reduced motor speed at 50 Hz from 60 Hz for a speed range from 3500 to 21,000 RPM shall be acceptable. The speeds

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

shall be controlled with eight pushbuttons provided in front of the base assembly. The pushbutton marked "OFF" shall stop the motor when depressed. Any of the remaining seven pushbuttons sequentially marked from 1 through 7, when depressed, shall provide the motor speed selected. The line cord shall have three conductors connected to a three prong grounded plug.

The stainless steel blade shall have a thickness of 0.047 inch. The blade shall be of such design as to produce four separate cutting edges which shall be sharpened on their leading edges. The blade shall be mounted at its center on a shaft, the distal end of which shall be designed to engage the motor shaft. The shaft bearing, as well as the manner in which the assembly is made for mounting in the container, shall result in a leakproof unit when mounted.

The container shall be made of stainless steel which shall be not less than 0.022 inch thick at any point. The capacity of the container shall be 32 oz. when measured at 3/4 inch from the top of the container. The blade assembly shall be securely mounted in the bottom of the container so that a leakproof unit results. The container shall fit within the four prongs of the base. A stainless steel handle permanently attached to the container shall provide easy handling. A two piece plastic cover shall be provided that will fit snugly on top of the container.

Finish. The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Finish shall provide a pleasing professional appearance and be free of gaps, scratches, chips and other imperfections.

Workmanship. The food blender shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One electric food blender, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one electric food blender.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one electric food blender.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6640-01-176-0700

MILITARY INTERESTS:

Custodians:

Air Force-03
Army-MD
Navy-MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6640-1851

Location: BLENDER14C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54944	2. DOCUMENT DATE (YYMMDD) 25 February 1994

3. DOCUMENT TITLE
FOOD BLENDER, ELECTRIC (110/220V, 50/60 Hz, AC)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — *90/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.