METRIC

A-A-54945

15 March 1994

COMMERCIAL ITEM DESCRIPTION

AIRWAY, PHARYNGEAL, Guedel

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers Guedel pharyngeal airways used in maintaining a patent airway.

Salient characteristics:

Shall be a clean, clear, disposable pharyngeal airway, Guedel design, fabricated of polyvinylchloride for use in maintaining a patent airway. The airway shall be of kink resistant construction with smoothly rounded edges for patient comfort. Shall have a cut-away flange to facilitate oral suctioning and to accommodate endotracheal tube positioning. The tube portion below the cut-away flange shall be a rigid bite block to help assure airway integrity. The airway shall conform to provisions of American National Standard Institutes's (ANSI) Z79.3 for style and design. The airway air inlet port shall conform to ANSI Z79.2 for tracheal tube connectors and adapters.

The airways shall be provided in the following sizes:

Size	Category	Length
0	Infant	50 mm
1	Child	60 mm
2	Small Adult	70 mm
3	Small/ Medium Adult	80 mm
4	Medium/ Large Adult	90 mm
5	Large Adult	100 mm

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Biocompatability and pyrogens. The airway shall be non-pyrogenic in accordance with the United States Pharmacopeia (USP) and meet requirements of Type II plastics Biocompatability when tested in accordance with the USP.

Workmanship. The airway shall be free from defects which detract from its appearance or impair its serviceability. Shall be free from blisters, shrink marks, cuts, tears and foreign material.

Unit. Package (PG). The airways shall be packaged in a suitable sealed container capable of maintaining sterility of contents unless package is damage or opened.

For size 0 and size 1, one package containing 10 airways of one type only, as specified, constitutes one unit.

For sizes 2 and 4, one package containing 30 airways, as specified, constitues one unit.

For size 3, one package cotaining 50 airways, as specified, constitutes one unit.

For size 5, one package containing 100 airways, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one package of airways. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one package of airways. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE 1: ANSI standards are available from the American National Institute Inc., 11 West 42nd Street, New York, NY 10036.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the current version before use.

NOTE 2: The following National Stock Numbers and its respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-01-233-1915	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Small Adult, 70 mm long, 30s	2
6515-01-266-3804	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Infant, 50 mm long, 10s	0
6515-01-310-8338	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Child, 60 mm long, 10s	1
6515-01-313-4854	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Small/Medium Adult, 80 mm long, 50s	3
6515-01-233-1913	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Medium/ Large Adult, 90 mm long, 30s	4
6515-01-308-8949	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, Large Adult, 100 mm long, 100s	5

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

A-A-54945

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5692
Location: airway/intergraph/vab

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54945	940315

3. DOCUMENT TITLE

AIRWAY, PHARYNGEAL, Guedel

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defénse Personnel Support Center ATTN: DPSC-MSE	215-737-2870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://dodiet.dla.mil - Downloaded: 2020-10-26T09:40Z
Check the source to verify validity of this electronic version before use.