MIL-P-2586A
28 March 1972
SUPERSEDING
MIL-P-2586
7 February 1951

MILITARY SPECIFICATION

PERIMETER, OPHTHALMOLOGICAL, SCHWEIGGER, HAND

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

* 1. SCOPE

1.1 Scope. This specification covers one type of hand perimeter of the Schweigger design.

* 2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-16T15:09Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Style, design, and dimensions. Perimeter shall be of "hand" type and shall consist of a hand support carrying an eye rest and a graduated metal arc, whose angular position is recorded by an indicator on a small protractor attached to the front of the support. Shall have a small mirror at the back of the protractor to serve as a fixation point. Each instrument shall be supplied with the following accessories: five color disks, a metal holder for same, and charts designed for perimeter furnished.

* 3.2.1 Hand Support. Shall consist of a well shaped wooden handle approximately 5-3/4 inch length, and 7/8 inch diameter in its widest portion, which shall be securely and permanently attached by socket and ferrule attachment to a curved metal plate approximately 2-1/4 inches long, 1/2 inch wide and 1/8 inch thick curved to conform to the arc of the metal frame to which it shall be firmly and securely riveted approximately 30° from bottom end. The metal frame shall be made of soft iron and shall be approximately 5/8 inch wide, 1/8 inch thick and 11 inches long, curved to form an arc of 90°. Proximally the frame shall be bent upon itself forming a 1 inch saddle on upper surface; to this saddle shall be firmly and securely attached the eye rest rod. Distally the frame shall have an opening approximately 1 inch from end, which shall accept the bolt passing through and supporting movable arc and protractor scale.

3.2.2 Eye rest rod. Shall be of brass 5/16 inch diameter. Bottom end shall be firmly bolted to saddle on frame. To its upper end shall be attached a properly shaped eye rest.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

3.2.3 Graduated metal arc. Shall be of same material as metal arc frame (3.2.1) but shall be a 180° arc. Shall be attached in midpoint to frame by a bolt with a split nut on protractor side. Attachment shall form a friction fit so that arc is adjustable but not loose. Arc shall be graduated from center outward on each arm in 10° major numbered graduations and 5° minor graduations. Graduations shall be on outer surface of arc and on one edge; shall be white lines and figures. As an alternate, the arc shall be fabricated of duraluminum metal. Arc degrees (i.e. numbers and location) shall be slotted into the metal and showing color of metal.

3.2.4 Protractor scale. Shall be a metal disk approximately two inch diameter and 3/32 inch thickness attached to front of frame (3.2.1) with its graduated face outward. Shall be immobile and shall be appropriately graduated. Shall have a moving pointer moving with the graduated arc so that the position of the latter, in degrees, is readily determinable on the scale. Scale shall be in 5° graduations; the major graduations being numbered at each 15°.

3.2.5 Mirror. Shall be a round mirror 1/2 inch diameter, attached securely to the middle of the graduated arc on inner surface, facing eye rest and acting as a fixation point for eye.

3.2.6 Accessories. The following accessories shall be furnished with each instrument.

a. Five colored metal disks.
b. A metal disk holder.
c. Perimeter charts (right and left).

3.2.6.1 Metal disks. Shall be five round metal disks approximately 5/8 inch diameter 1/32 inch thick with an integral handle approximately 7/16 inch long. Each disk shall have a colored square center on each side approximately 3/16 inch square. The following colors shall be furnished:

Red	White
Blue	Green
Yellow	

3.2.6.2 Disk holder. Shall be a metal rod approximately 14-1/4 inches long. The distal end shall be split for approximately 1/2 inch for fit to disk handles, the proximal end shall have a suitable round wood handle approximately 4 inches long and 3/8 inch in diameter.

Source: https://assist.dla.mil -- Downloaded: 202610-16T15:09Z
Check the source to verify that this is the current version before use.

* 3.2.6.3 Perimeter charts. Shall be of Schweigger type suitable for instrument furnished. Not less than 100 each right and left shall be furnished. Shall be printed on a good grade of white paper. Chart for left visual field shall be in accordance with requirements of the current Military Purchase Description for FSN 6515-355-1220 Chart, Vision Perimeter Testing, Schweigger, Left. Chart for right visual field shall be in accordance with requirements of the current Military Purchase Description for FSN 6515-355-1240 Chart, Vision Perimeter Testing, Schweigger, Right.

* 3.2.7 Finish. All metal except eye rest rod, shall be finished in a dull black enamel. No polished metal parts shall reflect light into examinee's eyes when eye is on eye rest. As an alternate, all metal shall be finished in a dull gray paint except the pointer, nut, mirror, and handle attachments shall be stainless steel. As an additional alternate, shall be a black anodized finish. The pointer shall be nickel plated brass and the handle attachment shall be bakelite.

* 3.3 Identification marking. The perimeter shall be permanently and legibly marked with the name or registered trademark of the contractor. Perimeters of foreign (other than U.S.) manufacture shall be permanently and legibly marked with the name of the country of origin together with the name or registered trademark of the contractor.

* 3.4 Workmanship. Workmanship shall be first class throughout. Perimeters shall be free from defects which detract from their appearance or may impair their serviceability.

* 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metals, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one perimeter.

Table I. Sampling for Examination

Type of Defect	Inspection level	AQL (Percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II.

Table II. Sampling for tests

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
1-Perimeter	Dimensions	3.2	----	S-1	*

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 202610-16T15:09Z
Check the source to verify that this is the current version before use.

4.2.3 Sample unit usage. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The perimeter shall be examined to determine compliance with all specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but shall not be limited to the defects listed in Table III.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Perimeter not complete.
103	Perimeter not free of sharp edges, rough areas, burrs, cracks, grind marks or overlaps.
Minor	
201	Identification marking not present, correct, complete or legible.
202	Finish not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each perimeter, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not applicable	18 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Perimeter, metal disk holder, and disks shall be wrapped in neutral tissue or other suitable anticorrosive material.

* 5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IC-2.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

* 5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; or PPP-B-621, class 2.

* 5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

* 5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements as specified in Table I of PPP-B-636 shall apply for fiberboard boxes.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

* 6. NOTES

6.1 Intended use. The perimeters are intended for use in visual field tests to discover restrictions of indirect vision.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Federal stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-16T15:09Z
Check the source to verify that this is the current version before use.

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

Federal Stock No.	Item Identification
6515-355-1200	PERIMETER, OPHTHALMOLOGICAL, Schweigger, Hand

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.4 Supersession data. This specification includes requirements of Purchase Description No. 4 dated 8 December 1966 for FSN 6515-355-1200 PERIMETER, OPHTHALMOLOGICAL, Schweigger, Hand.

6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and Contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1631

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-P-25956A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-01-16T15:09Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-16T15:09Z
Check the source to verify that this is the current version before use.