MIL-P-2586A
AMENDMENT-1
15 June 1972

MILITARY SPECIFICATION

PERIMETER, OPHTHALMOLOGICAL, SCHWEIGGER, HAND

This Amendment forms a part of Military Specification MIL-P-2586A, dated 28 March 1972, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

3.2.1 Hand support. In first line, insert "or aluminum rod" after "wooden". In line 7, delete period (.) and add "or screwed to arc." In line 8, delete period (.) and add "or aluminum approximately 3/4 inch wide, 1/8 inch thick and 11 inches long curved to form an arc of 90°."

3.2.2 Eye rest rod. At end of paragraph add the following: "As an alternate the eye rest shall be a part of the support frame and covered with a soft plastic at the eye rest."

Page 3:

3.2.3 Graduated metal arc. At end of paragraph delete period and add the following: "or graduations shall be produced by silk screen method, white on black anodized surface."

3.2.4 Protractor scale. In second line after "thickness" add: "or 1-3/4 diameter and 1/16 thickness".

3.2.5 Mirror. In first line insert "to 5/8" after "1/2".

3.2.6.2 Disk holder. In third line insert "or fiber" after "wood".

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1651

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T15:09Z
Check the source to verify that this is the current version before use.