MIL-P-2253A
15 JANUARY 1952
SUPERSEDING
MIL-P-2253
30 June 1930

MILITARY SPECIFICATION

PLATE, EYELID, JAEGER

This specification was approved by the Departments of the Army, the Navy, and the Air Force for the use of procurement services of the respective Departments.

1. SCOPE

1.1 This specification covers a double-ended, hard rubber plate suitably formed for use in ophthalmic surgery. and shall be of the type and design known to the surgical instrument industry and the medical profession as plate, lid, hard rubber, Jaeger.

2. APPLICABLE SPECIFICATIONS, STANDARDS, DRAWINGS, AND PUBLICATIONS

2.1 The following specifications and standards, of the issue in effect on date of invitation for bids, form a part of this specification:

SPECIFICATIONS

FEDERAL

NN-B-601 —Boxes, Wood - Cleated-Plywood, for Domestic Shipment.

NN-P-621 —Boxes, Wood, Nailed and Lock-Corner.

NN-B-631 —Boxes; Wood, Wire-bound (for Domestic Shipment).

MILITARY

JAN-P-105 —Packaging and Packing for Overseas Shipment—Boxes, Wood, Cleated, Plywood.

JAN-P-106 —Packaging and Packing for Overseas Shipment—Boxes; Wood, Nailed.

JAN-P-107 —Boxes, Wood, Wire-bound (overseas Type).

JAN-P-120 —Packaging and Packing for Overseas Shipment — Cartons, Folding, Paperboard.

JAN-P-125 —Packaging and Packing for Overseas Shipment — Barrier-Materials, Waterproof, Flexible.

JAN-P-133 —Packaging and Packing for Overseas Shipment—Boxes, Set-Up, Paperboard.

JAN-P-140 —Packaging and Packing for Overseas Shipment — Adhesives, Water-Resistant, Case-Liner.

STANDARDS

MILITARY

MIL-STD-105—Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129—Marking of Shipments.

(Copies of specifications, standards, and drawings, required by contractors in connection with specific procurement functions should be obtained from the procuring agency or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.—Lid plates shall be made of hard rubber, phenolformaldehyde resins, or equal.

3.2 Construction.—The plate, lid, Jaeger, shall be a thin double-ended instrument with a shaft somewhat narrowed. The ends shall be flattened to produce a spatulate effect, and shall have one surface slightly convex and the other slightly concave. These curvatures

Source: https://assist.dla.mil -- Downloaded: 20261O-16T14:37Z
Check the source to verify that this is the current version before use.

shall be so arranged that each side of the finished instrument shall present one convex and one concave surface.

3.3 Dimensions. — Dimensions shall approximate the following inches.

Length ..... 3½ inches.
Shaft ..... ½ by ⅛ inches.
Blades ..... 15/16 by 18 and ¾ by ⅛ inches.

3.4 Workmanship.—Workmanship shall be first class throughout. Lid plates shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.—Samples shall be submitted in accordance with instructions furnished by the procuring activity. In addition, inspection at contractor's facility by inspector shall be conducted in accordance with Standard MIL–STD–105.

4.2 Inspection.—Inspection may be made throughout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibilty for faulty workmanship or materials which may be discovered at any time prior to final acceptance. Final inspection of the finished article shall be made either at point of production or at point of delivery, as specified (see 6.1). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 Tests.—Tests shall be conducted in accordance with the applicable specifications, as necessary, to determine compliance with this specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 *Unit packaging.*—One plate constitutes one unit. Each unit shall be wrapped in kraft paper or packaged in an envelope of suitable size and design.

5.1.2 *Intermediate packaging.*—Thirty-six unit packages (36 units) shall be packaged in a carton of appropriate size constructed in accordance with Specification JAN–P–120 or JAN–P–133. Each carton shall be adequately secured to prevent accidental opening.

5.2 Packing.

5.2.1 *For domestic shipment.*—Forty intermediate cartons (1440 units) shall be overpacked in an exterior container constructed in accordance with Specification NN–B–601, NN–B–621, or NN–B–631.

5.2.2 For overseas shipment.—Forty intermediate cartons (1440 units) shall be overpacked in an exterior container designed for type 2 load, and constructed in accordance with Specification JAN–P–105, JAN–P–106, or JAN–P–107. Closure and strapping shall comply with appendix of applicable specification. Each box shall be lined with a waterproof barrier (bag) conforming to Specification JAN–P–125. Closures shall be completely sealed by a continuous seam of water-resistant adhesive at least ¾ inch wide and conforming to Specification JAN–P–140.

5.2.3 Packing variation permitted.—When the number of units in any shipment is not the exact multiple of the number of units specified to be contained in each shipping container, the units not constituting the nearest such multiple shall be packed in an exterior container of suitable size and design and so constructed as to insure safe delivery to specified destination.

5.3 Marking. — Each intermediate carton and each shipping container shall be marked in accordance with Standard MIL–STD–129.

Source: https://assist.dla.mil -- Downloaded: 202610-16T14:37Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. — Procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Sampling required (see 4.1).
c. Inspection required (see 4.2).
d. Place of final inspection (see 4.2).
e. Whether packing for domestic or overseas shipment is required (see 5.2).

6.2 This specification covers the following item appearing in the Armed Services Catalog of Medical Materiel:

3–561–000 Plate, Eyelid, Jaeger.

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

Custodian:
Army—Medical Service

Other interest:
Army—T
Navy—Md
Air Force.

Source: https://assist.dla.mil -- Downloaded: 20260-16T14:37Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-16T14:37Z
Check the source to verify that this is the current version before use.