W-H-191B
May 27, 1975
SUPERSEDING
Int. Fed. Spec. W-H-0019la(DSA-DM)
December 9, 1968

FEDERAL SPECIFICATION

HEATER, HEAT TREATMENT PAD

This specification was approved by the Commissioner, Federal Supply Services, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers one type and two sizes of an electric heater for immersion type heating pads (see 6.3 and 6.4).

1.2 Classification.

1.2.1 Sizes. Heaters shall be of the following sizes:

Size 1	Four pad capacity.
Size 2	Eight pad capacity.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

FEDERAL SPECIFICATIONS:

PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Shipping, Fiberboard.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:25Z
Check the source to verify that this is the currentversion before use.

Federal Standards:

Fed. Std. No. 66 - Steel: Chemical Composition and Hardenability.
Fed. Std. No. 123 - Marking for Domestic Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards, as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL-L-3661 - Lampholders, Indicator Lights, Indicator-Light Housings, and Indicator-Lights Lenses, General Specification For.
MIL-L-10547 - Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-M-7298 - Manual Technical: Commercial Equipment Preservation-packaging, Methods of.
MIL-P-37058 - Pad, Heat Treatment.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

American National Standards Institute (ANSI) Standard:

C-73.11 Caps and Receptacles, 125 Volt, 15 Amperes, 2 pole, 3 wire Grounding Type, General Purpose.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Underwriters' Laboratories, Inc., (UL) Standard:

UL 544 Standards for Safety-Medical and Dental Equipment

(Application for copies should be addressed to Underwriters' Laboratories, Inc., 1285 Walt Whitman Road, Melville, L.I., New York, 11746, 207 East Ohio Street, Chicago, Illinois, 60611, 1655 Scott Boulevard, Santa Clara, California 95050 and 333 Pfingsten Rd., Northbrook, IL 60062.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of UL 544. Proof shall consist of one of the following:

A. UL label and listing of the product, provided the contractor shall have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the Contracting Officer, upon the latter's request.

B. Laboratory report from a qualified independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The qualification of such laboratory shall be determined by the Government which reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches, and parts list, all appropriately identified. Compliance with this paragraph does not excuse compliance with all other requirements of the contract and specifications.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the currentversion before use.

3.2 Material.

3.2.1 Corrosion-resisting steel. Corrosion-resisting steel shall conform to class 302 or 304 of FED-STD-66.

3.3 Construction. The heater unit shall consist of a water compartment, electrical heating system, thermostatic control, immersion-type heating pads, and a rack for supporting the heating pads in water. Unit shall operate satisfactorily from 120 volts, 50-60 Hz current. Unless otherwise specified, dimensions of the water compartment, cover and base shall be maintained within ± 1/16 inch. All sides, and mating construction parts, shall be square within ± 1/64 inch.

3.3.1 Size 1.

3.3.1.1 Water compartment. Size 1 water compartment shall be fabricated from 0.025-inch thick (No. 24, United States Standard (U.S.S.) gage) corrosion-resisting steel sheet. Inside dimensions shall be 10-1/4 by 8-1/8 by 12-1/8 inches deep. Compartment wall shall be curled outward at the top around a 3/16-inch rod, to form a smooth, round hem. Corners shall be formed to a 1/2 inch ± 1/32 inch radius. The wall of the compartment shall be welded down the center of the back side. The bottom of the water compartment shall be fabricated of 0.0375-inch thick (No. 20, U.S.S. gage) corrosion-resisting steel sheet, with a 3/8-inch + 1/16, -0 inch flange, formed downward around its edge. The flanged bottom shall be slipped into and be welded to the bottom edge of the tank wall. There shall be two black, heat-resistant plastic carrying handles securely bolted to the sides, 1 inch from the top of the water compartment. The compartment shall be embossed to form a ridge, 7/16 inch +1/16, -0 inch wide and 3/16 inch + 1/16, -0 inch high, with the lower edge of the ridge 1 inch above the bottom edge of the tank wall.

3.3.1.2 Cover. Size 1 cover shall have a top cover fabricated from 0.031 inch thick (No. 22 U.S.S. gage) corrosion-resisting steel. Cover shall be of one-piece construction, and shall be formed with a 1/8-inch-high flat dome, the perimeter of which shall be 1/2 inch + 1/16, -0 inch inside the outer edge of the cover. The front and side edges of the cover shall be flanged downward 5/16 inch + 1/16, -0 inch. The rear of the cover shall be curled upward around a 3/16-inch-diameter rod which shall serve as a hinge pin, and which shall be threaded at each end to accept corrosion-resisting steel, Acorn Cap-nuts. The cover shall be attached to a one-piece formed channel, designed to accept the ends of the hinge rod, to serve as a stop when the cover is opened, and to catch condensate when the cover is opened. The entire cover assembly shall be attached to the tank by means of four corrosion-resisting steel machine screws and Acorn Cap-nuts. The cover shall be provided with heat-resistant, plastic knob.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:25Z
Check the source to verify that this is the currentversion before use.

3.3.1.3 Base. The Size 1 base shall consist of a 3-inch high strip of 0.0375-inch thick (No. 20, U.S.S. gage) corrosion-resisting steel. The corners of the base shall be formed to a 5/8-inch radius. The bottom of the base shall be hemmed inward with a 3/8-inch + 1/16, -0 inch hem. The top edge of the base shall be curved inward on a 3/16-inch + 1/16 inch, -0 inch radius to fit tightly against the tank wall at the bottom of the embossed ridge of the tank. The base shall be joined at the back in such a manner that the ends of the base shall be butted together to form a seam. The butted seam shall be secured by a piece of metal of adequate size, centrally located over the seam, inside the base, and held in place by means of spotwelding on one side, and two nuts and screws on the other side. Formed brackets shall be attached by spotwelding inside the base at each of the two sides, to which an inspection plate shall be attached by means of two corrosion-resisting steel machine screws and nuts. This shall provide a 2-1/2 inch space under the water compartment to house all wiring and electrical components. The base shall fit sufficiently tight against the tank wall to prevent its slipping or moving from position when the heater is lifted by its handle. The inspection plate shall be 0.050 inch thick (No. 16 B. & S. gage) aluminum alloy, non-dyed anodized.

3.3.2 Size 2.

3.3.2.1 Water compartment. The Size 2 water compartment shall be drawn of 0.0375 inch thick (No. 20 U.S.S. gage) corrosion-resisting steel sheet, and shall have the inside dimensions of 12 inches x 14 inches x 11-3/4 inches deep. Exterior dimensions of the assembled unit shall be approximately 17-1/2 inches wide, 15-1/2 inches front to back and 19-1/2 inches high. The compartment shall be of one piece, deep drawn, seamless construction. The top of the tank shall be formed outward into a half-inch flange which shall be welded to a corrosion-resisting steel top rim to form a recessed seat around the top edge of the tank to accept a double-wall, insulated, corrosion-resisting steel lid. The top rim of the unit shall consist of 0.050 inch thick (No. 18 U.S.S. gage) corrosion-resisting steel. It shall be roll formed, notched and formed into a rectangle, whose corners shall be welded, ground smooth, and polished to match the finish of the entire unit. The outside edge of this rim shall be a half-inch flange formed 90° downward. This flange shall fit down over the outside corrosion-resisting steel jacket of the unit. The outer sides of the compartment shall be covered with a minimum 1-1/2 inches of fiberglass insulation. The outside jacket shall consist of 0.0375 inch thick (No. 20 U.S.S. gage) corrosion-resisting steel, whose corners shall be press-formed into a 1 inch radius to fit and conform snugly to the inside surface of the downard flange of the rim. Each end of this jacket shall terminate at the rear of the unit to accept a corrosion-resisting steel panel which shall form the rear side of this jacket. This rear panel is attached to the jacket by half-inch inbound flanges, and subsequent spot welding of these inbound flanges to similar inbound flanges of the jacket. The compartment shall have a means of draining, consisting of a threaded brass or copper flange brazed in a recessed hole

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

in the bottom. Attached to this threaded drain (at the rear of the unit) shall be a 1/2 inch brass drain valve, and a 36 ± 2 inch length of 5/8 inch garden hose terminated with standard hose fittings. A rubber adapter for attachment of the hose to an unthreaded faucet nozzle shall also be provided.

3.3.2.2 Cover. The Size 2 water compartment shall have a top cover fabricated from 0.0375 inch thick (No. 20 U.S.S. gage) corrosion-resisting steel. This lid shall be double-wall construction, consisting of two corrosion-resisting steel pans, 1/2 inch deep. The top pan of the lid shall be die formed at the rear to form a hinge running across the entire width of the lid. The hinge shall accommodate a 3/16 inch diameter corrosion-resisting steel rod, which shall pass through a corrosion-resisting steel hinge and stop attached to the back of the top rim of the unit. This hinge rod shall be threaded at each end to accommodate Acorn, cap-nuts. The stop shall prevent the lid, when opened, from toppling to the rear of the unit. The half-inch spacing inside the double-wall cover shall be insulated with a 1/2 inch thick Dow Chemical Corp. Ethafoam 220, or equal. The cover shall be provided with a combination corrosion-resisting steel and heat resistant plastic handle having a minimum hand grip length of 6 inches.

3.3.2.3 Base. The base of the Size 2 unit shall consist of 0.109 inch thick (No. 12 U.S.S. gage) zinc plated steel plate, welded to the inbound flanges of the outer jacket. A 0.050 inch thick (No. 16 B. & S. gage) aluminum alloy, non-dyed, anodized plate shall cover the access opening in the base plate, providing access to the electrical wiring and components on the underside of the tank. There shall be a space of 5 inches between the bottom of the tank and the base plate of the unit. Attached to this bottom plate shall be four plastic feet, 3/4 inch in diameter, by 1/2 inch high.

3.3.3 Thermostat. The unit shall be provided with a reliable, rugged, remote control thermostat designed to operate automatically by expansion of fluid in a sensing bulb, pre-set to hold the water at 160° ± 5°. The sensing bulb of this thermostat shall be located in the water of the filled tank so that the above temperature shall be maintained; and in the event the unit is operated with no water in the tank, it shall prevent damage to the unit. The thermostat shall be held in place, attached to the heating element, by a two-part press-formed 0.050 inch thick (No. 18 U.S.S. gage) corrosion-resisting steel fitting which shall readily conduct the heat from the heating element to the sensing bulb in the event that the apparatus is operated with no water in the tank (see 6.7).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

3.3.4 Pilot light. Each heater shall be equipped with an NE-51 neon lamp conspicuously mounted in an enclosed panel assembly, complete with built-in resistor suitable for the operating voltage. Shall conform to MIL-L-3661, for LH76 lampholder and LC14RN lens. The pilot light shall glow when power is applied to the unit.

3.3.5 Line Cord. Each unit shall be equipped with not less than 8 feet of 3-conductor (type HSJ, No. 18 Awg, 90° C. with neoprene, 90° C insulated individual conductors) cord; terminating in a rubber male plug conforming to ANSI C73-11. Suitable strain relief shall be provided at the connection of line-cord to unit.

3.3.6 Switch. A double pole, toggle type ON-OFF switch will be located in the back panel at the rear of the unit.

3.3.7 Heater element. When tested as specified in 4.4.2, the fully loaded tank shall reach operating temperature within 1-1/2 hours for Size 1 and 2-1/2 hours for Size 2 (see 6.8).

3.3.8 Rack. Rack shall be fabricated of 1/8 inch diameter corrosion-resisting steel wire, and shall be of spot welded construction; ground smooth, and electropolished. The rack shall rest on the bottom of the tank without contacting the heating element, and shall be designed to hold four or eight standard size steam packs, as applicable, in a vertical position so as to expose the surfaces of the packs to the hot water for maximum absorption, and to facilitate easy withdrawal for use and returning to the hot water of the tank for re-heating. The rack shall be also equipped with corrosion-resisting steel inserts to accommodate Neck Contour Packs (see 6.2), one for Size 1 and two for Size 2.

3.3.9 Heat Treatment Pad. Each heater shall be provided with heat treatment pads; four for size 1 and eight for size 2. Pads shall conform to the latest requirements for type I pads of MIL-P-37058.

3.4 Dielectric strength. The complete heater unit shall withstand the dielectric strength test specified in 4.4.1.

3.5 Operation. When tested as specified in 4.4.3, each component of the unit shall operate properly and there shall be no sign of water leakage from the unit.

3.6 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the currentversion before use.

3.7 Welding. Corrosion-resisting steel shall be welded by the heliarc method. Welding shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides, or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free from slag, excessive undercut, or excessive spatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Fusion through the full thickness of the base metal, in areas other than those where 100 percent penetration is a requirement of joint design, is not acceptable when welding materials over 0.050 inch in thickness. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks which affect the quality, and shall have more strength than required for purpose intended.

3.8 Finish. Unless otherwise specified, corrosion-resisting steel areas exposed during use shall have a Number 4 Commercial finish. Exposed welds shall be ground and polished to blend with adjacent surfaces. All welds shall be passivated.

3.9 Identification marking. Marking shall be distinct and legible. Each heater shall have a corrosion-resisting metal nameplate, firmly and securely attached in a conspicuous location, bearing the following information:

Name or registered trademark of the manufacturer.

Electrical rating (voltage, amperage, Hz, type of current, etc.)

Marking secured by adhesives shall not be used.

3.10 Repair part. A spare heating element shall be furnished with each unit.

3.11 Workmanship. Workmanship shall be first class throughout. The unit shall be free from defects which detract from its appearance or may impair its serviceability. All exposed edges and corners shall be rounded to eliminate sharp edges.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw materials may be furnished in lieu of actual performance of such testing by the contractor; provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacutrer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

TABLE I

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-3	2.5

4.2.1 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105. The acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the currentversion before use.

rmed into a rectangle, whose corners shall be roll formed, notched and to

TABLE II

Characteristic	Requirement	Test Procedure	Inspection level
Operation	3.5	4.4.3	100%
Performance	3.3.3, 3.3.7	4.4.2	S-1
Dielectric strength	3.4	4.4.1	100%

4.3 Examination. The heaters shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects, as shown in Table III.

TABLE III

Categories	Defects*
Major	
101	Any component missing, incomplete.
102	Service data missing or does not include all specified information.
103	Exposed edges not free of burs and sharpness.
104	Finish not as specified; not free of scratches, and other blemishes.
105	Components and parts not free of cracks, tears, ruptures, and deformed surfaces.
106	Components, mating parts, loose or not properly aligned and fitted.
107	Seam or fillet welds; not free of burn holes, cracks, fractures; or otherwise not fused.
108	Spot welds not sound.
109	Exposed or accessible edges, corners, and surfaces not free of burs, sharpness, and sharp projections.
110	Solder or brazing not adherent.
111	Components in the electrical system not in accordance with requirements.
112	Electrical systems in units not consistent in layout not neatly wired.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the current version before use.

TABLE III

Categories	Defects* (Cont'd)
113	Insulation of wire connectors not clear of, or protected from, framework edges to avoid occurrence of shock hazard or shorts due to vibration.
114	Electrical wiring not free of cracks, breaks, and cuts in the insulation.
115	Electrical system not free of taut leads, loose or improperly fastened terminal connections, and poor solder joints.
Minor	
201	Compartment dimensions not within specified limits.
202	Fasteners not tight.
203	Marking not present, properly located, legible, correct or permanently affixed.
204	Service data not present in duplicate.
205	Unit not clean.
206	Welds not free of slags, undercuts. Scale or flux not removed, welds not free of pitting, splatter and discoloration.
207	Soldering or brazing not clean.

*Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The unit shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests.

4.4.1 Dielectric strength. After being heated to its normal operating temperature, the heater unit shall be subjected to a gradually increasing 60 Hz potential of 1100 volts, alternating current, applied between the noncurrent-carrying metal parts and the current-carrying parts and maintained at the maximum value for a period of one minute.

4.4.2 Performance. The unit shall be fully loaded with cold water (70° F. max.) and pads (see 3.3.9), and operated from 120 volts, 60 Hz current, in accordance with the manufacturer's directions. Room temperature shall be maintained between 70° and 80° F., throughout the test. The time for the water temperature to reach 160° F. shall be recorded to determine conformance to 3.3.7. Then the temperature of the water shall be measured every ten minutes for a period of one hour to determine conformance to 3.3.3. Temperature measurements shall be made to an accuracy of ± 1/2° F. The functioning of the thermostatic control, without water in the tank, shall also be tested for conformance to 3.3.3.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:25Z
Check the source to verify that this is the current version before use.

4.4.3 Operation. The unit shall be operated from 120 volts, 60 Hz, to determine that each component is functioning properly, and that there is no leakage of water from the tank or drain.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.3).

5.1.1 Unit of issue. One heater of the type specified, containing 4 pads or 8 pads along with a spare element and service data consitutes one unit of issue.

5.1.2 Level A. Each unit shall be packaged in accordance with MIL-P-116, method III.

5.1.3 Level C. Units shall be packaged in accordance with contractors commercial practice which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.3).

5.2.1 Level A. Each unit packaged method III, shall be further packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636.

5.2.2 Level B. Each unit packaged method III, shall be further packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601; domestic type or PPP-B-636, type CF, class domestic. Fiberboard box shall conform to the special requirements as specified in Table I of PPP-B-636. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard boxes shall be closed in accordance with PPP-B-636, method II.

5.2.3 Level C. Each unit shall be packed in a substantial commercial container of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery call for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with FED-STD-123.

5.3.2 Military activities. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The heater unit and pads are intended for use in medical facilities for the application of moist-heat to patients.

6.2 Accessories. Four rectangular (standard size) immersion pads, NSN 6530-00-711-3010, are furnished with each Size 1 unit; eight pads are furnished with each Size 2 unit. Neck-contour pads are available under NSN 6530-00-865-2288; cotton covers for the neck-contour pads are available under NSN 6530-00-890-1214.

6.3 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 1.2.1).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.4 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification	Size
6530-00-711-3000	HEATER, HEAT TREATMENT PAD, Four Pad, 115 volt, 50 Hz, AC	1
6530-00-926-9063	HEATER, HEAT TREATMENT PAD, Eight Pad, 115 volt, 60 Hz, AC	2

6.5 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

6.6 Thermostat. Thermostat is available from Chattanooga Pharmacal Co.; Catalog number E202A.

6.7 Heater element. Heater element is available from Chattanooga Pharmacal Co.; Catalog number E-201.

Cusodians:

Army - MD
Navy - MS
Air Force - 13

Preparing activity:

DSA-DM

Civil Agency Coordinating Activities:

GSA-FSS
VA-DMS

Project No. 6530-1318

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See Section 2 of this specification to obtain extra copies and other documents referencend herein. Price 40 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

W-H-191B HEATER, HEAT TREATMENT PAD

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:25Z
Check the source to verify that this is the current version before use.