MIL-T-37596(DSA-DM)
28 July 1976

MILITARY SPECIFICATION

TEST STRIPS AND COLOR CHART, URINARY BILIRUBIN,

BLOOD, GLUCOSE, KETONE, NITRITE, pH, PROTEIN, AND UROBILINOGEN

This specification is approved for use by the Defense Personnel Support Center, Defense Supply Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Test Strips and Color Chart, Urinary Bilirubin, Blood, Glucose, Ketone, Nitrite, pH, Protein and Urobilinogen (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-C-186 - Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals. (See 6.3)

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:05Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be test strips, each strip approximately 1/5 inch (5.08 mm) wide by 3-1/4 inches (82.6 mm) long with eight separate and distinctly divided bibulous paper areas which are impregnated with chemicals, enzymes or mixtures thereof, in each such area, which produce sensitive and reliable qualitative colorimetric determination of glucosuria, proteinuria, urinary pH, ketone in urine, bilirubin in urine, urobilinogen in urine, occult blood and nitrite in urine or the specific absence of glucose, protein, urinary ketones, bilirubin, urobilinogen, occult blood and nitrite when moistened with urine.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:05Z
Check the source to verify that this is the current version before use.

3.1.1 Each impregnated area shall be separated so that the dye of the test areas shall not bleed across the area spacing.

3.2 Shall be suitable for use as a rapid dip and read test for urinary glucose, blood, ketone, protein, bilirubin, urobilinogen, pH and nitrite.

3.3 Impregnation of test areas. The separate test areas shall be impregnated with specific composition for each test, to produce the anticipated color response charted in the incremented table provided on the label of the immediate container. The quality of color which is developed shall be evenly distributed and not spotty to the degree that the reader cannot interpret test results.

3.3.1 The strips shall be tested as specified in 4.2.

3.4 Expiration dating period. The expiration dating period shall be not less than 18 months.

3.4.1 Delivery. Not more than three (3) months of the expiration dating period shall have elapsed at the time of delivery to the Government.

3.5 Workmanship. The material and its containers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:05Z
Check the source to verify that this is the current version before use.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Tests. The strips shall be tested by appropriate, accurate, and reproducible method(s) as used for commercial release. A copy of the test results shall be submitted to Defense Personnel Support Center, Directorate of Medical Materiel, Attention: Quality Assurance Branch, DPSC-ATQD, 2800 South 20th Street, Philadelphia, PA 19101.

4.3 Packaging inspection. The inspection of the intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and the unit of product shall be the unit of issue. The inspection lot shall be the number of units of issue of each batch, lot, control number, etc., offered for inspection at one time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

5. Packaging.

5.1 Packaging requirements. The requirements for packaging shall be in accordance with all applicable requirements of PPP-C-186 (see 6.3) and as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.2.1 Unit of issue. One bottle containing 100 strips, as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate container. The immediate container shall be a commercially available tightly closed immediate container, with dessicant, considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.2 Unit package. At the option of the contractor, a unit package may be furnished. If a unit package is furnished, each bottle shall be packaged as specified in 5.2.5 of PPP-C-186.

5.2.2.3 Intermediate package. Twenty-four units of issue shall be packaged as specified in 5.2.7.1 of PPP-C-186.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:05Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.3.1 Level A. Six intermediate packages (144 units) shall be packed in an exterior container as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Six intermediate packages shall be packed in an exterior container as specified in 5.4.3.1.1 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act, and shall include the information required below.

5.4.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the expiration date.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the National Stock Number.

NOTES: (1) The national stock number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Labeling adhesion test shall not apply.
(4) If a unit package is not supplied, the National Stock Number shall appear on the immediate container label.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:05Z
Check the source to verify that this is the current version before use.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type I Shelf Life Marking as specified in MIL-STD-129 shall be shown. Marking shall include the legend: "STORE AT CONTROLLED ROOM TEMPERATURE 15°-30°C(59°-86°F)".

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.4 of PPP-C-186. Type I Shelf Life Marking as specified in MIL-STD-129 shall be shown: Life Marking as specified in MIL-STD-129 shall be shown: "STORE AT CONTROLLED ROOM TEMPERATURE 15°-30°C(59°-86°F)".

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock Number	Item Identification
6505-01-022-2401	TEST STRIPS AND COLOR CHART, URINARY BILIRUBIN, BLOOD, GLUCOSE, KETONE, NITRITE, pH, PROTEIN AND UROBILINOGEN, 100s

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superseded by a revision.

Preparing activity:
Defense Supply Agency - DM

Project No. 6505-1741

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:05Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE MIL-T-37596(DSA-DM) TEST STRIPS AND COLOR CHART, URINARY BILIRUBIN, BLOOD, GLUCOSE, KETONE, NITRITE, pH, PROTEIN, AND URO-BILINOGEN

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:05Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20th STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:05Z
Check the source to verify that this is the current version before use.