A-A-51264B
30 May 1989
SUPERSEDING
A-A-51264A
30 April 1986

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, GOWN AND TOWEL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one size of a sterile, disposable, reinforced, operating gown with towel, packaged in two different quantities, suitable for preventing bacterial migration into the sterile field.

Salient characteristics.

The surgical operating gown with towel shall be sterile, large size, disposable, and nonwoven.

The gown shall conform to the following minimum dimensions in inches:

LOCATION	DIMENSION (MINIMUM)
Chest	53
Length, Front (Center)	42
Sweep	60
Sleeve (measured from neck to cuff)	30
Belt Length	15

The surgical operating gowns shall be full length with a generous overlapping back closed with plastic snap fasteners or hook and pile fastener at the neck area. The snap fasteners shall be applied to a nonwoven ribbon made from the body material. The ribbon shall be sewn to the gown; three (3) male snaps on the left flap and two (2) female snaps on the right flap. The gown shall have four (4) tie tapes; two (2) on the outside and two (2) on the inside, approximately at waist level. The tie tapes shall be attached to the gown. The gown shall have a nonwoven neck binding sewn to neck of gown body.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

The gown shall have full length raglan sleeves, or full length set-in sleeves attached by sewing. The sleeves shall have wristlets made of tubular knit polyester fabric.

The gown shall be fabricated from a soft, pliable, fluid resistant nonwoven material. The material shall have good resistance to dry-particle penetration. The material shall have high breathability, and shall be comfortable to wear for long periods of time. The color of the material shall be a light pastel shade of blue, green or blue-green. The gown shall meet the following requirements:

Hydrohead 35 cm minimum
Spray impact 3 grams maximum
Bacterial filtration efficiency (BFE) 90 percent

Each gown shall be supplied with a scrim re-enforced, nonwoven absorbent towel. The towel shall have minimum dimensions of 15 by 19-1/2 inches.

The gowns and towels shall be folded and overwrapped in a white nonwoven or plastic fabric. The folded and wrapped gowns and towels shall be packaged in a white Kraft heat-sealed bag which is tear resistant and which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

The surgical gowns and towels in their immediate container shall be sterile.

Workmanship. The operating gown with towel shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 32 large gowns with towels for NSN 6532-01-187-7768 or 28 large gowns with towels for NSN 6532-01-146-4269, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:53Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one (1) gown with towel as specified in this description.

Defects.*

Any component missing, broken, or not as specified.

Any hole, cut, tear, or perforation.

Gown and/or towel not free of grease, dirt or foreign material.

Snap fasteners, hook and pile fasteners missing, not properly attached or otherwise not useable.

Any spot, stain, or streak.

Wrong color.

*Examination shall not be limited to the defects listed above.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one operating gown with towel. Unless otherwise specified, sampling for tests shall be conducted at Inspection Level S-1, with an AQL of 2.5 (percent defective).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

6532-01-146-4269
6532-01-187-7768

Ordering data. Intermediate/exterior quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

Project No. 6532-1188
Location:
CID/SURGICAL/GOWN/H15-3

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.