A-A-51265
9 December 1985

COMMERCIAL ITEM DESCRIPTION

STONE, SHARPENING

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification SS-S-736C, Type VI.

This commercial item description covers a natural (novaculite) grooved sharpening stone.

Salient characteristics:

The sharpening stone shall be a natural (novaculite) grooved stone suitable for use in sharpening the cutting edges of dental instruments. The stone shall be furnished in hard (slow-cutting) grit and shall not be impregnated with oil. Four (4) machined, parallel, longitudinal grooves shall be spaced uniformly across the width of the stone. Sharpening stones shall possess sufficient strength to withstand normal service without chipping, breaking, or otherwise becoming damaged. The stone shall be in accordance with figure 1.

Workmanship. The sharpening stone shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One sharpening stone, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

DIMENSIONS						
Length ± 3/16	Width ± 1/16	Thickness ± 1/16	Diameter of 4 Grooves (Millimeters and Inches)			
Inches 4	Inches 1-1/2	Inches 1/2	+25%-0 1 mm (0.0394)	+20%-0 2 mm (0.0787)	± 10% 4 mm (0.1575)	± 5% 8 mm (0.3150)




FIGURE 1

Stone, Sharpening

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for test purposes shall be one sharpening stone. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Examination. Samples shall be examined for breaks or chips and for defects affecting function or serviceability.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-00-536-2150 STONE, SHARPENING, Dental Instrument, 4 by 1-1/2 by 1/2 inches

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

A-A-51265

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6520-1798

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.