A-A-51267A
19 September 1986

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

The General Services Administration has authorized the use of this commercial item description in preference to A-A-0051267(DM), dated 9 December 1985.

This commercial item description covers a sterile, disposable, individually sealed, delivery room surgical pack for use in vaginal obstetrical procedures and newborn infant care.

Salient characteristics:

The delivery room drape pack shall be sterile, disposable, and fabricated of nonwoven material for creating a sterile field specifically for vaginal obstetrical procedures and for care of the delivered infant. The pack shall consist of at least the following components in the listed approximate dimensions:

a. One overwrap/tablecover, moisture-resistant, 44 by 78 inches
b. One set-up cover, moisture-resistant, 44 by 75 inches
c. One absorbent towel, 14 by 24 inches or 19-3/4 by 20 inches
d. Two utility drapes, moisture-resistant, 14 by 24 inches (optional)
e. One under-buttock drape, moisture-resistant, 36 by 56 inches with a 10 inch cuff and impervious fabric reinforcement
f. Two leggings, 29 by 48 inches with 6 inch cuffs
g. One abdominal drape, moisture-resistant, 31 by 57 inches with an impervious fabric reinforcement.
h. Two sanitary napkins, absorbent, 3 by 10-1/2 inches
i. Ten sponges, radiopaque, 12-ply absorbent gauze, 4 by 4 inch
j. One gown, obstetrical-surgical, moisture-resistant
k. One placenta basin, rigid plastic, 8-3/4 by 12-1/4 by 3-1/2 inches
l. One syringe, bulb
m. One infant blanket, 30 by 30 inches
n. One umbilical cord clamp, plastic
o. One vaginal packing sponge, absorbent, radiopaque, 4-1/2 by 18 inches

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

The overwrap/tablecover shall be fan-folded for ease in opening on the delivery room back table. It shall be impervious to fluids and moisture.

The surgical pack shall have a set-up cover that is fan-folded and water repellent.

The towels shall be highly absorbent to sufficiently dry arms after a surgical scrub.

The utility drapes (optional) shall be water repellent.

The surgical pack shall have a fan-folded drape to place under the buttocks. The drape shall have a 10 inch cuff for aseptic draping. The drape shall have a reinforced area, 18 by 32 inches or 18 by 24 inches. The reinforcement area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids.

The leggings shall be cuffed and fan-folded for easy and aseptic application.

The surgical pack shall have a water repellent abdominal drape with a 12 by 18 inch reinforced area. The reinforcement area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids.

The surgical pack shall contain two highly absorbent sanitary napkins, approximately 3 inches wide by 10-1/2 inches long.

The 4 by 4 inch radiopaque sponges (optional) shall be 12-ply, USP, Type VII, and be fabricated of an absorbent gauze material.

The surgical gown shall be large or extra large size. The gown shall completely cover the back of the individual gowned. It shall be water repellent and shall be of a soft, comfortable, memory-free material with stockinette cuffs that conform to wrists. The gown shall have a snap or tie neck closure.

The surgical pack shall contain a 2 or 3 oz ear and ulcer syringe for aspiration of fluid and mucous from the mouth of the newborn. Suction shall be 120/140 mm mercury.

The surgical pack shall contain an infant blanket fabricated of a cotton flannel or a soft polyester material, approximately 30 by 30 inch (76.2 by 76.2 cm). all edges shall be finished on all sides.

A plastic umbilical cord clamp shall be included in the surgical pack.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

The pack shall contain one radiopaque vaginal packing sponge, approximately 4-1/2 by 18 inches (11.4 by 45.7 cm). It shall have at least a 6 inch (15.2 cm) braided textile loop to maintain visibility of the pack.

All nonwoven items shall be fabricated of a soft material with minimal memory for good drapeability. The fabric shall resist tearing,. pilling and linting.

The pack shall be sterile and individually wrapped in a suitable tear resistant, heat sealed package which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

Each pack shall also be wrapped in a heat sealed exterior dust cover.

All pack components shall be clearly listed on the exterior wrapper.

Workmanship. The surgical packs shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing four (4) or five (5) individually sealed packs, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one (1) pack. Unless otherwise specified, sampling for examination shall be conducted at Inspection Level II, AQL 2.5, and sampling for tests shall be conducted at Inspection Level S-1, acceptance number is zero.

Sterility test. The surgical pack shall be tested for sterility in accordance with the method specified in the U.S. Pharmacopeial Convention Publication: Pharmacopeia of the United States, Mack Publishing Co., Easton, PA 18042 (USP), for purified cotton, gauze, surgical dressings and related material using inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as listed below. Product containing a biological indicator shall be marked or labeled, so that it clearly differentiates from product intended for product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20. For sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature Degrees C
Steam	B.stearothermophilus	55 - 65
Dry Heat	B.stearothermophilus	55 - 65
	B.subtilis var. niger (or globigii)	30 - 35
Gas	B.subtilis var. niger (or globigii)	30 - 35
Radiation	B. pumilus	30 - 35

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method, incubation and number of test units shall be as indicated herein for inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Alternate method. Suppliers conducting sterility tests in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a Government contract. "Dosimetric Release" as approved by FDA for gamma radiation sterilization is acceptable without prior approval by the contracting officer.

Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the USP.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE:
1) The following National Stock Number is covered by this document:

6530-01-146-4262

2) For purposes of bid evaluation each bidder shall submit their quotation based on a single surgical pack.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 59719

Project No. 6530-1915

Location: CID/SURG-PACK-DISPOS-DEL-RM/H15

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.