COMMERCIAL ITEM DESCRIPTION

RING, LABORATORY APPARATUS SUPPORT

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-R-370.

This commercial item description covers one type of laboratory apparatus support ring.

Salient characteristics:

The laboratory apparatus support ring shall be a closed metal ring suitable for use with laboratory apparatus support stands having rods up to 1/2 inch in diameter.

Shall have the following dimensions as listed in Table I.

TABLE I

Size	Outer Diameter* (inches)	Inner Diameter* (inches)
1	3	2-1/2
2	4	3-1/2
3	5	4-3/8

*Tolerance, ± 1/32 inch.

Shall be fabricated of a suitable metal material that will withstand laboratory procedures. The finish shall be in gray or black Japan enamel. Shall be provided with an integral clamping device consisting of 2 planar surfaces at right angles to each other to accommodate a 1-1/2 inch long thumb screw. The thumb screws shall be fabricated of a corrosion-resistant metal. Shall be designed so that different diameter rings shall be concentric when clamped to the same support rod.

The distance from the center of the ring to the center of the support rod shall be a nominal 4-1/2 inches. Support rings shall have no sharp edges or casting defects. The support rings shall be level ±3° when attached to a horizontal support rod.

The inner periphery of the ring having a circular cross section may be provided with three equally spaced projections. If rectangular, the top inner peripheral edge shall be beveled to hold round bottom objects.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.
Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Workmanship. The support ring shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One ring, of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-2, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-440-1315	RING, LABORATORY APPARATUS SUPPORT, Size 1
6640-00-440-1325	RING, LABORATORY APPARATUS SUPPORT, Size 2
6640-00-440-1340	RING, LABORATORY APPARATUS SUPPORT, Size 3

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1343

Location S3-8-12

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 — 606-036/45071

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.