A-A-51277
21 January 1986

COMMERCIAL ITEM DESCRIPTION

TEST KIT, ENTEROBACTERIACEAE DIFFERENTIATION

This Commercial Item Description covers a miniaturization of standard biochemical determinations used in the identification of Enterobacteriaceae and other gram-negative bacteria.

Salient characteristics:

Shall be a miniaturization of standard biochemical determinations containing dehydrated media, for identification of Enterobacteriaceae and other gram-negative bacteria from a single colony of organisms suspended in a saline solution. The dehydrated media shall be reconstituted by adding a bacterial suspension.

Shall be in the form of a test strip that holds a series of micro tubes/wells, containing 20 different dehydrated differential media for performing the following 23 biomedical tests:

1. Beta-galactosidase (ONPG disc)
2. Citrate reductions
3. Glucose fermentation
4. Mannitol oxidation
5. Inositol oxidation
6. Sorbitol oxidation
7. Rhamnose oxidation
8. Sucrose oxidation
9. Melibiose oxidation
10. Amygdalin oxidation
11. Arabinose oxidation
12. Motility
13. Gelatin liquification
14. H_2S production
15. Indole formation (Kovac's reagent)
16. Arginine Dehydrolase
17. Lysine Decarboxylase
18. Ornithine Decarboxylase
19. Tryptophane Deaminase
20. Urea reduction
21. Catalase (H_2O_2)
22. Nitrate reduction
23. Voges-Proskauer

Shall be supplied with an incubation tray with a lid.

Shall be supplied with instructions, a differential chart, and report sheet(s) to assist in the identification of organisms.

After a test strip has been prepared and innoculated in accordance with the manufacturer's instructions, organism identification shall be facilitated by using the differential chart, or a reference manual that contains an interpretation scheme for use in identifying bacteria of the family Enterobacteriaceae and other clinically significant gram-negative in an 18-24 identification procedure.

Not less than 16 months of the expiration dating period shall remain at the time of delivery to the Government.

Workmanship. The Enterobacteriaceae differentiation test kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One kit consisting of 25 sets of media strips together with 25 incubation trays/covers, 25 report sheet instructions and a differential chart, as specified, constitutes one unit.

FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Testing. Tests shall be conducted on the media strips, to check the positive and negative biochemical reactions of the media to different viable cultures that include the following:

1. Klebsiella pneumoniae (ATCC 13883)
2. Enterobacter cloacae (ATCC 13047)
3. Proteus vulgaris (ATCC 13315)
4. Pseudomonas aeruginosa (ATCC 10145)

The tests shall be conducted in accordance with manufacturer's instructions. The biochemical reactions shall be recorded and compared with a suitable reference for expected biochemical reactions.

The sampling for testing shall be in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6550-01-046-2912

The following National Stock Number covers a reference manual for identifying bacteria of the family Enterobacteriaceae and other clinically significant gram-negative bacteria.

6550-01-167-2617

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 53719

Project No. 6550-0035

Location S3-8-9

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.