A-A-51280A
17 June 1988
SUPERSEDING
A-A-51280
9 December 1985

COMMERCIAL ITEM DESCRIPTION

NEUROSURGICAL HEADREST SYSTEM

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a Neurosurgical Headrest System suitable for attachment to a field operating table (See Notes).

Salient characteristics.

The neurosurgical headrest system shall be of the Mayfield design and shall consist of the following components (See Notes).

Neurosurgical Headrest Attachment Unit
General Purpose Neurosurgical Headrest
Horseshoe Headrest
Crossbar Attachment
Skull Clamp
Transitional Member
Skull Clamp Adapter
Table Adapter
Skull Clamp Pin
Locknut, 3/8" in Diameter

Attachment unit. Shall be adjustable and shall incorporate a single locking lever that shall provide firm positioning and eliminate troublesome set screws and wing nuts. Shall allow the head to be quickly repositioned laterally or vertically by simply releasing the locking lever. The tension on the lever shall be easily adjustable. The attachment unit shall be used with the general purpose headrest, the horseshoe and crossbar attachment.

AMSC N/A FSC 6530
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

General purpose headrest. Shall be of Mayfield design and shall attach to the attachment unit to be used in procedures which require the supine position. Floating rubber cups mounted on a half moon frame shall conform to the shape of the head. The caterpillar tread action of the cups shall provide firm fixation. The more firmly the head is placed against the cups, the more securely it shall be locked into position. The head shall be repositioned up or down during the procedure by simply releasing the locking lever or the attachment unit and changing the position.

Horseshoe headrest. Shall be of Mayfield design for use in all prone position procedures. The horseshoe headrest shall attach to the attachment unit by means of a transitional member and Gardner clamp adapter. Shall be padded with soft, pliable sponge rubber for equalized weight distribution and incorporate quick change pads. The horseshoe width shall be quickly adjusted to accomodate various head sizes.

Crossbar attachment. Shall be of Mayfield design and used when the patient is in the sitting position for procedures such as Posterior Fossa Craniotomy or a Posterior Cervical Laminectomy. The crossbar shall attach to the operation table and the Neurosurgical Attachment Unit shall attach to the crossbar. Shall be supplied with a Gardner skull clamp adapter to facilitate the attachment of the skull clamp.

Skull clamp. Shall provide three-point pin fixation to enable rigid skeletal fixation for supine prone, lateral, or sitting positions during neurosurgical procedures. Shall have a mechanism to allow insertion of three removable cone shaped corrosion-resisting pins to adult sizes.

Width of the skull clamp. Shall be adjustable by means of a ratchet with a quick release knob. Shall have a large calibrated adjusting knob to permit precise positioning of pins and selection of applied pressure. The skull clamp shall be supplied with three adult cone-shaped pins. The clamp shall be capable of withstanding pressure and temperatures required for sterilization.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Transitional member. Shall be a three-inch member to be used for various attachments to the operating room table. One end of the transitional member shall be designed to be inserted into the top section of the attachment unit . The other end shall have a sunburst connection designed to mate with various Mayfield attachments.

Skull clamp adapter. Shall be of Gardner design. Shall fit the Mayfield transitional member to allow adaptation of all Mayfield skull clamps.

Shampaine table adapter. Shall enable the attachment unit to adapt to the shampaine table.

Skull clamp pin. Shall be of the Mayfield design.

3/8" diameter locknut. Shall be suitable for use with the headrest attachment unit.

Except for headrest padding, all of the above items shall be fabricated of a suitable corrosion-resisting steel, properly hardened and passivated.

This system shall be supplied with a storage board, designed to hold all parts of the Mayfield headrest and skull clamp system described above.

Workmanship. The headrest system shall be free from from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One headrest system, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6530-01-208-2408	HEADREST NEUROSURGICAL SYSTEM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Note 2: The following components are compatible with the Neurosurgical System covered in this specification:

NSN	ITEM IDENTIFICATION
6530-01-139-9439	ATTACHMENT UNIT, NEUROSURGICAL HEADREST
6530-01-139-9441	HEADREST, NEUROSURGICAL, GENERAL PURPOSE
6530-01-139-9440	HEADREST, NEUROSURGICAL, HORSESHOE
6530-01-139-9442	CROSSBAR ATTACHMENT, NEUROSURGICAL
6530-01-139-7858	CLAMP, SKULL, NEUROSURGICAL
6530-01-172-1116	TRANSITIONAL MEMBER, NEUROSURGICAL HEADREST
6530-01-166-7768	ADAPTER, SKULL CLAMP, GARDNER
6530-01-169-3869	TABLE ADAPTER, SHAMPAINE
6530-01-166-0152	SKULL CLAMP PIN
6530-01-166-3511	LOCKNUT, 3/8" DIAMETER
6530-00-142-9239	TABLE, OPERATING, FIELD

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQASs

Project No. 6530-2007

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-PSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51280A	NEUROSURGICAL HEADREST SYSTEM

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City, State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM CUT ALONG THIS LINE.)

Source: https://assist.dla.mil – Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.