A-A-51284
16 December 1985

COMMERCIAL ITEM DESCRIPTION

WASH BOTTLE, LABORATORY, (Polyethylene)

The General Services Administration has authorized the use of this commercial item description in preference to Type II, Styles A, B and C Wash Bottle of Federal Specification NNN-W-25B.

This commercial item description covers a plastic wash bottle for laboratory use.

Salient characteristics:

The laboratory wash bottle shall conform to one of the following styles and capacities:

Style A. The style A wash bottle shall be constructed with the outlet tube passing through the center of the screwcap closure.

The bottle and screwcap liner (if required) shall be fabricated of polyethylene. The screwcap and outlet tube shall be fabricated of polyethylene or polypropylene or polyallomer.

The bottle shall be wide- or narrow-mouthed, cylindrical, and of a diameter suitable for convenient laboratory handling. The mouth of the bottle shall be threaded to receive a screwcap. The wall of the bottle shall be of sufficient thickness to provide the required strength and durability for general laboratory use.

The outlet tube, with the screwcap in place on the bottle, shall extend from within 5 mm. of the bottom of the bottle upward through the center of the liner and the cap for a convenient distance, and then bend outward and downward from the vertical axis to form an angle which is less than 90°. The outlet tube shall terminate in a drawn-out tip of 2-3 centimeters (cm.), suitable for producing a small diameter stream. The wall of the outlet tube shall be of sufficient thickness to provide the necessary strength and durability required for general laboratory use. The inside diameter of the outlet tube shall be such as will allow a steady, controllable stream of liquid to be dispensed. The screwcap shall be properly threaded to mate and form a tight fit with the mouth of the bottle. Screwcap may be furnished with a liner, but liner is not required.

As an alternate, the screwcap and the external portion of the outlet tube, may be molded as one integral piece with the internal portion of the outlet tube as a second piece inserted into the screwcap to form a continuous passage.

Style B. The style B wash bottle shall be constructed with a removable side outlet tube passing through a boss in the shoulder of the bottle. The bottle shall have a screwcap closure.

The bottle, boss, and outlet tube shall be fabricated of polyethylene. The screwcap shall be fabricated of polypropylene.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

The bottle shall be narrow-mouthed, cylindrical, and of a diameter suitable for convenient laboratory handling. The mouth of the bottle shall be threaded to receive the screwcap. The wall of the bottle shall be of sufficient thickness to provide the required strength and durability during general laboratory use.

The boss shall be integral with the bottle or shall form a tight friction fit with the hole in the shoulder of the bottle. A uniform hole running the entire length of the boss shall be provided to allow insertion of the outlet tube through the boss.

The outlet tube shall be flexible throughout its entire length. The tube shall extend from within 5 mm. of the bottom of the bottle upward through and above the boss for a convenient distance, and then bend outward and downward from the vertical axis to form an angle which is less than 90°. The tube shall terminate in a drawn-out tip of 2-3 cm., suitable for producing a small diameter stream. The wall of the outlet tube shall be of sufficient thickness to provide the necessary strength and durability required for general laboratory use. The inside diameter of the outlet tube shall be such as will allow a steady, controllable stream of liquid to be dispensed.

The screwcap shall be properly threaded to mate and form a tight fit with the mouth of the bottle.

Style C. The style C wash bottle shall be constructed with an integral side outlet tube. The bottle shall have a screwcap closure.

The bottle and outlet tube shall be fabricated of polyethylene. the screwcap shall be fabricated of polypropylene.

The bottle shall be wide-or narrow-mouthed, cylindrical, and of a diameter suitable for convenient laboratory handling. The wall of the bottle shall be of sufficient thickness to provide the required strength and durability during general laboratory use. The mouth of the bottle shall be threaded to receive the screwcap.

The outlet tube shall be molded in and shall be integral with the bottle. The tube shall open into the side of the bottle just above the junction of the side and the bottom. A portion of the tube outside the bottle shall be arced, and of curvature such that the top of the arc is below the top of the mouth of the bottle, and the end of the arc forms an angle of $135^{\circ} \pm 5^{\circ}$ with the vertical axis. The tube shall terminate in a straight, rounded-off tip, or a tapered cone tip. A small orifice at the end of the tip shall allow for dispensing of a small diameter stream of liquid.

The screwcap shall be properly threaded to mate and form a tight fit with the mouth of the bottle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

The fit of the wash bottle component parts shall be such that there is no leakage when the wash bottle is used in various positions including the upright, horizontal and inverted positions.

The wash bottles shall be provided in 250, 500, and 1000 milliliter (ml) capacities.

Workmanship. The wash bottle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One wash bottle, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one wash bottle.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-299-8493	250 ml
6640-00-314-2097	500 ml
6640-00-855-8762	1000 ml

Ordering data; (Intermediate/exterior package quantities, and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1314
Locarion S3-8-26

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/45204

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.