A-A-51286
18 December 1985

COMMERCIAL ITEM DESCRIPTION

WEDGE SET, DENTAL, (Wood)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-W-36421B.

This commercial item description covers a wedge set suitable for use in dental operative procedures.

Salient characteristics:

Wedges shall be suitable for use in exerting a lateral pressure and enforcing a separation of teeth between the proximal surfaces of adjoining teeth. Wedges shall be fabricated of high quality, seasoned or kiln dried white birchwood. The wedge set shall consist of an assortment of not less than 500 wedges. Wedges shall be of triangular cross section and shall be longitudinally tapered to a point or slightly blunted. All surfaces of the wedges shall be clean, smoothly finished, and free from splinters. Wedges shall be furnished in the shapes or sizes, and quantities listed in Table I. Wedges shall also conform to Figure 1.

TABLE I. Shapes or sizes, and quantities.

Shape or size	Quantity (minimum)1/
Small	200
Large	75
Wide base	50
Slim Jim	150

1/ The remaining 25 wedges may be supplied in any one size or assortment of sizes.

Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended. Wedges shall be furnished clean and ready for use.


Workmanship. Wedge set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Box (BX). One box containing 500 wedges, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.



DIMENSIONS				
SHAPE OR SIZE	H (±1/2)	L	ANGLE a	ANGLE b
SMALL	3 1/2	11 3/4 - 12 1/2	29 - 33	11 1/2 - 13 1/2
LARGE	4	14 - 16	29 - 33	11 1/2 - 13 1/2
WIDE BASE	4	14 - 16	47 - 51	10 1/2 - 14 1/2
SLIM JIM	3	14 1/2 - 16	47 - 51	8 - 10

NOTE:
DIMS. H AND L ARE IN MILLIMETERS.
ANGLES a AND b ARE IN DEGREES.

FIGURE 1. Wedge set, dental, wood.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one box of wedges. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-00-584-3333	WEDGE SET, DENTAL, Wood, Set of 500

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

Project No. 6520-1873

48245

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.