A-A-51289
18 December 1985

COMMERCIAL ITEM DESCRIPTION

MAGNIFIER, (Dental Laboratory)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers headband-mounted loupes providing binocular magnification.

Salient characteristics:

The magnifier shall be headband-mounted loupes providing binocular magnification, so designed as to permit magnification of work leaving both hands free and the wearing of ordinary eyeglasses and bifocals.

The magnifier shall have a continuous headband adjustable to head size by a knurled knob at back of head. Clockwise turns of the knob shall make the headband smaller and counter-clockwise turns shall make it larger. A comfort band shall be fastened to the front portion of the headband with four equally spaced finger snaps. The male portion of the snaps shall be permanently mounted on the outside of the headband and the female portion shall be permanently mounted on the outside portion of the comfort band. The comfort band shall be cushioned with a strip of suitable batting material permanently glued to inside fold of the band. The visor portion of the magnifier shall be fabricated in one piece from lightweight, durable, impact-resistant material and shall afford protective shielding to the right, left, and over the eyes. The visor arms extending past the side shielding shall be attached to the headband on each side with an adjustable pivot assembly. The raising and lowering of the visor shall be controlled by two tension thumb nuts at the pivot points. The lens plate shall be a continuous piece and shall be attached to the front of the visor with two round head, slotted screws with appropriate nuts. The lens plate and the front portion of visor shall afford protection from the direction of the work. For design purposes, the magnifier shall be as shown in Figure 1.

The lens plate shall be equipped with two 4 diopter optical glass lenses, each with a focal length of approximately 10 inches providing a magnification of 2X. The lenses shall be securely retained by the lens plate.

Dimensions of the magnifier shall be as follows: Minimum inside circumference of headband shall be 20-1/2 ± 1/8 inches. The total adjustment for size shall be 4 ± 1/8 inches. The width of the headband, excluding knurled knob portion, shall be 1-1/8 ± 1/16 inches. The lens plate shall be rectangularly shaped, 4 x 1-1/2 ± 1/32 inches, with rounded corners, shal have two openings for lenses, and shall have two screw holes 3-1/8 ± 1/32 inches apart for fastenings. The thickness shall be stepped at screw holes for fitting into and onto the visor. The lenses shall be rectangularly shaped with rounded corners and shall be 1-1/2 x 1-1/8 ± 1/32 inches, excluding bevel. The comfort band shall be 11-1/2 ± 1/4 inches long and shall have sufficient inside width to cover the headband.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.




Figure 1. Magnifier, Dental Laboratory

The magnifier shown is typical in design and is not intended to restrict or preclude the purchase of magnifiers which are otherwise in accordance with the requirements of this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Material. The headband, visor, and lens plate frames shall be fabricated from plastic materials suitable for the purposes intended. The knurled knob, pivot assemblies, finger snaps, and other hardware shall be fabricated from corrosion-resisting metals or metals treated for corrosion resistance. The lenses shall be fabricated from suitable optical glass. The comfort band shall be fabricated from leather suitably surface treated and suitable for the purpose intended.

Suitable instructions on care and cleaning of loupes shall be included with each unit of issue.

Workmanship. The magnifier shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One magnifier, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the rdetermination of two or more inspection characteristics. The unit of product for inspection purposes shall be one magnifier. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-01-082-4897 MAGNIFIER, DENTAL, LABORATORY, Headband-Mounted

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

49052

Project No. 6520-1892

Location: S2-12-15

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.