A-A-51293
12 December, 1985

COMMERCIAL ITEM DESCRIPTION

CART, MEDICINE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a corrosion-resisting steel medicine cart designed to provide a complete unit with all accessories necessary for use by nursing personnel in distributing and administering medication, oral and parenteral, to patients in military hospitals.

Salient characteristics:

Shall be a medicine carriage designed and constructed in such a manner that it will provide a complete unit with all accessories necessary for use by nursing personnel in preparing, distributing, and administering medication, oral and parenteral, to patients in wards of military hospitals.

Shall be constructed of approximately 20 gauge corrosion-resisting steel rthroughout, with suitable corrosion-resisting hardware.

Shall consist essentially of a carriage mounted on 4 tubular uprights or legs on top of which is included a flat table-type top and appropriate encircling apron and a removable or hinged stepped-up medicine tray with 3 or 4 steps, each step with circular cutout openings, to accommodate standard 1 oz. medicine glasses. Each riser shall contain holders to hold securely an identification card for each medicine glass behind it. The tray shall hold at least 24 medicine glasses.

The carriage shall be approximately 36 inches high from the floor to the flat table top; the height of the carriage from the floor to the top step of the stepped-up medicine tray shall be not more than 44 inches. The length of the table top shall be not less than 20 inches, the width not less than 16 inches and the tubular uprights or legs shall be approximately 1 inch outside diameter or 1 inch square. There shall be a guard rail or fence extending above the flat table top along at least the two ends and across the back.

A rectangular stainless steel tray with cover shall be provided. This tray shall be approximately 9 inches long, 5 inches wide, and 2 inches deep. The cover shall be provided with an appropriate handle. The tray shall be suitable for the purpose of disposing of used hypodermic syringes after injection.

There shall be provided a stainless steel pitcher with hinged cover and handle. The pitcher shall have a capacity of approximately 2 quarts. The pitcher shall be provided with a suitable pouring lip.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

There shall be a drawer located immediately beneath the flat table top. The depth (front-to-back) of the drawer shall be commensurate with the depth of table top; the length of the drawer shall be not less than 16 inches. There shall be an adequate strap handle provided on the front of the drawer.

There shall be a syringe rack of corrosion-resisting steel or suitable durable plastic, designed to fit snugly and securely on the bottom of the drawer.

The rack shall be constructed to hold at least 12 standard 5 cc. and 2 cc. hypodermic syringes and attached needles securely in an acceptable aseptic fashion. The rack shall be so fabricated that its thorough cleaning will not be a difficult or impossible operation. The rack shall also be equipped with individual clip holders to accommodate patient identification cards. The rack shall be equipped with an appropriate handle or other lifting device so that it may be easily and quickly removed or inserted.

The drawer shall be so designed and constructed that when closed and with the carriage in motion, it will not inadvertently open. The drawer shall be provided with a suitable type of locking device to eliminate the possibility of the drawer being completely pulled out by the user; the drawer, however, should be so constructed that it can be removed from the carriage for cleaning.

The carriage shall be equipped with a lower shelf. The lower shalf shall include an adequate encircling apron.

There shall be a lower drawer included immediately beneath the lower shelf, similar to the upper drawer. The height of the lower drawer shall be approximately 6 inches and there shall be no accommodations for the syringe rack.

There shall be a guard rail located approximately 3 inches above the lower shelf completely encircling the carriage.

The entire carriage shall be mounted on suitable ball-bearing, swivel-type casters, each caster to be at least 4 inches in diameter.

Shall be so constructed that all corners shall be rounded and edges smooth so that there will be no danger of cutting the hands of the operator by sharp cornerss, or tearing of clothing of an operator who may inadvertently brush against the item in passing.

Workmanship. The medicine cart shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cart, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-774-1100

Ordering data. Intermediate/Exterior package Quantities and labeling and marking must be specified in the and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

A-A-51293

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 50372

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6530-1816

Location H1-17-13

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.