A-A-51295
12 December, 1985

COMMERCIAL ITEM DESCRIPTION

TUBING, STERILIZATION

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-T-37027.

This Commercial Item Description covers heat sealable polyethylene tubing used for packaging medical supplies for gas sterilization and storage in military medical facilities.

Salient characteristics:

Shall be extruded, low density, lay flat, transparent polyethylene tubing suitable for packaging medical supplies for ethylene oxide gas sterilization and storage. The tubing shall be supplied in four sizes, each 75 feet in length.

Size 1 - 3 inches wide
Size 2 - 5 inches wide
Size 3 - 10 inches wide
Size 4 - 20 inches wide

Width tolerance shall be ± 1/8 inch for size 1 and 2 and ± 1/4 inch for size 3 and 4.

The low density polyethylene film shall be extruded from Sinclair-Koppers Dylan 1000 F resin, Gulf Inc's 4524, or equal.

The polyethylene film shall have a minimum ethylene oxide gas permeability of 100,000 cc per mil thickness, per square meter, per 24 hours, at 70 degrees F, and 29.92 inches of hq atmospheric pressure.

The polyethylene film shall be treated with an internal antistatic additive such that 1×10^{13} ohms surface resistance, maximum, shall be measured when tested in accordance with ASTM D257.

The nominal single film thickness shall be 0.0030 inch. The walls of the tubing shall be free of defects which will interfere with its proper use and function.

The tubing shall be tinted green to identify it with gas sterilization.

Instructions for using the tubing shall be provided with each roll including directions that tubing should not be used if it has had contact with strong acids or chlorinated and aromatic solvents at temperatures over 122 degrees Fahrenheit, to preclude chemical reaction and possible toxic by-products.

The tubing shall be furnished in a suitable, easy-to-use, dispenser type container.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

Workmanship. The tubing shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Roll (Ro). One roll of tubing of the size specified, and supplied in a dispenser type container, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6530-00-104-7552 (Size 1)
6530-00-104-7553 (Size 2)
6530-00-104-7564 (Size 3)
6530-00-104-7584 (Size 4)

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 05470

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DH

Project No. 6530-1811

Location H1-16-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.