COMMERCIAL ITEM DESCRIPTION

WAX, DENTAL, (Boxing)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification U-W-00138A.

This commercial item description covers dental boxing wax.

Salient characteristics:

The boxing wax shall be suitable for use as a frame to contain gypsum material in dental prosthetic procedures.

The wax shall be tacky, dark red in color, and shall cohere with finger pressure at room temperature. The wax shall be furnished in sheets from a minimum of 10-7/8 inches to a maximum of 12 inches long, 1-1/2 ± 1/8 inches wide, and 0.05 ± 0.01 inches thick. The layers of wax sheets shall be separated with silicon paper or other similar material.

The wax shall have a softening point between 150.0 degrees Fahrenheit (^{o}F.) and 172.0^{o}F. and shall be determined by method ASTM Designation E28-67 (Apparatus, Ring & Ball, Softening Point By).

The breaking point of the wax shall be determined as follows: a 6 ± 1/8 inch sample at 70.0^{o} ± 2.0^{o}F. shall be bent in 10 seconds at a uniform rate until it is bent double with the two halves of the specimen in contact for their entire length. The wax shall not fracture.

The working properties of the wax shall be determined by practical performance tests as follows: shall be pliable at 70.0^{o} + 2.0^{o}F., yet stiff enough to retain position and shape at 95.8^{o} ± 2.0^{o}F., and shall be capable of sealing easily to a plaster model with hand manipulation at 70.0^{o} ± 2.0^{o}F.

The unit of issue shall contain a minimum of one pound of boxing wax. The weight shall be determined by standard measuring instruments.

Material. The boxing wax shall be a uniform mixture of waxes, and other ingredients, such as resins, combined in proportions that will result in a material possessing the physical properties and characteristics specified above.

Workmanship. The wax shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Box (BX). One box containing one pound of wax, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. The unit of product for inspection purposes shall be one box of wax.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-00-567-2000	WAX, DENTAL, (Boxing)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

Project No. 6520-1883

Location: S2-12-14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.