A-A-51297
18 December 1985

COMMERCIAL ITEM DESCRIPTION

WIRE, NONELECTRICAL (Corrosion-Resisting Steel)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a nonelectrical wire, of corrosion-resisting steel, suitable for use in suturing and ligating in surgical procedures.

Salient characteristics:

The wire shall be suitable for use in suturing and ligating the intraoral tissues. The wire shall be a round corrosion-resisting steel wire of type AISI 304 and a minimum of 18 inches in length. The wire shall have a diameter of 0.020 inch with a minus tolerance of 0.0005 inch. The wire shall have a minimum tensile strength of 80,000 pounds per square inch (p.s.i.) and a maximum tensile strength of 120,000 p.s.i. in accordance with method ASTM E 8. The wire shall have a minimum elongation of 35 percent in a 10-inch gage length. The wire shall show no corrosion when subjected to the boil test of Federal Specification GG-I-526. The wire shall be furnished in the bright annealed or dead soft temper and shall have a bright, clean and smooth finish. The wire shall receive a passivation treatment after final processing.

Workmanship. The wire shall be free from cuts, scratches, burrs, crimps, or sharp edges which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty wires, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQ 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Chemical analysis shall be conducted in accordance with Methods 111 or 112 of Fed-Std-151.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6520-00-569-4150	WIRE, NONELECTRICAL, Corrosion-Resisting Steel, Round, 24 Gage, 18 Inch Length, 50s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

48225

Project No. 6520-1890

Location: S2-12-13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.