A-A-51301A
24 January 1992
SUPERSEDING
A-A-51301
20 December 1985

COMMERCIAL ITEM DESCRIPTION

FOOTWEAR COVERS, OPERATING ROOM

(Disposable)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for A-A-30042B which is cancelled.

This Commercial Item Description covers a disposable footwear cover for use by hospital personnel or visitors in operating rooms.

Salient Characteristics:

Type 1. The operating room footwear cover shall be a non-conductive, low-cut, moisture-resistant, disposable shoe cover suitable for one-time, short-term use. The footwear cover shall be fabricated of non-woven, moisture-proof, spun bonded olefin fabric, with an elastic binding stitched to the cover opening, and an elastic band stretched into a full length sole seam, or as an alternate, an elastic band stitched from the front of the sole seam over the toe of the cover to the elasticized opening. The construction shall be closed-back style, and be such that the cover can be worn over any shoe greater than men's size 13, held firmly to the shoe by the elastic binding and band.

The spun bonded olefin shall have no known irritating or toxic elements, shall be free of lint and objectionable odor, and shall have a minimum breaking strength of six pounds in the machine direction and nine pounds in the cross direction. The cover shall be blue, green, aqua, or other pastel shade.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

The elastic binding shall be stitched to the cover opening, so that the opening shall measure approximately nine inches in diameter when manually stretched to its extreme opening without breaking. The stitching shall be free of defects. The elastic binding and the elastic band shall be of a suitable width and sufficient elasticity and strength to secure the cover to the shoe with a snug fit during use, without deterioration of the cover.

Thread breaks, skip, and run-off shall be overstitched not less than 1/2 inch. Skipped stitches or thread breaks shall be repaired. The thread tension shall be maintained, so that there shall not be any loose or tight stitching, and the lock shall be embedded in the materials sewn together. All thread ends shall be trimmed to a maximum length of one quarter inch. The cover shall not contain any metal, spot, stain, non-specified hole, abraded area, tear, cut mend, patch, scorch, or needle chew.

Type 2. The footwear cover shall be a disposable, non-conductive, nonskid, nonwoven synthetic material, without tears or flaws in the material, or excessive threads. The cover shall be moisture resistant with an elasticized top which allows the cover to hold firmly to the shoe by means of the elastic binding and band. The cover shall have an opening to allow for "one size fits all" with a secure, soft, and comfortable fit.

Workmanship. The operating room footwear cover shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

Type 1. Package (PG). One package of 100 pairs of footwear covers, as specified, constitutes one unit.

Type 2. Package (PG). One package of 300 pairs of footwear covers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level II. The unit of product for examination purposes shall be one pair of footwear covers.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one pair of footwear covers.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6532-01-152-2796 (Type 1)
6532-01-301-2291 (Type 2)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA-MPQAS

Project Number: 6532-1220

Location: FOOTMC13.DOC

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.