COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS
16 Gage Catheter, 5-1/4 inches Long, 50s

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, non-toxic, pyrogen-free, radiopaque, disposable, catheter and needle placement unit, suitable for the administration of intravenous medication during special procedures.

Salient characteristics:

Materials. The indwelling catheter shall be fabricated from clear or translucent, pliant plastic suitable in all respects for the purpose intended. The catheter shall be radiopaque throughout its length. The catheter, and all plastic components coming in contact with intravenous medication and patient's blood shall contain no more than 5 parts per million of lead. The finished catheter shall be sterile, pyrogen-free, and meet the requirements of the USP for class IV plastics. The catheter shall meet the sterility requirement when tested in accordance with the USP.

The needle cannula shall be fabricated from austenitic corrosion-resisting steel in accordance with ASTM F899.

The catheter Luer hub, needle hub and flow control plug shall be fabricated from suitable plastics meeting the requirements of the USP for class II plastics.

Catheter and needle point protector shall be fabricated from a suitable plastic.

Construction. The unit shall permit aseptic insertion by surgeon without assistance and without the necessity of hand scrubbing or wearing of sterile gloves. The unit shall consist of an outer plastic 16 gage catheter permanently secured to a plastic Luer lock hub.

The catheter hub shall be in accordance with ANSI/HIMA MD70.1. The catheter length shall be 5-1/4 ± 1/8 inches.

The inner needle cannula shall be permanently secured to a transparent hub that shall accept male, tapered devices that are in accordance with ANSI/HIMA MD70...

The hub of the needle bevel shall extend beyond the catheter tip when the needle is fully inserted into the catheter.

The transparent needle hub shall possess raised, ribbed-surfaces or other features that shall provide for good gripping technique. A flow control plug shall be assembled to the needle hub.

The needle cannula shall make a free sliding fit within the catheter.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Each unit shall be provided with a component necessary to protect the entire length of the catheter and needle cannula during handling.

Catheters shall not pull out of, of loosen from, the catheter hub when ubjected to an axial force of 3.0 pounds. Catheters shall not exhibit leakage exceeding one clinging drop (hydrostatic test) or one clinging air bubble (pneumatic test) when tested with appropriate fixtures and taper test gages. The criteria shall be as follows:

Hydrostatic test - 20 pounds per square inch gage (psig) pressure applied for 2 minutes.

Pneumatic test - 10 psig applied for 1 minute.

The needle cannula shall not pull out of, or loosen from, the needle hub when subjected to an axial force of 10 pounds.

Precautionary statement. A caution note shall appear on the label that shall state the following: "Do not reinsert needle in catheter if venipuncture is not successful" (or similar wording).

Workmanship. The catheter and needle placement unit shall be free from defects which detract from its appearance or impair their serviceability.

Unit. Package (PG). One package containing fifty catheter and needle units, each unit individually packaged in a container capable of maintaining sterility of contents unless package is damaged or opened, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one package of catheter and needle units.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to contents and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Number is covered by this document:

6515-00-216-5107

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51302

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

59480

Project No. 6515-3440

Location S1-10-7

U S GOVERNMENT PRINTING OFFICE 1985 — 605 036/40673

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.