A-A-51303
19 December 1985

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS, 16 Gage Catheter, 8 Inches Long, 50s

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, pyrogen-free, radiopaque, disposable, catheter and needle unit suitable for long term intravenous therapy.

Salient characteristics

Materials. The indwelling catheter shall be fabricated from pliant fluorinated ethylene propylene (FEP), of the best quality and suitable in all respects for the purpose intended. The catheter and all plastic components coming in contact with intravenous medication and patient's blood shall contain no more than 5 parts per million of lead. The finished catheter shall be sterile, radiopaque pyrogen-free, non-toxic and meet the requirements of the USP for class IV plastics. The catheter shall meet the sterility requirements when tested in accordance with the USP.

The needle cannula shall be fabricated from class 3 type 304 austenitic corrosion-resistant steel of the composition listed in accordance with ASTM F899.

The catheter Luer hub and flow control plug shall be fabricated from suitable plastics meeting the requirements of the USP for class II plastics.

Catheter and needle point protector shall be fabricated from a suitable plastic.

Construction. The unit shall permit aseptic insertion by the surgeon without assistance and without the necessity of hand-scrubbing or wearing sterile gloves. Shall consist of an outer plastic 16 gauge catheter permanently secured to a plastic Luer-lock hub. The catheter hub shall be in accordance with ANSI/HIMA MD70.1. The catheter length shall be 8 ± 1/4 inches.

The heel of the needle bevel shall extend beyond the catheter tip when the needle is fully inserted into the catheter. A flow control or self-venting plug shall be assembled to the needle hub.

The needle cannula shall make a free-sliding fit within the catheter. Needle point shall be well defined.

Catheters shall not pull out or loosen from the catheter hub when subjected to an axial force of 3.0 pounds. Catheters shall not exhibit leakage exceeding one clinging drop (hydrostatic test) when tested with appropriate fixtures and taper test gauges. The criteria shall be as follows:

Hydrostatic test Twenty pounds per square inch gauge (psig) pressure applied for two minutes.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:54Z
Check the source to verify that this is the current version before use.

Pneumatic test: Ten psig pressure applied for one minute. The needle cannula shall not pull out of, or loosen from, the needle hub when subjected to an axial force of ten pounds.

Precautionary statement. A caution note shall appear on the label that shall state the following "Once needle is withdrawn do not reinsert into catheter" (or similar wording).

Instructions. Instructions for use shall be provided in each unit package.

Workmanship. The catheter and needle shall be free from defects which detract from its appearance or impair its serviceability

Unit. Package (PG). One package containing fifty individually packaged catheter and needle units, each unit individually packaged in a container capable of maintaining sterility of contents unless package is damaged or opened, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one package of catheter and needle units.

Source: https://assist.dla.mil -- Downloaded: 2026-04-26T08:54Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-161-7004

MILITARY INTERESTS

Custodians.

Army-MD
Navy-MS
Air Force-CD

PREPARING ACTIVITY

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS

51815

Project No. 6515-3442

Location S1-10-8

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:54Z
Check the source to verify that this is the current version before use.