A-A-51304A
14 February 1991
SUPERSEDING
A-A-51304
20 December 1985

COMMERCIAL ITEM DESCRIPTION

CHEST, X-RAY FILM PROTECTIVE

(Steel)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Interim Federal Specification GG-C-00225 which is cancelled.

This Commercial Item Description covers a steel x-ray film protective chest.

Salient Characteristics:

The x-ray film protective chest shall be suitable for the storage of x-ray film and cassettes. The chest shall be furnished in the following two types:

Type I - Mounted on 4 casters, with minimum inside dimensions of 19 x 10 x 8 inches.

Type II - Without casters, with minimum inside dimensions of 18 x 15-3/4 x 11-1/4 inches.

The chest shall consist of an inner shell, a lead barrier layer, an outer shell, and a cover.

Type I Chest.

The inner shell shall be 24 gage (0.025 inch) sheet steel minimum. Sections of the inner shell shall be welded together The inner shell shall be completely covered on the end, sides, and bottom with 1/16 inch thick sheet lead minimum. The sheet lead shall be completely free of pinholes.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The seams and joints of the chest shall have lead protection equivalent to the body material. The outer shell shall be 20 gage (0 037 inch) sheet steel minimum Sections of the outer shell shall be welded together

The chest outer shell shall have four casters mounted on the bottom The casters shall be so located that when turned to the worst possible position, and with the cover open, the empty chest shall not be unstable The casters shall be welded in place, or bolted. Sheet metal, or self-tapping screws of any type, shall not be acceptable Bolts used shall not be smaller than size #8-32, and there shall be at least three bolts for each caster.

Type II Chest.

The inner shell and the outer shell shall both be 20 gage (0 037 inch) sheet steel minimum. A minimum 1/16 inch thick sheet of lead shall be located between the inner and outer shells The sheet lead shall be completely free of pinholes. Seams and joints of the chest shall have lead protection equivalent to the body material.

Type I and Type II Chests.

The cover for both chest types shall consist of a minimum 24 gage (0 025 inch) sheet steel inner surface, a minimum 20 gage (0.037 inch) sheet steel outer surface, and a 1/16 inch thick lead sheet between the inner and outer surfaces. The cover shall be mounted to the body of the chest by means of a continuous hinge, or hinges of the 5-knuckle type, 2-1/2 inch size. The hinges shall be bolted or spot-welded in position. Sheet metal, or self-tapping screws of any type, are not acceptable. The cover shall be held open at an angle of 120 degrees plus or minus 10 degrees by means of a check. A drawer pull shall be welded or bolted to the cover.

All seam and fillet welds shall be continuous, even, and smooth. Spot welds shall not be at more than 2 inch centers. All exposed seam and fillet welds shall be ground smooth to eliminate traces of weld marks, and shall be polished to blend into adjacent surfaces. All exposed spot welds shall be ground smooth, or filled and sanded smooth, so that they will not be visible after the final finish is applied.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The chest shall be equipped with 3 lift handles, one on each side, and one on the hinged cover.

All inside and outside exposed surfaces of the chest shall be degreased and finished with baked enamel paint or an epoxy powder coating If enamel is used, a primer coat shall be applied to all external and internal exposed surfaces as well as unexposed surfaces inside the walls of the chest The final chest finish shall be the manufacturer's standard color and shall be hard, smooth, and highly resistant to abrasion, cracking, and chipping

Workmanship The x-ray film protective chest shall be free from defects which detract from its appearance or impair its serviceability.

Unit Each (EA) One chest of the type specified, constitutes one unit.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4 0 (percent defective) and an inspection level of S-1 The unit of product for examination purposes shall be one x-ray film protective chest

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2 5 (percent defective) and an inspection level of S-3 The unit of product for test purposes shall be one x-ray film protective chest

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE. The following National Stock Numbers are covered by this document.

6525-00-601-2120 (Type I)
6525-00-601-2124 (Type II)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

A-A-51304A

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6525-1263

Location CHESTCID 13A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-51304A	2. DOCUMENT DATE (YYMMDD): 14 February 1991

3. DOCUMENT TITLE: CHEST, X-RAY FILM PROTECTIVE, (Steel)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSTE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.