A-A-51307
20 December 1985

COMMERCIAL ITEM DESCRIPTION

DISH, CULTURE, PETRI,

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-D-36425B.

This commercial item description covers a sterile, plastic, disposable Petri culture dish, with a tight lid, for use as a container for a filter disk used in membrane filtering technique for bacteriological analysis of water under field conditions.

Salient characteristics:

The Petri dish shall be complete, top and bottom, sterile and disposable. The Petri dish shall be a suitable container for a Disk, Filtering, Microporous, (NSN 6640-00-299-8692), used with Water Testing Kit, Bacteriological, (NSN 6665-000-682-4765), for bacteriological analysis of water in the field.

The Petri dish shall be fabricated of optically clear, scratch-free, polystyrene, free of toxic lubricants and other additives which would adversely affect the growth of bacterial cultures.

The Petri dish shall consist of a top section (cover) and a bottom section. The top section shall consist of a flat, circular top merging into a continuous wall, slightly tapered to fit telescopically with the upstanding wall of the bottom section, forming a tight seal by means of a wedge fit. The bottom section shall have a flat, circular bottom and an upstanding continuous wall, slightly thicker lower part, tapering upward to fit the top section of the dish. An annular shoulder shall be formed on the lower part of the bottom section to provide a secure closure of the Petri dish. The bottom section shall be constructed so that it sits flat on a plane and level surface, and provides for a uniform media depth.

The bottom section shall be provided with 3 or 4 radial lugs, slits or depressions to aid in opening the closed dish.

The closed dimensions shall be 2-3/32 $\pm$ 1/16 inches in diameter and 0.035 to 0.050 inches in height.

The inside of the Petri dish shall be sterile. The method and efficacy of sterilization shall be acceptable to the Food and Drug Administration (FDA).

Workmanship. The Petri dish shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). Twenty dishes (tops and bottoms), shall be packaged in a sealed container which will maintain sterility until manually opened. One package containing twenty-five containers of twenty dishes (500 dishes), as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one Petri dish (top and bottom, complete).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein:

Falcon Labware Part No. 1006

The following National Stock Number is covered by this document:

6640-01-030-9012

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

56436

Project No. 6640-1362

Location S3-8-20

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51307	Dish, Culture, Petri

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording.

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.