A-A-51308
12 March 1986

COMMERCIAL ITEM DESCRIPTION

SPLINT, ARM and SPLINT, LEG (Pneumatic)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-S-37025A.

This commercial item description covers transparent, plastic, mouth inflatable, pneumatic splints, with full length zipper closure, for the emergency treatment of fractures or superficial bleeding of the full adult arm or leg.

Salient characteristics.

The splint shall be adult size, double-wall plastic construction, with full length zipper closure, and a valve with mouthpiece for inflation by mouth. The splint shall be one of the following types:

Type I - Splint, Arm, Pneumatic, Adult, Zipper Closure.
Type II - Splint, Leg, Open Toe, Pneumatic, Adult, Zipper Closure.

The splint shall be constructed in accordance with DPSC Drawing No. 23187 to the revision in effect on the date of solicitation.

Materials used shall contain no ingredients known to be skin sensitizers or irritants.

The splint shall be constructed of clear, transparent film conforming to the requirements of Federal Specification L-P-375 Type II Class 1, except for the following paragraphs of L-P-375:

3.3.2 Type II, Table II. Delete the requirements for Stiffness and Blocking. Cold Crack Test delete -40 degrees and substitute 0 degrees.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

3.7 Finish. Delete entirely and substitute:

"3.7 Finish. The film shall be double polished, conforming to a 60 degree specular gloss of 140 units on each surface when tested on a dull background as specified in ASTM-D-2457."

The outer wall shall be 12 gage (0.012 inch nominal), the inner wall shall be 8 gage (0.008 inch nominal), and the zipper reinforcing strips shall be between 0.008 and 0.020 inch material thickness.

The zipper closure shall conform to the requirements of Federal Specification V-F-106 except it shall be fabricated of a polyester coil with white enameled die cast zinc slide and end stops, size L, LS or LM. The arm splint shall be Type I Style 3, nonseparating, nonreversible. The leg splint shall be Type IV Style 8, separating, nonreversible. Tape shall be 9/16 inch wide, white open weave (Leno) cotton or white cotton impregnated with a clear vinyl plastic.

The leg splint zipper pull shall include a thong fabricated of nylon cord conforming to Type IA (coreless) of MIL-C-5040. The Core may be removed. The ends of the cord shall be tied by means of a non-slippable knot to form a loop 2 inches ± 1/4 inch in length. Strands of cord ends shall be fused.

Grommets on the leg splint shall be fabricated from polyvinyl chloride. Shall be Halkey-Roberts Inc. No. 44GF or No. 44GH, or equal.

The valve for opening and closing the air inflation passage shall be a screw type with knurled external member to which a plastic mouthpiece is assembled. The valve housing shall be fabricated from a polyvinyl chloride compatible with the splint plastic. The valves shall be Halkey-Roberts No. 40 ASE or equal for the arm splint and No. 240 ASE or equal for the leg splint.

The inner and outer walls, zipper reinforcing strips, zipper tapes and valve assembly shall be heat sealed to make an air tight end item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Each splint shall be loosely folded in half lengthwise in such a manner that the fold is not creased. A cushioning medium such as plastic air bubbles or other suitable material may be placed between the folds of the splint to prevent the folded edge from being creased. With the zipper in the closed position, the folded splint shall be inserted into the immediate container which shall be an air-tight bag fabricated of polyethylene having a minimum thickness of 0.002 inch. The bag shall be heat sealed after excess air has been evacuated.

Instructions for use shall be permanently marked on the immediate container. In addition, the following statement shall be located below the instructions for use in 3/16 minimum capital letters:

> "CAUTION - USE IN AIRCRAFT REQUIRES MONITORING THE SPLINT PRESSURE TO PREVENT NEUROVASCULAR INJURY. LIMIT INFLATION PRESSURE TO ADEQUATE FRACTURE SUPPORT ONLY."

The following information shall be printed in 1/8 inch minimum letters:

> "The aneroid pressure gage of a standard sphygmomanometer should be available for monitoring the splint pressure during ascent and descent. Splints air transported should be inflated to provide adequate support of the fractured limb (below 50 mm Hg). Over-inflation is to be specifically avoided."

As an alternate, the above markings shall be legibly printed on a separate sheet of paper which shall be inserted in the immediate container in such a manner as to permit ready identification.

Leakage Test. The splint shall be tested for leakage by closing the zipper inflating the splint to a pressure of 60 mm Hg, closing the valve and immersing in water. There shall be no air bubble escape from the immersed splint.

Reference specification. Where Military or Federal Specifications, ASTM Stnadards or drawings are referenced herein, the issue of that specification standard or drawing in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Workmanship. The splint shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One splint, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for tests purposes shall be one splint.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one splint.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-935-6592 SPLINT, ARM, Pneumatic, Adult, Zipper Closure
6515-00-935-6593 SPLINT, LEG, Pneumatic, Adult, Zipper Closure

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51308

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PBS
FDA-MPQAS

Project No. 6515-3526

Location: CID/SPLINT/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.